

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

ProjectID: 98 Morse Ave Nutley

Lab Code: ACE

SDG No.: Q2486

Calibration Sequence : FG070925		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	164717398	96893	FG016264.D
850	83171210	97848	FG016265.D
340	35101863	103241	FG016266.D
170	18639349	109643	FG016267.D
85	9352448	110029	FG016268.D
AVG RF : 103531		% RSD : 6.032	AVG RT : 15.201