

Surrogate Summary

SDG No.: Q2486

Client: Sciacca General Contractors, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2486-02	VOC	1,2-Dichloroethane-d4	50	49.6	99		63	155
		Dibromofluoromethane	50	52.3	105		70	134
		Toluene-d8	50	51.1	102		74	123
		4-Bromofluorobenzene	50	52.3	105		17	146
VY0703SBL01	VY0703SBL01	1,2-Dichloroethane-d4	50	52.7	105		63	155
		Dibromofluoromethane	50	52.7	105		70	134
		Toluene-d8	50	52.1	104		74	123
		4-Bromofluorobenzene	50	55.3	111		17	146
VY0703SBS01	VY0703SBS01	1,2-Dichloroethane-d4	50	48.4	97		63	155
		Dibromofluoromethane	50	50.6	101		70	134
		Toluene-d8	50	51.1	102		74	123
		4-Bromofluorobenzene	50	47.0	94		17	146
VY0703SBSD01	VY0703SBSD01	1,2-Dichloroethane-d4	50	52.1	104		63	155
		Dibromofluoromethane	50	53.0	106		70	134
		Toluene-d8	50	52.1	104		74	123
		4-Bromofluorobenzene	50	49.1	98		17	146