



## **CASE NARRATIVE**

### **Walsh Construction Company II, LLC**

**Project Name: Construction of Shafts 17B-18B - PN 220084**

**Project # N/A**

**Order ID # Q2487**

**Test Name: VOC-TCLVOA-10**

#### **A. Number of Samples and Date of Receipt:**

24 Solid samples were received on 07/01/2025.

#### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: Ammonia, COD, Corrosivity, Cyanide, Diesel Range Organics, EPH\_NF, Gasoline Range Organics, Herbicide, Hexavalent Chromium, Ignitability, Mercury, Metals ICP-TAL, METALS TAL+CN, Oil and Grease, Paint Filter, PCB, Pesticide-TCL, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SPLP BNA, SPLP Extraction, SPLP Herbicide, SPLP ICP Metals, SPLP Mercury, SPLP Pesticide, SPLP VOA, SPLP ZHE Ext, SPLP-FULL, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Metals+Cu+Ni+Zn, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, TCLP Metals Group1, TKN, TPH GC, Trivalent Chromium, TS, TVS and VOC-TCLVOA-10. This data package contains results for VOC-TCLVOA-10.

#### **C. Analytical Techniques:**

The analysis performed on instrument MSVOA\_Y were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868. The analysis of VOC-TCLVOA-10 was based on method 8260D.

#### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for G1(4.5) [1,2-Dichloroethane-d4 - 62%] sample was reanalyzed to confirm the failure and reported.

The Internal Standards Areas met the acceptable requirements except for G2(2.5), G2(2.5)RE, G1(4.5) and G1(4.5)RE sample was reanalyzed to confirm the failure and reported.

The Retention Times were acceptable for all samples.

The RPD met criteria.



The Blank Spike for {VY0707SBS01} with File ID: VY022948.D met requirements for all samples except for Methylene Chloride[142%] are failing high and associate sample having hit of acetone but below CRQL therefore no corrective action taken.

The Blank Spike Duplicate met requirements for all samples.  
The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration File ID VY022903.D met the requirements except for Acetone are failing high and associate sample having hit of acetone but below CRQL therefore no corrective action taken.

The Tuning criteria met requirements.

**E. Additional Comments:**

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

The soil samples results are based on a dry weight basis.  
Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature \_\_\_\_\_