

Initial and Continuing Calibration Verification

Client: Walsh Construction Company II, LLC
Project: Construction of Shafts 17B-18B - PN 220084

SDG No.: Q2487
RunNo.: LB136387

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.097	0.099	98	90-110	07/07/2025
Sample ID: CCV1 Cyanide	mg/L	0.25	0.25	100	90-110	07/07/2025
Sample ID: CCV2 Cyanide	mg/L	0.25	0.25	100	90-110	07/07/2025
Sample ID: CCV3 Cyanide	mg/L	0.25	0.25	100	90-110	07/07/2025

Initial and Continuing Calibration Verification

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Project: Construction of Shafts 17B-18B - PN 220084

RunNo.: LB136397

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.493	0.5	99	90-110	07/08/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.496	0.5	99	90-110	07/08/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.494	0.5	99	90-110	07/08/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.496	0.5	99	90-110	07/08/2025

Initial and Continuing Calibration Verification

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Project: Construction of Shafts 17B-18B - PN 220084

RunNo.: LB136399

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.962	50	102	95-105	05/28/2025
Sample ID: CCV1 COD	mg/L	50.962	50	102	95-105	07/08/2025
Sample ID: CCV2 COD	mg/L	49.946	50	100	95-105	07/08/2025
Sample ID: CCV3 COD	mg/L	50.962	50	102	95-105	07/08/2025

Initial and Continuing Calibration Verification

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Project: Construction of Shafts 17B-18B - PN 220084

RunNo.: LB136405

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.99	1	99	90-110	07/09/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.94	1	94	90-110	07/09/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.96	1	96	90-110	07/09/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	07/09/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.96	1	96	90-110	07/09/2025
Sample ID: CCV5 Ammonia as N	mg/L	0.99	1	99	90-110	07/09/2025

Initial and Continuing Calibration Verification

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Project: Construction of Shafts 17B-18B - PN 220084

RunNo.: LB136408

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TKN	mg/L	4.9	5	98	90-110	07/09/2025
Sample ID: CCV1 TKN	mg/L	4.9	5	98	90-110	07/09/2025
Sample ID: CCV2 TKN	mg/L	4.9	5	98	90-110	07/09/2025
Sample ID: CCV3 TKN	mg/L	4.9	5	98	90-110	07/09/2025
Sample ID: CCV4 TKN	mg/L	4.9	5	98	90-110	07/09/2025
Sample ID: CCV5 TKN	mg/L	4.9	5	98	90-110	07/09/2025

Initial and Continuing Calibration Verification

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Project: Construction of Shafts 17B-18B - PN 220084

RunNo.: LB136408

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------