

Initial and Continuing Calibration Blank Summary

Client: Walsh Construction Company II, LLC
Project: Construction of Shafts 17B-18B - PN 220084

SDG No.: Q2487
RunNo.: LB136387

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/07/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/07/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/07/2025
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/07/2025

Initial and Continuing Calibration Blank Summary

Client: Walsh Construction Company II, LLC
Project: Construction of Shafts 17B-18B - PN 220084

SDG No.: Q2487
RunNo.: LB136397

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	07/08/2025
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	07/08/2025
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	07/08/2025
Sample ID: CCB3 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	07/08/2025

Initial and Continuing Calibration Blank Summary

Client: Walsh Construction Company II, LLC
Project: Construction of Shafts 17B-18B - PN 220084

SDG No.: Q2487
RunNo.: LB136399

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB COD	mg/L	< 5.0000	5.0000	U	1.50	10	05/28/2025
Sample ID: CCB1 COD	mg/L	< 5.0000	5.0000	U	1.50	10	07/08/2025
Sample ID: CCB2 COD	mg/L	< 5.0000	5.0000	U	1.50	10	07/08/2025
Sample ID: CCB3 COD	mg/L	< 5.0000	5.0000	U	1.50	10	07/08/2025

Initial and Continuing Calibration Blank Summary

Client: Walsh Construction Company II, LLC
Project: Construction of Shafts 17B-18B - PN 220084

SDG No.: Q2487
RunNo.: LB136405

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	07/09/2025
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	07/09/2025
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	07/09/2025
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	07/09/2025
Sample ID: CCB4 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	07/09/2025
Sample ID: CCB5 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	07/09/2025

Initial and Continuing Calibration Blank Summary

Client: Walsh Construction Company II, LLC
Project: Construction of Shafts 17B-18B - PN 220084

SDG No.: Q2487
RunNo.: LB136408

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 TKN	mg/L	< 0.2500	0.2500	U	0.11	0.5	07/09/2025
Sample ID: CCB1 TKN	mg/L	< 0.2500	0.2500	U	0.11	0.5	07/09/2025
Sample ID: CCB2 TKN	mg/L	< 0.2500	0.2500	U	0.11	0.5	07/09/2025
Sample ID: CCB3 TKN	mg/L	< 0.2500	0.2500	U	0.11	0.5	07/09/2025
Sample ID: CCB4 TKN	mg/L	< 0.2500	0.2500	U	0.11	0.5	07/09/2025
Sample ID: CCB5 TKN	mg/L	< 0.2500	0.2500	U	0.11	0.5	07/09/2025

Initial and Continuing Calibration Blank Summary

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Project: Construction of Shafts 17B-18B - PN 220084

RunNo.: LB136408

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------