

Laboratory Control Sample Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q2487
Project:	Construction of Shafts 17B-18B - PN 220084	Run No.:	LB136399

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136399BS							
COD	mg/Kg	2500	2600		104	1	90-110	07/08/2025

Laboratory Control Sample Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q2487
Project:	Construction of Shafts 17B-18B - PN 220084	Run No.:	LB136422

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136422BS							
Oil and Grease	mg/Kg	100	94.9		95	1	80-120	07/10/2025

Laboratory Control Sample Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q2487
Project:	Construction of Shafts 17B-18B - PN 220084	Run No.:	LB136408

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168731BS							
TKN	mg/Kg	250	245		98	1	90-110	07/09/2025

Laboratory Control Sample Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q2487
Project:	Construction of Shafts 17B-18B - PN 220084	Run No.:	LB136397

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168732BS							
Hexavalent Chromium	mg/Kg	20	19.9		100	1	84-110	07/08/2025

Laboratory Control Sample Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q2487
Project:	Construction of Shafts 17B-18B - PN 220084	Run No.:	LB136387

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168749BS							
Cyanide	mg/Kg	5	4.80		96	1	85-115	07/07/2025

Laboratory Control Sample Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q2487
Project:	Construction of Shafts 17B-18B - PN 220084	Run No.:	LB136405

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168757BS							
Ammonia as N	mg/Kg	50	49.1		98	1	90-110	07/09/2025