



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Contract: WALS01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Aluminum	109	100	109	65 - 135	P	07/07/2025	13:49	LB136390
	Antimony	52.7	50.0	106	65 - 135	P	07/07/2025	13:49	LB136390
	Arsenic	23.0	20.0	115	65 - 135	P	07/07/2025	13:49	LB136390
	Barium	97.5	100	98	65 - 135	P	07/07/2025	13:49	LB136390
	Beryllium	6.07	6.0	101	65 - 135	P	07/07/2025	13:49	LB136390
	Cadmium	7.08	6.0	118	65 - 135	P	07/07/2025	13:49	LB136390
	Calcium	2020	2000	101	65 - 135	P	07/07/2025	13:49	LB136390
	Chromium	11.8	10.0	118	65 - 135	P	07/07/2025	13:49	LB136390
	Cobalt	30.6	30.0	102	65 - 135	P	07/07/2025	13:49	LB136390
	Copper	22.1	20.0	111	65 - 135	P	07/07/2025	13:49	LB136390
	Iron	117	100	117	65 - 135	P	07/07/2025	13:49	LB136390
	Lead	13.7	12.0	114	65 - 135	P	07/07/2025	13:49	LB136390
	Magnesium	2100	2000	105	65 - 135	P	07/07/2025	13:49	LB136390
	Manganese	21.9	20.0	110	65 - 135	P	07/07/2025	13:49	LB136390
	Nickel	42.1	40.0	105	65 - 135	P	07/07/2025	13:49	LB136390
	Potassium	1880	2000	94	65 - 135	P	07/07/2025	13:49	LB136390
	Selenium	19.2	20.0	96	65 - 135	P	07/07/2025	13:49	LB136390
	Silver	10.6	10.0	106	65 - 135	P	07/07/2025	13:49	LB136390
	Sodium	2070	2000	104	65 - 135	P	07/07/2025	13:49	LB136390
	Thallium	42.1	40.0	105	65 - 135	P	07/07/2025	13:49	LB136390
	Vanadium	42.5	40.0	106	65 - 135	P	07/07/2025	13:49	LB136390
	Zinc	46.7	40.0	117	65 - 135	P	07/07/2025	13:49	LB136390
CRI01	Aluminum	99.5	100	100	65 - 135	P	07/08/2025	14:22	LB136400
	Antimony	49.2	50.0	98	65 - 135	P	07/08/2025	14:22	LB136400
	Arsenic	20.6	20.0	103	65 - 135	P	07/08/2025	14:22	LB136400
	Barium	84.3	100	84	65 - 135	P	07/08/2025	14:22	LB136400
	Beryllium	5.81	6.0	97	65 - 135	P	07/08/2025	14:22	LB136400
	Cadmium	5.85	6.0	98	65 - 135	P	07/08/2025	14:22	LB136400
	Calcium	1910	2000	95	65 - 135	P	07/08/2025	14:22	LB136400
	Chromium	11.1	10.0	111	65 - 135	P	07/08/2025	14:22	LB136400
	Cobalt	29.0	30.0	97	65 - 135	P	07/08/2025	14:22	LB136400
	Copper	20.8	20.0	104	65 - 135	P	07/08/2025	14:22	LB136400
	Iron	117	100	117	65 - 135	P	07/08/2025	14:22	LB136400
	Lead	12.3	12.0	102	65 - 135	P	07/08/2025	14:22	LB136400
	Magnesium	1980	2000	99	65 - 135	P	07/08/2025	14:22	LB136400
	Manganese	20.8	20.0	104	65 - 135	P	07/08/2025	14:22	LB136400

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Contract: WALS01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Nickel	39.6	40.0	99	65 - 135	P	07/08/2025	14:22	LB136400
	Potassium	1940	2000	97	65 - 135	P	07/08/2025	14:22	LB136400
	Selenium	22.5	20.0	113	65 - 135	P	07/08/2025	14:22	LB136400
	Silver	9.90	10.0	99	65 - 135	P	07/08/2025	14:22	LB136400
	Sodium	2000	2000	100	65 - 135	P	07/08/2025	14:22	LB136400
	Thallium	38.8	40.0	97	65 - 135	P	07/08/2025	14:22	LB136400
	Vanadium	37.3	40.0	93	65 - 135	P	07/08/2025	14:22	LB136400
	Zinc	43.2	40.0	108	65 - 135	P	07/08/2025	14:22	LB136400
CRI01	Aluminum	114	100	114	65 - 135	P	07/09/2025	10:26	LB136407
	Antimony	53.6	50.0	107	65 - 135	P	07/09/2025	10:26	LB136407
	Arsenic	14.2	20.0	71	65 - 135	P	07/09/2025	10:26	LB136407
	Barium	104	100	104	65 - 135	P	07/09/2025	10:26	LB136407
	Beryllium	6.33	6.0	105	65 - 135	P	07/09/2025	10:26	LB136407
	Cadmium	5.51	6.0	92	65 - 135	P	07/09/2025	10:26	LB136407
	Calcium	1930	2000	96	65 - 135	P	07/09/2025	10:26	LB136407
	Chromium	10.7	10.0	107	65 - 135	P	07/09/2025	10:26	LB136407
	Cobalt	29.1	30.0	97	65 - 135	P	07/09/2025	10:26	LB136407
	Copper	24.1	20.0	120	65 - 135	P	07/09/2025	10:26	LB136407
	Iron	119	100	119	65 - 135	P	07/09/2025	10:26	LB136407
	Lead	11.5	12.0	96	65 - 135	P	07/09/2025	10:26	LB136407
	Magnesium	2060	2000	103	65 - 135	P	07/09/2025	10:26	LB136407
	Manganese	22.6	20.0	113	65 - 135	P	07/09/2025	10:26	LB136407
	Nickel	38.2	40.0	96	65 - 135	P	07/09/2025	10:26	LB136407
	Potassium	1710	2000	86	65 - 135	P	07/09/2025	10:26	LB136407
	Selenium	22.8	20.0	114	65 - 135	P	07/09/2025	10:26	LB136407
	Silver	11.6	10.0	116	65 - 135	P	07/09/2025	10:26	LB136407
	Sodium	1830	2000	92	65 - 135	P	07/09/2025	10:26	LB136407
	Thallium	38.4	40.0	96	65 - 135	P	07/09/2025	10:26	LB136407
	Vanadium	32.0	40.0	80	65 - 135	P	07/09/2025	10:26	LB136407
	Zinc	43.1	40.0	108	65 - 135	P	07/09/2025	10:26	LB136407
CRI01	Aluminum	85.6	100	86	65 - 135	P	07/10/2025	19:01	LB136434
	Antimony	51.2	50.0	102	65 - 135	P	07/10/2025	19:01	LB136434
	Arsenic	20.6	20.0	103	65 - 135	P	07/10/2025	19:01	LB136434
	Barium	96.3	100	96	65 - 135	P	07/10/2025	19:01	LB136434
	Beryllium	6.31	6.0	105	65 - 135	P	07/10/2025	19:01	LB136434
	Cadmium	5.33	6.0	89	65 - 135	P	07/10/2025	19:01	LB136434
	Calcium	2070	2000	104	65 - 135	P	07/10/2025	19:01	LB136434
	Chromium	10.5	10.0	105	65 - 135	P	07/10/2025	19:01	LB136434

Metals
- 2b -
CRDL STANDARD FOR AA & ICP

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Contract: WALS01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cobalt	28.4	30.0	94	65 - 135	P	07/10/2025	19:01	LB136434
	Copper	20.6	20.0	103	65 - 135	P	07/10/2025	19:01	LB136434
	Iron	116	100	116	65 - 135	P	07/10/2025	19:01	LB136434
	Lead	12.4	12.0	103	65 - 135	P	07/10/2025	19:01	LB136434
	Magnesium	2140	2000	107	65 - 135	P	07/10/2025	19:01	LB136434
	Manganese	21.9	20.0	110	65 - 135	P	07/10/2025	19:01	LB136434
	Nickel	37.1	40.0	93	65 - 135	P	07/10/2025	19:01	LB136434
	Potassium	1770	2000	89	65 - 135	P	07/10/2025	19:01	LB136434
	Selenium	16.7	20.0	84	65 - 135	P	07/10/2025	19:01	LB136434
	Silver	10.5	10.0	105	65 - 135	P	07/10/2025	19:01	LB136434
	Sodium	1780	2000	89	65 - 135	P	07/10/2025	19:01	LB136434
	Thallium	37.7	40.0	94	65 - 135	P	07/10/2025	19:01	LB136434
	Vanadium	28.3	40.0	71	65 - 135	P	07/10/2025	19:01	LB136434
	Zinc	44.6	40.0	112	65 - 135	P	07/10/2025	19:01	LB136434
CRI01	Aluminum	126	100	126	65 - 135	P	07/14/2025	14:26	LB136469
	Antimony	49.5	50.0	99	65 - 135	P	07/14/2025	14:26	LB136469
	Arsenic	16.7	20.0	83	65 - 135	P	07/14/2025	14:26	LB136469
	Barium	99.1	100	99	65 - 135	P	07/14/2025	14:26	LB136469
	Beryllium	6.42	6.0	107	65 - 135	P	07/14/2025	14:26	LB136469
	Cadmium	5.57	6.0	93	65 - 135	P	07/14/2025	14:26	LB136469
	Calcium	2030	2000	101	65 - 135	P	07/14/2025	14:26	LB136469
	Chromium	10.6	10.0	106	65 - 135	P	07/14/2025	14:26	LB136469
	Cobalt	29.7	30.0	99	65 - 135	P	07/14/2025	14:26	LB136469
	Copper	22.4	20.0	112	65 - 135	P	07/14/2025	14:26	LB136469
	Iron	114	100	114	65 - 135	P	07/14/2025	14:26	LB136469
	Lead	9.89	12.0	82	65 - 135	P	07/14/2025	14:26	LB136469
	Magnesium	2160	2000	108	65 - 135	P	07/14/2025	14:26	LB136469
	Manganese	20.7	20.0	104	65 - 135	P	07/14/2025	14:26	LB136469
	Nickel	39.2	40.0	98	65 - 135	P	07/14/2025	14:26	LB136469
	Potassium	1540	2000	77	65 - 135	P	07/14/2025	14:26	LB136469
	Selenium	20.4	20.0	102	65 - 135	P	07/14/2025	14:26	LB136469
	Silver	11.9	10.0	119	65 - 135	P	07/14/2025	14:26	LB136469
	Sodium	1800	2000	90	65 - 135	P	07/14/2025	14:26	LB136469
	Thallium	36.6	40.0	91	65 - 135	P	07/14/2025	14:26	LB136469
	Vanadium	37.6	40.0	94	65 - 135	P	07/14/2025	14:26	LB136469
	Zinc	40.6	40.0	102	65 - 135	P	07/14/2025	14:26	LB136469
CRA	Mercury	0.22	0.2	109	70 - 130	CV	07/16/2025	12:06	LB136499



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Contract: WALS01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	--------------------	---------------	---------------------------	---	------------------	------------------	---------------
