

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Contract: WALS01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	241000	255000	94	216000	294000	07/07/2025	13:53	LB136390
	Antimony	5.00			-50	50	07/07/2025	13:53	LB136390
	Arsenic	-4.19			-20	20	07/07/2025	13:53	LB136390
	Barium	-3.69	6.0	62	-94	106	07/07/2025	13:53	LB136390
	Beryllium	0.58			-6	6	07/07/2025	13:53	LB136390
	Cadmium	-0.042	1.0	4	-5	7	07/07/2025	13:53	LB136390
	Calcium	231000	245000	94	208000	282000	07/07/2025	13:53	LB136390
	Chromium	53.6	52.0	103	42	62	07/07/2025	13:53	LB136390
	Cobalt	0.58			-30	30	07/07/2025	13:53	LB136390
	Copper	10.9	2.0	544	-18	22	07/07/2025	13:53	LB136390
	Iron	101000	101000	100	85600	116500	07/07/2025	13:53	LB136390
	Lead	-5.45			-12	12	07/07/2025	13:53	LB136390
	Magnesium	244000	255000	96	216000	294000	07/07/2025	13:53	LB136390
	Manganese	6.77	7.0	97	-13	27	07/07/2025	13:53	LB136390
	Nickel	15.9	2.0	795	-38	42	07/07/2025	13:53	LB136390
	Potassium	133			0	0	07/07/2025	13:53	LB136390
	Selenium	8.20			-20	20	07/07/2025	13:53	LB136390
	Silver	-6.96			-10	10	07/07/2025	13:53	LB136390
	Sodium	101			0	0	07/07/2025	13:53	LB136390
	Thallium	-0.81			-40	40	07/07/2025	13:53	LB136390
	Vanadium	4.26			-40	40	07/07/2025	13:53	LB136390
	Zinc	3.42			-40	40	07/07/2025	13:53	LB136390
ICSAB01	Aluminum	238000	247000	96	209000	285000	07/07/2025	14:10	LB136390
	Antimony	622	618	101	525	711	07/07/2025	14:10	LB136390
	Arsenic	99.1	104	95	88.4	120	07/07/2025	14:10	LB136390
	Barium	455	537	85	437	637	07/07/2025	14:10	LB136390
	Beryllium	472	495	95	420	570	07/07/2025	14:10	LB136390
	Cadmium	992	972	102	826	1120	07/07/2025	14:10	LB136390
	Calcium	227000	235000	97	199000	271000	07/07/2025	14:10	LB136390
	Chromium	558	542	103	460	624	07/07/2025	14:10	LB136390
	Cobalt	501	476	105	404	548	07/07/2025	14:10	LB136390
	Copper	495	511	97	434	588	07/07/2025	14:10	LB136390
	Iron	103000	99300	104	84400	114500	07/07/2025	14:10	LB136390
	Lead	41.5	49.0	85	37	61	07/07/2025	14:10	LB136390
	Magnesium	238000	248000	96	210000	286000	07/07/2025	14:10	LB136390
	Manganese	473	507	93	430	584	07/07/2025	14:10	LB136390
	Nickel	1010	954	106	810	1100	07/07/2025	14:10	LB136390
	Potassium	-260			0	0	07/07/2025	14:10	LB136390
	Selenium	53.0	46.0	115	26	66	07/07/2025	14:10	LB136390
	Silver	208	201	104	170	232	07/07/2025	14:10	LB136390
	Sodium	16.7			0	0	07/07/2025	14:10	LB136390
	Thallium	78.7	108	73	68	148	07/07/2025	14:10	LB136390

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Contract: WALS01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	459	491	94	417	565	07/07/2025	14:10	LB136390
	Zinc	1050	952	110	809	1095	07/07/2025	14:10	LB136390
ICSA	Aluminum	240000	255000	94	216000	294000	07/07/2025	14:15	LB136390
	Antimony	9.95			-50	50	07/07/2025	14:15	LB136390
	Arsenic	-16.2			-20	20	07/07/2025	14:15	LB136390
	Barium	-48.6	6.0	810	-94	106	07/07/2025	14:15	LB136390
	Beryllium	-0.36			-6	6	07/07/2025	14:15	LB136390
	Cadmium	-2.96	1.0	296	-5	7	07/07/2025	14:15	LB136390
	Calcium	236000	245000	96	208000	282000	07/07/2025	14:15	LB136390
	Chromium	153	52.0	294	42	62	07/07/2025	14:15	LB136390
	Cobalt	1.25			-30	30	07/07/2025	14:15	LB136390
	Copper	24.1	2.0	1205	-18	22	07/07/2025	14:15	LB136390
	Iron	99500	101000	98	85600	116500	07/07/2025	14:15	LB136390
	Lead	0.64			-12	12	07/07/2025	14:15	LB136390
	Magnesium	244000	255000	96	216000	294000	07/07/2025	14:15	LB136390
	Manganese	11.2	7.0	160	-13	27	07/07/2025	14:15	LB136390
	Nickel	70.7	2.0	3535	-38	42	07/07/2025	14:15	LB136390
	Potassium	-7370			0	0	07/07/2025	14:15	LB136390
	Selenium	-53.4			-20	20	07/07/2025	14:15	LB136390
	Silver	-8.88			-10	10	07/07/2025	14:15	LB136390
	Sodium	-118			0	0	07/07/2025	14:15	LB136390
	Thallium	-40.5			-40	40	07/07/2025	14:15	LB136390
	Vanadium	8.77			-40	40	07/07/2025	14:15	LB136390
	Zinc	12.7			-40	40	07/07/2025	14:15	LB136390
ICSAB	Aluminum	244000	247000	99	209000	285000	07/07/2025	14:19	LB136390
	Antimony	588	618	95	525	711	07/07/2025	14:19	LB136390
	Arsenic	94.5	104	91	88.4	120	07/07/2025	14:19	LB136390
	Barium	474	537	88	437	637	07/07/2025	14:19	LB136390
	Beryllium	493	495	100	420	570	07/07/2025	14:19	LB136390
	Cadmium	970	972	100	826	1120	07/07/2025	14:19	LB136390
	Calcium	242000	235000	103	199000	271000	07/07/2025	14:19	LB136390
	Chromium	587	542	108	460	624	07/07/2025	14:19	LB136390
	Cobalt	489	476	103	404	548	07/07/2025	14:19	LB136390
	Copper	515	511	101	434	588	07/07/2025	14:19	LB136390
	Iron	102000	99300	103	84400	114500	07/07/2025	14:19	LB136390
	Lead	47.0	49.0	96	37	61	07/07/2025	14:19	LB136390
	Magnesium	250000	248000	101	210000	286000	07/07/2025	14:19	LB136390
	Manganese	490	507	97	430	584	07/07/2025	14:19	LB136390
	Nickel	1000	954	105	810	1100	07/07/2025	14:19	LB136390
	Potassium	-8010			0	0	07/07/2025	14:19	LB136390
	Selenium	19.1	46.0	42	26	66	07/07/2025	14:19	LB136390
	Silver	227	201	113	170	232	07/07/2025	14:19	LB136390

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Contract: WALS01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB	Sodium	-111			0	0	07/07/2025	14:19	LB136390
	Thallium	48.2	108	45	68	148	07/07/2025	14:19	LB136390
	Vanadium	527	491	107	417	565	07/07/2025	14:19	LB136390
	Zinc	1030	952	108	809	1095	07/07/2025	14:19	LB136390
ICSA01	Aluminum	234000	255000	92	216000	294000	07/08/2025	14:27	LB136400
	Antimony	2.92			-50	50	07/08/2025	14:27	LB136400
	Arsenic	5.77			-20	20	07/08/2025	14:27	LB136400
	Barium	-9.39	6.0	156	-94	106	07/08/2025	14:27	LB136400
	Beryllium	0.60			-6	6	07/08/2025	14:27	LB136400
	Cadmium	-0.031	1.0	3	-5	7	07/08/2025	14:27	LB136400
	Calcium	223000	245000	91	208000	282000	07/08/2025	14:27	LB136400
	Chromium	54.3	52.0	104	42	62	07/08/2025	14:27	LB136400
	Cobalt	0.62			-30	30	07/08/2025	14:27	LB136400
	Copper	5.66	2.0	283	-18	22	07/08/2025	14:27	LB136400
	Iron	101000	101000	100	85600	116500	07/08/2025	14:27	LB136400
	Lead	-6.52			-12	12	07/08/2025	14:27	LB136400
	Magnesium	234000	255000	92	216000	294000	07/08/2025	14:27	LB136400
	Manganese	7.53	7.0	108	-13	27	07/08/2025	14:27	LB136400
	Nickel	16.6	2.0	830	-38	42	07/08/2025	14:27	LB136400
	Potassium	126			0	0	07/08/2025	14:27	LB136400
	Selenium	9.28			-20	20	07/08/2025	14:27	LB136400
	Silver	-5.57			-10	10	07/08/2025	14:27	LB136400
	Sodium	51.9			0	0	07/08/2025	14:27	LB136400
	Thallium	4.20			-40	40	07/08/2025	14:27	LB136400
	Vanadium	7.70			-40	40	07/08/2025	14:27	LB136400
	Zinc	2.83			-40	40	07/08/2025	14:27	LB136400
ICSAB01	Aluminum	233000	247000	94	209000	285000	07/08/2025	14:31	LB136400
	Antimony	611	618	99	525	711	07/08/2025	14:31	LB136400
	Arsenic	102	104	98	88.4	120	07/08/2025	14:31	LB136400
	Barium	443	537	82	437	637	07/08/2025	14:31	LB136400
	Beryllium	456	495	92	420	570	07/08/2025	14:31	LB136400
	Cadmium	963	972	99	826	1120	07/08/2025	14:31	LB136400
	Calcium	223000	235000	95	199000	271000	07/08/2025	14:31	LB136400
	Chromium	543	542	100	460	624	07/08/2025	14:31	LB136400
	Cobalt	489	476	103	404	548	07/08/2025	14:31	LB136400
	Copper	474	511	93	434	588	07/08/2025	14:31	LB136400
	Iron	101000	99300	102	84400	114500	07/08/2025	14:31	LB136400
	Lead	40.0	49.0	82	37	61	07/08/2025	14:31	LB136400
	Magnesium	233000	248000	94	210000	286000	07/08/2025	14:31	LB136400
	Manganese	464	507	92	430	584	07/08/2025	14:31	LB136400
	Nickel	978	954	102	810	1100	07/08/2025	14:31	LB136400
	Potassium	86.8			0	0	07/08/2025	14:31	LB136400

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Contract: WALS01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Selenium	56.6	46.0	123	26	66	07/08/2025	14:31	LB136400
	Silver	206	201	102	170	232	07/08/2025	14:31	LB136400
	Sodium	27.5			0	0	07/08/2025	14:31	LB136400
	Thallium	83.9	108	78	68	148	07/08/2025	14:31	LB136400
	Vanadium	454	491	92	417	565	07/08/2025	14:31	LB136400
	Zinc	1030	952	108	809	1095	07/08/2025	14:31	LB136400
ICSA	Aluminum	234000	255000	92	216000	294000	07/08/2025	14:35	LB136400
	Antimony	-26.7			-50	50	07/08/2025	14:35	LB136400
	Arsenic	-19.2			-20	20	07/08/2025	14:35	LB136400
	Barium	-159	6.0	2650	-94	106	07/08/2025	14:35	LB136400
	Beryllium	1.31			-6	6	07/08/2025	14:35	LB136400
	Cadmium	-0.58	1.0	58	-5	7	07/08/2025	14:35	LB136400
	Calcium	232000	245000	95	208000	282000	07/08/2025	14:35	LB136400
	Chromium	145	52.0	279	42	62	07/08/2025	14:35	LB136400
	Cobalt	0.80			-30	30	07/08/2025	14:35	LB136400
	Copper	20.3	2.0	1015	-18	22	07/08/2025	14:35	LB136400
	Iron	98400	101000	97	85600	116500	07/08/2025	14:35	LB136400
	Lead	-7.42			-12	12	07/08/2025	14:35	LB136400
	Magnesium	240000	255000	94	216000	294000	07/08/2025	14:35	LB136400
	Manganese	15.8	7.0	226	-13	27	07/08/2025	14:35	LB136400
	Nickel	73.7	2.0	3685	-38	42	07/08/2025	14:35	LB136400
	Potassium	-699			0	0	07/08/2025	14:35	LB136400
	Selenium	67.4			-20	20	07/08/2025	14:35	LB136400
	Silver	-8.36			-10	10	07/08/2025	14:35	LB136400
	Sodium	53.5			0	0	07/08/2025	14:35	LB136400
	Thallium	-7.84			-40	40	07/08/2025	14:35	LB136400
	Vanadium	19.3			-40	40	07/08/2025	14:35	LB136400
	Zinc	13.1			-40	40	07/08/2025	14:35	LB136400
ICSAB	Aluminum	237000	247000	96	209000	285000	07/08/2025	14:39	LB136400
	Antimony	611	618	99	525	711	07/08/2025	14:39	LB136400
	Arsenic	75.0	104	72	88.4	120	07/08/2025	14:39	LB136400
	Barium	302	537	56	437	637	07/08/2025	14:39	LB136400
	Beryllium	474	495	96	420	570	07/08/2025	14:39	LB136400
	Cadmium	954	972	98	826	1120	07/08/2025	14:39	LB136400
	Calcium	235000	235000	100	199000	271000	07/08/2025	14:39	LB136400
	Chromium	569	542	105	460	624	07/08/2025	14:39	LB136400
	Cobalt	482	476	101	404	548	07/08/2025	14:39	LB136400
	Copper	518	511	101	434	588	07/08/2025	14:39	LB136400
	Iron	98100	99300	99	84400	114500	07/08/2025	14:39	LB136400
	Lead	35.6	49.0	73	37	61	07/08/2025	14:39	LB136400
	Magnesium	242000	248000	98	210000	286000	07/08/2025	14:39	LB136400
	Manganese	483	507	95	430	584	07/08/2025	14:39	LB136400

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Contract: WALS01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB	Nickel	988	954	104	810	1100	07/08/2025	14:39	LB136400
	Potassium	-1170			0	0	07/08/2025	14:39	LB136400
	Selenium	97.6	46.0	212	26	66	07/08/2025	14:39	LB136400
	Silver	211	201	105	170	232	07/08/2025	14:39	LB136400
	Sodium	-11.1			0	0	07/08/2025	14:39	LB136400
	Thallium	53.8	108	50	68	148	07/08/2025	14:39	LB136400
	Vanadium	456	491	93	417	565	07/08/2025	14:39	LB136400
	Zinc	1000	952	105	809	1095	07/08/2025	14:39	LB136400
ICSA01	Aluminum	245000	250000	98	216000	294000	07/09/2025	10:31	LB136407
	Antimony	23.3			-50	50	07/09/2025	10:31	LB136407
	Arsenic	5.46			-20	20	07/09/2025	10:31	LB136407
	Barium	8.02	6.0	134	-94	106	07/09/2025	10:31	LB136407
	Beryllium	1.50			-6	6	07/09/2025	10:31	LB136407
	Cadmium	-0.43	1.0	43	-5	7	07/09/2025	10:31	LB136407
	Calcium	242000	240000	101	208000	282000	07/09/2025	10:31	LB136407
	Chromium	52.2	52.0	100	42	62	07/09/2025	10:31	LB136407
	Cobalt	27.0			-30	30	07/09/2025	10:31	LB136407
	Copper	-4.74	2.0	237	-18	22	07/09/2025	10:31	LB136407
	Iron	96800	100000	97	85600	116500	07/09/2025	10:31	LB136407
	Lead	-5.64			-12	12	07/09/2025	10:31	LB136407
	Magnesium	242000	260000	93	216000	294000	07/09/2025	10:31	LB136407
	Manganese	3.37	7.0	48	-13	27	07/09/2025	10:31	LB136407
	Nickel	13.5	2.0	675	-38	42	07/09/2025	10:31	LB136407
	Potassium	10.6			0	0	07/09/2025	10:31	LB136407
	Selenium	-11.2			-20	20	07/09/2025	10:31	LB136407
	Silver	-2.41			-10	10	07/09/2025	10:31	LB136407
	Sodium	119			0	0	07/09/2025	10:31	LB136407
	Thallium	16.7			-40	40	07/09/2025	10:31	LB136407
	Vanadium	1.75			-40	40	07/09/2025	10:31	LB136407
	Zinc	10.9			-40	40	07/09/2025	10:31	LB136407
ICSAB01	Aluminum	245000	250000	98	209000	285000	07/09/2025	10:35	LB136407
	Antimony	621	620	100	525	711	07/09/2025	10:35	LB136407
	Arsenic	98.6	100	99	88.4	120	07/09/2025	10:35	LB136407
	Barium	529	540	98	437	637	07/09/2025	10:35	LB136407
	Beryllium	514	500	103	420	570	07/09/2025	10:35	LB136407
	Cadmium	984	970	101	826	1120	07/09/2025	10:35	LB136407
	Calcium	241000	230000	105	199000	271000	07/09/2025	10:35	LB136407
	Chromium	555	540	103	460	624	07/09/2025	10:35	LB136407
	Cobalt	506	480	105	404	548	07/09/2025	10:35	LB136407
	Copper	479	510	94	434	588	07/09/2025	10:35	LB136407
	Iron	98200	99000	99	84400	114500	07/09/2025	10:35	LB136407
	Lead	38.6	49.0	79	37	61	07/09/2025	10:35	LB136407

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Contract: WALS01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Magnesium	243000	250000	97	210000	286000	07/09/2025	10:35	LB136407
	Manganese	498	510	98	430	584	07/09/2025	10:35	LB136407
	Nickel	966	950	102	810	1100	07/09/2025	10:35	LB136407
	Potassium	9.40			0	0	07/09/2025	10:35	LB136407
	Selenium	40.2	46.0	87	26	66	07/09/2025	10:35	LB136407
	Silver	209	200	104	170	232	07/09/2025	10:35	LB136407
	Sodium	100			0	0	07/09/2025	10:35	LB136407
	Thallium	101	110	92	68	148	07/09/2025	10:35	LB136407
	Vanadium	489	490	100	417	565	07/09/2025	10:35	LB136407
	Zinc	1030	950	108	809	1095	07/09/2025	10:35	LB136407
ICSA	Aluminum	256000	250000	102	216000	294000	07/09/2025	10:39	LB136407
	Antimony	56.4			-50	50	07/09/2025	10:39	LB136407
	Arsenic	-38.5			-20	20	07/09/2025	10:39	LB136407
	Barium	3.35	6.0	56	-94	106	07/09/2025	10:39	LB136407
	Beryllium	3.59			-6	6	07/09/2025	10:39	LB136407
	Cadmium	-4.72	1.0	472	-5	7	07/09/2025	10:39	LB136407
	Calcium	253000	240000	105	208000	282000	07/09/2025	10:39	LB136407
	Chromium	30.5	52.0	59	42	62	07/09/2025	10:39	LB136407
	Cobalt	120			-30	30	07/09/2025	10:39	LB136407
	Copper	39.8	2.0	1990	-18	22	07/09/2025	10:39	LB136407
	Iron	108000	100000	108	85600	116500	07/09/2025	10:39	LB136407
	Lead	24.1			-12	12	07/09/2025	10:39	LB136407
	Magnesium	259000	260000	100	216000	294000	07/09/2025	10:39	LB136407
	Manganese	25.6	7.0	366	-13	27	07/09/2025	10:39	LB136407
	Nickel	-4.00	2.0	200	-38	42	07/09/2025	10:39	LB136407
	Potassium	59.1			0	0	07/09/2025	10:39	LB136407
	Selenium	91.3			-20	20	07/09/2025	10:39	LB136407
	Silver	-6.28			-10	10	07/09/2025	10:39	LB136407
	Sodium	-123			0	0	07/09/2025	10:39	LB136407
	Thallium	-21.0			-40	40	07/09/2025	10:39	LB136407
	Vanadium	-159			-40	40	07/09/2025	10:39	LB136407
	Zinc	21.5			-40	40	07/09/2025	10:39	LB136407
ICSAB	Aluminum	239000	250000	96	209000	285000	07/09/2025	10:44	LB136407
	Antimony	588	620	95	525	711	07/09/2025	10:44	LB136407
	Arsenic	95.9	100	96	88.4	120	07/09/2025	10:44	LB136407
	Barium	555	540	103	437	637	07/09/2025	10:44	LB136407
	Beryllium	497	500	99	420	570	07/09/2025	10:44	LB136407
	Cadmium	885	970	91	826	1120	07/09/2025	10:44	LB136407
	Calcium	236000	230000	103	199000	271000	07/09/2025	10:44	LB136407
	Chromium	527	540	98	460	624	07/09/2025	10:44	LB136407
	Cobalt	563	480	117	404	548	07/09/2025	10:44	LB136407
	Copper	517	510	101	434	588	07/09/2025	10:44	LB136407

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Contract: WALS01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB	Iron	101000	99000	102	84400	114500	07/09/2025	10:44	LB136407
	Lead	15.6	49.0	32	37	61	07/09/2025	10:44	LB136407
	Magnesium	245000	250000	98	210000	286000	07/09/2025	10:44	LB136407
	Manganese	558	510	109	430	584	07/09/2025	10:44	LB136407
	Nickel	883	950	93	810	1100	07/09/2025	10:44	LB136407
	Potassium	-58.8			0	0	07/09/2025	10:44	LB136407
	Selenium	-15.1	46.0	33	26	66	07/09/2025	10:44	LB136407
	Silver	181	200	90	170	232	07/09/2025	10:44	LB136407
	Sodium	-8.12			0	0	07/09/2025	10:44	LB136407
	Thallium	109	110	99	68	148	07/09/2025	10:44	LB136407
	Vanadium	305	490	62	417	565	07/09/2025	10:44	LB136407
	Zinc	991	950	104	809	1095	07/09/2025	10:44	LB136407
ICSA01	Aluminum	247000	250000	99	216000	294000	07/10/2025	19:11	LB136434
	Antimony	21.5			-50	50	07/10/2025	19:11	LB136434
	Arsenic	2.10			-20	20	07/10/2025	19:11	LB136434
	Barium	5.81	6.0	97	-94	106	07/10/2025	19:11	LB136434
	Beryllium	1.15			-6	6	07/10/2025	19:11	LB136434
	Cadmium	-0.49	1.0	49	-5	7	07/10/2025	19:11	LB136434
	Calcium	242000	240000	101	208000	282000	07/10/2025	19:11	LB136434
	Chromium	48.9	52.0	94	42	62	07/10/2025	19:11	LB136434
	Cobalt	25.0			-30	30	07/10/2025	19:11	LB136434
	Copper	-9.61	2.0	480	-18	22	07/10/2025	19:11	LB136434
	Iron	104000	100000	104	85600	116500	07/10/2025	19:11	LB136434
	Lead	0.44			-12	12	07/10/2025	19:11	LB136434
	Magnesium	252000	260000	97	216000	294000	07/10/2025	19:11	LB136434
	Manganese	2.89	7.0	41	-13	27	07/10/2025	19:11	LB136434
	Nickel	16.4	2.0	820	-38	42	07/10/2025	19:11	LB136434
	Potassium	5.28			0	0	07/10/2025	19:11	LB136434
	Selenium	4.66			-20	20	07/10/2025	19:11	LB136434
	Silver	-1.13			-10	10	07/10/2025	19:11	LB136434
	Sodium	100			0	0	07/10/2025	19:11	LB136434
	Thallium	17.8			-40	40	07/10/2025	19:11	LB136434
	Vanadium	4.92			-40	40	07/10/2025	19:11	LB136434
	Zinc	11.0			-40	40	07/10/2025	19:11	LB136434
ICSAB01	Aluminum	244000	250000	98	209000	285000	07/10/2025	19:15	LB136434
	Antimony	612	620	99	525	711	07/10/2025	19:15	LB136434
	Arsenic	92.1	100	92	88.4	120	07/10/2025	19:15	LB136434
	Barium	505	540	94	437	637	07/10/2025	19:15	LB136434
	Beryllium	505	500	101	420	570	07/10/2025	19:15	LB136434
	Cadmium	972	970	100	826	1120	07/10/2025	19:15	LB136434
	Calcium	239000	230000	104	199000	271000	07/10/2025	19:15	LB136434
	Chromium	548	540	102	460	624	07/10/2025	19:15	LB136434

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Contract: WALS01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Cobalt	497	480	104	404	548	07/10/2025	19:15	LB136434
	Copper	468	510	92	434	588	07/10/2025	19:15	LB136434
	Iron	98700	99000	100	84400	114500	07/10/2025	19:15	LB136434
	Lead	43.4	49.0	89	37	61	07/10/2025	19:15	LB136434
	Magnesium	249000	250000	100	210000	286000	07/10/2025	19:15	LB136434
	Manganese	491	510	96	430	584	07/10/2025	19:15	LB136434
	Nickel	971	950	102	810	1100	07/10/2025	19:15	LB136434
	Potassium	0.96			0	0	07/10/2025	19:15	LB136434
	Selenium	52.3	46.0	114	26	66	07/10/2025	19:15	LB136434
	Silver	217	200	108	170	232	07/10/2025	19:15	LB136434
	Sodium	88.1			0	0	07/10/2025	19:15	LB136434
	Thallium	99.2	110	90	68	148	07/10/2025	19:15	LB136434
	Vanadium	488	490	100	417	565	07/10/2025	19:15	LB136434
	Zinc	983	950	104	809	1095	07/10/2025	19:15	LB136434
ICSA	Aluminum	243000	250000	97	216000	294000	07/10/2025	19:20	LB136434
	Antimony	-0.94			-50	50	07/10/2025	19:20	LB136434
	Arsenic	-11.1			-20	20	07/10/2025	19:20	LB136434
	Barium	-44.9	6.0	748	-94	106	07/10/2025	19:20	LB136434
	Beryllium	3.62			-6	6	07/10/2025	19:20	LB136434
	Cadmium	-3.58	1.0	358	-5	7	07/10/2025	19:20	LB136434
	Calcium	243000	240000	101	208000	282000	07/10/2025	19:20	LB136434
	Chromium	46.3	52.0	89	42	62	07/10/2025	19:20	LB136434
	Cobalt	105			-30	30	07/10/2025	19:20	LB136434
	Copper	-19.0	2.0	950	-18	22	07/10/2025	19:20	LB136434
	Iron	112000	100000	112	85600	116500	07/10/2025	19:20	LB136434
	Lead	41.5			-12	12	07/10/2025	19:20	LB136434
	Magnesium	257000	260000	99	216000	294000	07/10/2025	19:20	LB136434
	Manganese	29.5	7.0	421	-13	27	07/10/2025	19:20	LB136434
	Nickel	4.08	2.0	204	-38	42	07/10/2025	19:20	LB136434
	Potassium	-56.8			0	0	07/10/2025	19:20	LB136434
	Selenium	-5.16			-20	20	07/10/2025	19:20	LB136434
	Silver	-23.6			-10	10	07/10/2025	19:20	LB136434
	Sodium	-285			0	0	07/10/2025	19:20	LB136434
	Thallium	30.2			-40	40	07/10/2025	19:20	LB136434
	Vanadium	-108			-40	40	07/10/2025	19:20	LB136434
	Zinc	-12.9			-40	40	07/10/2025	19:20	LB136434
ICSAB	Aluminum	248000	250000	99	209000	285000	07/10/2025	19:24	LB136434
	Antimony	729	620	118	525	711	07/10/2025	19:24	LB136434
	Arsenic	19.2	100	19	88.4	120	07/10/2025	19:24	LB136434
	Barium	460	540	85	437	637	07/10/2025	19:24	LB136434
	Beryllium	531	500	106	420	570	07/10/2025	19:24	LB136434
	Cadmium	925	970	95	826	1120	07/10/2025	19:24	LB136434

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Contract: WALS01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB	Calcium	248000	230000	108	199000	271000	07/10/2025	19:24	LB136434
	Chromium	554	540	103	460	624	07/10/2025	19:24	LB136434
	Cobalt	565	480	118	404	548	07/10/2025	19:24	LB136434
	Copper	475	510	93	434	588	07/10/2025	19:24	LB136434
	Iron	111000	99000	112	84400	114500	07/10/2025	19:24	LB136434
	Lead	109	49.0	222	37	61	07/10/2025	19:24	LB136434
	Magnesium	263000	250000	105	210000	286000	07/10/2025	19:24	LB136434
	Manganese	531	510	104	430	584	07/10/2025	19:24	LB136434
	Nickel	933	950	98	810	1100	07/10/2025	19:24	LB136434
	Potassium	-72.5			0	0	07/10/2025	19:24	LB136434
	Selenium	17.3	46.0	38	26	66	07/10/2025	19:24	LB136434
	Silver	207	200	104	170	232	07/10/2025	19:24	LB136434
	Sodium	-448			0	0	07/10/2025	19:24	LB136434
	Thallium	131	110	119	68	148	07/10/2025	19:24	LB136434
	Vanadium	444	490	91	417	565	07/10/2025	19:24	LB136434
	Zinc	975	950	103	809	1095	07/10/2025	19:24	LB136434
ICSA01	Aluminum	258000	250000	103	216000	294000	07/14/2025	14:37	LB136469
	Antimony	20.8			-50	50	07/14/2025	14:37	LB136469
	Arsenic	2.40			-20	20	07/14/2025	14:37	LB136469
	Barium	6.88	6.0	115	-94	106	07/14/2025	14:37	LB136469
	Beryllium	1.18			-6	6	07/14/2025	14:37	LB136469
	Cadmium	-0.081	1.0	8	-5	7	07/14/2025	14:37	LB136469
	Calcium	248000	240000	103	208000	282000	07/14/2025	14:37	LB136469
	Chromium	53.4	52.0	103	42	62	07/14/2025	14:37	LB136469
	Cobalt	-17.9			-30	30	07/14/2025	14:37	LB136469
	Copper	-6.35	2.0	318	-18	22	07/14/2025	14:37	LB136469
	Iron	99000	100000	99	85600	116500	07/14/2025	14:37	LB136469
	Lead	4.49			-12	12	07/14/2025	14:37	LB136469
	Magnesium	256000	260000	98	216000	294000	07/14/2025	14:37	LB136469
	Manganese	7.27	7.0	104	-13	27	07/14/2025	14:37	LB136469
	Nickel	17.4	2.0	870	-38	42	07/14/2025	14:37	LB136469
	Potassium	6.93			0	0	07/14/2025	14:37	LB136469
	Selenium	-9.51			-20	20	07/14/2025	14:37	LB136469
	Silver	4.04			-10	10	07/14/2025	14:37	LB136469
	Sodium	99.1			0	0	07/14/2025	14:37	LB136469
	Thallium	21.9			-40	40	07/14/2025	14:37	LB136469
	Vanadium	0.58			-40	40	07/14/2025	14:37	LB136469
	Zinc	11.6			-40	40	07/14/2025	14:37	LB136469
ICSAB01	Aluminum	248000	250000	99	209000	285000	07/14/2025	14:41	LB136469
	Antimony	626	620	101	525	711	07/14/2025	14:41	LB136469
	Arsenic	96.3	100	96	88.4	120	07/14/2025	14:41	LB136469
	Barium	508	540	94	437	637	07/14/2025	14:41	LB136469

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Contract: WALS01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Beryllium	515	500	103	420	570	07/14/2025	14:41	LB136469
	Cadmium	996	970	103	826	1120	07/14/2025	14:41	LB136469
	Calcium	241000	230000	105	199000	271000	07/14/2025	14:41	LB136469
	Chromium	555	540	103	460	624	07/14/2025	14:41	LB136469
	Cobalt	473	480	98	404	548	07/14/2025	14:41	LB136469
	Copper	485	510	95	434	588	07/14/2025	14:41	LB136469
	Iron	97200	99000	98	84400	114500	07/14/2025	14:41	LB136469
	Lead	46.0	49.0	94	37	61	07/14/2025	14:41	LB136469
	Magnesium	246000	250000	98	210000	286000	07/14/2025	14:41	LB136469
	Manganese	492	510	96	430	584	07/14/2025	14:41	LB136469
	Nickel	998	950	105	810	1100	07/14/2025	14:41	LB136469
	Potassium	6.60			0	0	07/14/2025	14:41	LB136469
	Selenium	35.1	46.0	76	26	66	07/14/2025	14:41	LB136469
	Silver	217	200	108	170	232	07/14/2025	14:41	LB136469
	Sodium	118			0	0	07/14/2025	14:41	LB136469
	Thallium	104	110	94	68	148	07/14/2025	14:41	LB136469
	Vanadium	483	490	99	417	565	07/14/2025	14:41	LB136469
	Zinc	1010	950	106	809	1095	07/14/2025	14:41	LB136469
ICSA	Aluminum	248000	250000	99	216000	294000	07/14/2025	14:46	LB136469
	Antimony	-0.92			-50	50	07/14/2025	14:46	LB136469
	Arsenic	-91.2			-20	20	07/14/2025	14:46	LB136469
	Barium	15.9	6.0	265	-94	106	07/14/2025	14:46	LB136469
	Beryllium	1.94			-6	6	07/14/2025	14:46	LB136469
	Cadmium	-2.48	1.0	248	-5	7	07/14/2025	14:46	LB136469
	Calcium	246000	240000	102	208000	282000	07/14/2025	14:46	LB136469
	Chromium	58.2	52.0	112	42	62	07/14/2025	14:46	LB136469
	Cobalt	95.0			-30	30	07/14/2025	14:46	LB136469
	Copper	-3.84	2.0	192	-18	22	07/14/2025	14:46	LB136469
	Iron	107000	100000	107	85600	116500	07/14/2025	14:46	LB136469
	Lead	10.7			-12	12	07/14/2025	14:46	LB136469
	Magnesium	258000	260000	99	216000	294000	07/14/2025	14:46	LB136469
	Manganese	-1.10	7.0	16	-13	27	07/14/2025	14:46	LB136469
	Nickel	37.9	2.0	1895	-38	42	07/14/2025	14:46	LB136469
	Potassium	-3.54			0	0	07/14/2025	14:46	LB136469
	Selenium	-111			-20	20	07/14/2025	14:46	LB136469
	Silver	19.6			-10	10	07/14/2025	14:46	LB136469
	Sodium	-141			0	0	07/14/2025	14:46	LB136469
	Thallium	89.6			-40	40	07/14/2025	14:46	LB136469
	Vanadium	-33.3			-40	40	07/14/2025	14:46	LB136469
	Zinc	12.3			-40	40	07/14/2025	14:46	LB136469
ICSAB	Aluminum	250000	250000	100	209000	285000	07/14/2025	14:50	LB136469
	Antimony	584	620	94	525	711	07/14/2025	14:50	LB136469

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC

SDG No.: Q2487

Contract: WALS01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB	Arsenic	-5.02	100	5	88.4	120	07/14/2025	14:50	LB136469
	Barium	482	540	89	437	637	07/14/2025	14:50	LB136469
	Beryllium	524	500	105	420	570	07/14/2025	14:50	LB136469
	Cadmium	940	970	97	826	1120	07/14/2025	14:50	LB136469
	Calcium	244000	230000	106	199000	271000	07/14/2025	14:50	LB136469
	Chromium	567	540	105	460	624	07/14/2025	14:50	LB136469
	Cobalt	572	480	119	404	548	07/14/2025	14:50	LB136469
	Copper	521	510	102	434	588	07/14/2025	14:50	LB136469
	Iron	107000	99000	108	84400	114500	07/14/2025	14:50	LB136469
	Lead	46.2	49.0	94	37	61	07/14/2025	14:50	LB136469
	Magnesium	256000	250000	102	210000	286000	07/14/2025	14:50	LB136469
	Manganese	512	510	100	430	584	07/14/2025	14:50	LB136469
	Nickel	975	950	103	810	1100	07/14/2025	14:50	LB136469
	Potassium	-6.48			0	0	07/14/2025	14:50	LB136469
	Selenium	-65.3	46.0	142	26	66	07/14/2025	14:50	LB136469
	Silver	208	200	104	170	232	07/14/2025	14:50	LB136469
	Sodium	133			0	0	07/14/2025	14:50	LB136469
	Thallium	147	110	134	68	148	07/14/2025	14:50	LB136469
	Vanadium	348	490	71	417	565	07/14/2025	14:50	LB136469
	Zinc	1000	950	105	809	1095	07/14/2025	14:50	LB136469