284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LABORATORY CONTROL SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC **SDG No.:** Q2487

Contract: WALS01 **Lab Code:** ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168714BS							
Aluminum	mg/Kg	100	95.7		96	80 - 120	P
Antimony	mg/Kg	40.0	40.7		102	80 - 120	P
Arsenic	mg/Kg	40.0	39.5		99	80 - 120	P
Barium	mg/Kg	10.0	10.3		103	80 - 120	P
Beryllium	mg/Kg	10.0	9.46		95	80 - 120	P
Cadmium	mg/Kg	10.0	9.84		98	80 - 120	P
Calcium	mg/Kg	50.0	50.0	J	100	80 - 120	P
Chromium	mg/Kg	20.0	20.1		100	80 - 120	P
Cobalt	mg/Kg	10.0	9.99		100	80 - 120	P
Copper	mg/Kg	15.0	15.5		103	80 - 120	P
Iron	mg/Kg	150	152		101	80 - 120	P
Lead	mg/Kg	50.0	48.3		97	80 - 120	P
Magnesium	mg/Kg	100	94.5	J	94	80 - 120	P
Manganese	mg/Kg	10.0	9.68		97	80 - 120	P
Nickel	mg/Kg	25.0	25.2		101	80 - 120	P
Potassium	mg/Kg	500	501		100	80 - 120	P
Selenium	mg/Kg	100	98.0		98	80 - 120	P
Silver	mg/Kg	3.8	3.58		94	80 - 120	P
Sodium	mg/Kg	150	158		105	80 - 120	P
Thallium	mg/Kg	100	82.6		83	80 - 120	P
Vanadium	mg/Kg	15.0	14.1		94	80 - 120	P
Zinc	mg/Kg	10.0	10.3		103	80 - 120	P

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LABORATORY CONTROL SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC **SDG No.:** Q2487

Contract: WALS01 **Lab Code:** ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168892BS							
Mercury	mg/Kg	0.25	0.24		98	80 - 120	CV