

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Contract: WALS01
 ProjectID: Construction of Shafts 17B-18B - PN 220084
 Lab Code: ACE SDG No.: Q2487

Calibration Sequence : FG061325		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	121978597	121979	FG016063.D
500	61365081	122730	FG016064.D
200	26234834	131174	FG016065.D
100	13396741	133967	FG016066.D
50	7145066	142901	FG016067.D
AVG RF : 130550		% RSD : 6.625	AVG RT : 15.1972