

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Contract: WALS01
ProjectID: Construction of Shafts 17B-18B - PN 220084
Lab Code: ACE SDG No.: Q2487

Calibration Sequence : FG070925		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	107457449	107457	FG016264.D
500	53927299	107855	FG016265.D
200	23004292	115021	FG016266.D
100	12319443	123194	FG016267.D
50	6076250	121525	FG016268.D
AVG RF : 115010		% RSD : 6.414	AVG RT : 15.201