

DIESEL RANGE ORGANICS CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Alliane Contract: WALS01
ProjectID: Construction of Shafts 17B-18B - PN 220
Lab Code: ACE SDG No.: Q2487
DataFile: FG016245.D Analyst Name: YP\AJ Analyst Date: 07-08-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
500	61154300	122309	130550	6.313