

Continuing Calibration Report for SequenceID : FC070725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069375.D

Aliphatic C9-C12	8408520.000	60.000	3.302	6.602	140142.000	146679.627	4.457
Aliphatic C12-C16	5794746.000	40.000	6.603	10.006	144868.650	159210.532	9.008
Aliphatic C16-C21	8151741.000	60.000	10.007	13.376	135862.350	154434.323	12.026
Aliphatic C21-C28	9574432.000	80.000	13.377	17.041	119680.400	133966.368	10.664
Aliphatic C28-C40	10859754.000	120.000	17.042	22.022	90497.950	92124.417	1.766
Aliphatic EPH	42789193.000	360.000	3.302	22.022	118858.869	128354.160	7.398