

Continuing Calibration Report for SequenceID : FC070725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069385.D

Aliphatic C9-C12	8356333.000	60.000	3.302	6.602	139272.217	146679.627	5.050
Aliphatic C12-C16	5755442.000	40.000	6.603	10.006	143886.050	159210.532	9.625
Aliphatic C16-C21	8081610.000	60.000	10.007	13.376	134693.500	154434.323	12.783
Aliphatic C21-C28	9567952.000	80.000	13.377	17.041	119599.400	133966.368	10.724
Aliphatic C28-C40	11143762.000	120.000	17.042	22.022	92864.683	92124.417	-0.804
Aliphatic EPH	42905099.000	360.000	3.302	22.022	119180.831	128354.160	7.147