

Continuing Calibration Report for SequenceID : FE070725AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FE054704.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 8370943.000  | 60.000  | 3.322  | 6.953  | 139515.717 | 136055.079 | -2.544 |
| Aliphatic C12-C16 | 5771729.000  | 40.000  | 6.954  | 10.404 | 144293.225 | 140568.128 | -2.650 |
| Aliphatic C16-C21 | 8399108.000  | 60.000  | 10.405 | 13.782 | 139985.133 | 144393.643 | 3.053  |
| Aliphatic C21-C28 | 11772203.000 | 80.000  | 13.783 | 17.452 | 147152.538 | 144651.717 | -1.729 |
| Aliphatic C28-C40 | 15996736.000 | 120.000 | 17.453 | 22.469 | 133306.133 | 138647.492 | 3.852  |
| Aliphatic EPH     | 50310719.000 | 360.000 | 3.322  | 22.469 | 139751.997 | 140720.791 | 0.688  |