

Continuing Calibration Report for SequenceID : FE070825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE054727.D

Aliphatic C9-C12	8667506.000	60.000	3.323	6.956	144458.433	136055.079	-6.176
Aliphatic C12-C16	6020302.000	40.000	6.957	10.409	150507.550	140568.128	-7.071
Aliphatic C16-C21	8748005.000	60.000	10.410	13.786	145800.083	144393.643	-0.974
Aliphatic C21-C28	12264455.000	80.000	13.787	17.458	153305.688	144651.717	-5.983
Aliphatic C28-C40	17283781.000	120.000	17.459	22.482	144031.508	138647.492	-3.883
Aliphatic EPH	52984049.000	360.000	3.323	22.482	147177.914	140720.791	-4.589