

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB062325

Review By	yogesh	Review On	6/23/2025 3:07:11 PM
Supervise By	mohammad	Supervise On	6/25/2025 2:41:46 AM
SubDirectory	FB062325	HP Acquire Method	HP Processing Method FB062325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24656,PP24658,PP24659,PP24660,PP24661		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24662		

Sr#	SampleID	Data File Name	Date-Time	Operator	Status
1	5 GRO STD	FB031909.D	23 Jun 2025 12:17	YP/AJ	Ok
2	10 GRO STD	FB031910.D	23 Jun 2025 12:44	YP/AJ	Ok
3	20 GRO STD	FB031911.D	23 Jun 2025 13:12	YP/AJ	Ok
4	50 GRO STD	FB031912.D	23 Jun 2025 13:40	YP/AJ	Ok
5	100 GRO STD	FB031913.D	23 Jun 2025 14:07	YP/AJ	Ok
6	FB062325GROICV	FB031914.D	23 Jun 2025 14:48	YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB070925

Review By	yogesh	Review On	7/9/2025 3:32:13 PM
Supervise By	mohammad	Supervise On	7/11/2025 1:36:50 AM
SubDirectory	FB070925	HP Acquire Method	HP Processing Method FB062325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24656,PP24658,PP24659,PP24660,PP24661		
CCC Internal Standard/PEM	PP24705,PP24706,PP24707		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24662		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	20 PPB GRO STD	FB032033.D	9 Jul 2025 8:12	YP/AJ	Ok
2	VBF0709S1	FB032034.D	9 Jul 2025 9:18	YP/AJ	Ok
3	VBF0709S2	FB032035.D	9 Jul 2025 9:45	YP/AJ	Ok
4	BSF0709S1	FB032036.D	9 Jul 2025 10:13	YP/AJ	Ok
5	Q2487-09	FB032037.D	9 Jul 2025 10:58	YP/AJ	Ok
6	Q2487-10	FB032038.D	9 Jul 2025 11:26	YP/AJ	Ok
7	Q2487-11	FB032039.D	9 Jul 2025 11:53	YP/AJ	Ok
8	Q2487-12	FB032040.D	9 Jul 2025 12:20	YP/AJ	Ok
9	Q2487-13	FB032041.D	9 Jul 2025 12:48	YP/AJ	Not Ok
10	Q2487-14	FB032042.D	9 Jul 2025 13:15	YP/AJ	Not Ok
11	BSF0709S2	FB032043.D	9 Jul 2025 13:42	YP/AJ	Ok
12	20 PPB GRO STD	FB032044.D	9 Jul 2025 15:30	YP/AJ	Ok,M
13	Q2514-02	FB032045.D	9 Jul 2025 16:15	YP/AJ	Ok
14	Q2514-01	FB032046.D	9 Jul 2025 16:42	YP/AJ	Ok
15	Q2514-08	FB032047.D	9 Jul 2025 17:10	YP/AJ	Ok
16	Q2514-07	FB032048.D	9 Jul 2025 17:37	YP/AJ	Ok
17	Q2514-03	FB032049.D	9 Jul 2025 18:05	YP/AJ	Ok
18	Q2514-04	FB032050.D	9 Jul 2025 18:32	YP/AJ	Ok
19	Q2514-05	FB032051.D	9 Jul 2025 19:00	YP/AJ	Ok
20	Q2514-06	FB032052.D	9 Jul 2025 19:27	YP/AJ	Ok,M
21	Q2514-10	FB032053.D	9 Jul 2025 19:54	YP/AJ	Ok

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB070925

Review By	yogesh	Review On	7/9/2025 3:32:13 PM
Supervise By	mohammad	Supervise On	7/11/2025 1:36:50 AM
SubDirectory	FB070925	HP Acquire Method	HP Processing Method FB062325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24656,PP24658,PP24659,PP24660,PP24661		
CCC Internal Standard/PEM	PP24705,PP24706,PP24707		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24662		

22	Q2514-09	FB032054.D	9 Jul 2025 20:21	YP/AJ	Ok
23	Q2487-13	FB032055.D	9 Jul 2025 20:49	YP/AJ	Ok
24	Q2487-14	FB032056.D	9 Jul 2025 21:16	YP/AJ	Ok
25	BSF0709S3	FB032057.D	9 Jul 2025 21:43	YP/AJ	Ok
26	20 PPB GRO STD	FB032058.D	9 Jul 2025 22:10	YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB071025

Review By	yogesh	Review On	7/10/2025 3:17:40 PM
Supervise By	mohammad	Supervise On	7/11/2025 9:10:17 AM
SubDirectory	FB071025	HP Acquire Method	HP Processing Method FB062325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24656,PP24658,PP24659,PP24660,PP24661		
CCC Internal Standard/PEM	PP24710,PP24711,PP24712		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24662		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	20 PPB GRO STD	FB032059.D	10 Jul 2025 9:08	YP/AJ	Ok,M
2	VBF0710S1	FB032060.D	10 Jul 2025 9:50	YP/AJ	Ok
3	VBF0710S2	FB032061.D	10 Jul 2025 10:18	YP/AJ	Ok
4	BSF0710S1	FB032062.D	10 Jul 2025 10:45	YP/AJ	Ok
5	Q2487-15	FB032063.D	10 Jul 2025 11:25	YP/AJ	Not Ok
6	Q2487-16	FB032064.D	10 Jul 2025 11:53	YP/AJ	Not Ok
7	Q2543-01	FB032065.D	10 Jul 2025 12:36	YP/AJ	Ok
8	Q2543-02	FB032066.D	10 Jul 2025 13:04	YP/AJ	Not Ok
9	Q2543-03	FB032067.D	10 Jul 2025 13:31	YP/AJ	Ok
10	Q2543-04	FB032068.D	10 Jul 2025 13:59	YP/AJ	Ok
11	Q2487-15	FB032069.D	10 Jul 2025 14:27	YP/AJ	Ok
12	BSF0710S2	FB032070.D	10 Jul 2025 14:54	YP/AJ	Ok
13	20 PPB GRO STD	FB032071.D	10 Jul 2025 15:22	YP/AJ	Ok
14	Q2487-16	FB032072.D	10 Jul 2025 16:24	YP/AJ	Ok
15	Q2543-02	FB032073.D	10 Jul 2025 16:52	YP/AJ	Ok
16	Q2529-01	FB032074.D	10 Jul 2025 17:20	YP/AJ	Ok
17	Q2529-02	FB032075.D	10 Jul 2025 17:48	YP/AJ	Ok
18	Q2529-03	FB032076.D	10 Jul 2025 18:15	YP/AJ	Not Ok
19	Q2529-04	FB032077.D	10 Jul 2025 18:43	YP/AJ	Ok
20	Q2529-05	FB032078.D	10 Jul 2025 19:10	YP/AJ	Ok
21	Q2529-06	FB032079.D	10 Jul 2025 19:38	YP/AJ	Not Ok

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB071025

Review By	yogesh	Review On	7/10/2025 3:17:40 PM
Supervise By	mohammad	Supervise On	7/11/2025 9:10:17 AM
SubDirectory	FB071025	HP Acquire Method	HP Processing Method FB062325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24656,PP24658,PP24659,PP24660,PP24661		
CCC Internal Standard/PEM	PP24710,PP24711,PP24712		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24662		

22	Q2529-07	FB032080.D	10 Jul 2025 20:05	YP/AJ	Ok
23	Q2529-08	FB032081.D	10 Jul 2025 20:33	YP/AJ	Not Ok
24	Q2529-09	FB032082.D	10 Jul 2025 21:00	YP/AJ	Not Ok
25	Q2529-10	FB032083.D	10 Jul 2025 21:28	YP/AJ	Ok
26	BSF0710S3	FB032084.D	10 Jul 2025 21:55	YP/AJ	Ok,M
27	20 PPB GRO STD	FB032085.D	10 Jul 2025 23:17	YP/AJ	Ok,M

M : Manual Integration