

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/10/2025

Initial Calibration Date(s): 06/18/2025

06/18/2025

Continuing Calib Time: 09:48

Initial Calibration Time(s): 11:25

13:29

GC Column: RTX-CLP ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|------------------|------------|-----------|------------------------|-------|------------|
| DICAMBA          | 7.53       | 7.54      | 7.44                   | 7.64  | 0.01       |
| 2,4-DCAA         | 7.34       | 7.35      | 7.25                   | 7.45  | 0.01       |
| DICHLORPROP      | 8.24       | 8.25      | 8.15                   | 8.35  | 0.01       |
| 2,4-D            | 8.47       | 8.49      | 8.39                   | 8.59  | 0.02       |
| 2,4,5-TP(Silvex) | 9.36       | 9.38      | 9.28                   | 9.48  | 0.02       |
| 2,4,5-T          | 9.65       | 9.67      | 9.57                   | 9.77  | 0.02       |
| 2,4-DB           | 10.23      | 10.25     | 10.15                  | 10.35 | 0.02       |
| Dinoseb          | 11.45      | 11.47     | 11.37                  | 11.57 | 0.03       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/10/2025

Initial Calibration Date(s): 06/18/2025

06/18/2025

Continuing Calib Time: 09:48

Initial Calibration Time(s): 11:25

13:29

GC Column: RTX-CLP2 ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|------------------|------------|-----------|------------------------|-------|------------|
| DICAMBA          | 7.97       | 7.97      | 7.87                   | 8.07  | 0.00       |
| 2,4-DCAA         | 7.77       | 7.77      | 7.67                   | 7.87  | 0.00       |
| DICHLORPROP      | 8.69       | 8.69      | 8.59                   | 8.79  | 0.00       |
| 2,4-D            | 9.03       | 9.03      | 8.93                   | 9.13  | 0.00       |
| 2,4,5-TP(Silvex) | 9.93       | 9.93      | 9.83                   | 10.03 | 0.00       |
| 2,4,5-T          | 10.36      | 10.36     | 10.26                  | 10.46 | 0.00       |
| 2,4-DB           | 10.93      | 10.93     | 10.83                  | 11.03 | 0.00       |
| Dinoseb          | 11.31      | 11.31     | 11.21                  | 11.41 | 0.00       |