

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/10/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030977.D **Time Analyzed:** 02:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.651	9.568	9.768	750.960	712.500	5.4
2,4,5-TP(Silvex)	9.357	9.275	9.475	694.030	712.500	-2.6
2,4-D	8.470	8.386	8.586	674.330	705.000	-4.4
2,4-DB	10.229	10.147	10.347	816.110	712.500	14.5
2,4-DCAA	7.335	7.249	7.449	690.790	750.000	-7.9
DICAMBA	7.525	7.439	7.639	664.390	705.000	-5.8
DICHLORPROP	8.238	8.153	8.353	628.000	705.000	-10.9
Dinoseb	11.446	11.366	11.566	669.630	705.000	-5.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/10/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030977.D **Time Analyzed:** 02:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.359	10.259	10.459	652.640	712.500	-8.4
2,4,5-TP(Silvex)	9.933	9.833	10.033	640.710	712.500	-10.1
2,4-D	9.027	8.927	9.127	639.890	705.000	-9.2
2,4-DB	10.927	10.827	11.027	634.340	712.500	-11.0
2,4-DCAA	7.768	7.670	7.870	647.960	750.000	-13.6
DICAMBA	7.971	7.872	8.072	629.820	705.000	-10.7
DICHLORPROP	8.691	8.591	8.791	613.040	705.000	-13.0
Dinoseb	11.312	11.211	11.411	611.410	705.000	-13.3