

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/10/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030980.D **Time Analyzed:** 09:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.652	9.568	9.768	767.030	712.500	7.7
2,4,5-TP(Silvex)	9.356	9.275	9.475	706.790	712.500	-0.8
2,4-D	8.470	8.386	8.586	689.810	705.000	-2.2
2,4-DB	10.229	10.147	10.347	835.670	712.500	17.3
2,4-DCAA	7.336	7.249	7.449	701.710	750.000	-6.4
DICAMBA	7.525	7.439	7.639	677.750	705.000	-3.9
DICHLORPROP	8.238	8.153	8.353	639.810	705.000	-9.2
Dinoseb	11.445	11.366	11.566	678.890	705.000	-3.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/10/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030980.D **Time Analyzed:** 09:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.360	10.259	10.459	657.200	712.500	-7.8
2,4,5-TP(Silvex)	9.934	9.833	10.033	652.700	712.500	-8.4
2,4-D	9.027	8.927	9.127	643.630	705.000	-8.7
2,4-DB	10.928	10.827	11.027	639.290	712.500	-10.3
2,4-DCAA	7.769	7.670	7.870	654.250	750.000	-12.8
DICAMBA	7.972	7.872	8.072	635.160	705.000	-9.9
DICHLORPROP	8.691	8.591	8.791	616.740	705.000	-12.5
Dinoseb	11.313	11.211	11.411	618.480	705.000	-12.3