

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
 Data File : PS030741.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jun 2025 12:13
 Operator : AR\AJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 06/19/2025
 Supervised By :mohammad ahmed 06/20/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 13:28:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 18 13:26:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.349	7.770	2695.0E6	788.0E6	735.040m	750.000
Target Compounds							
1) T	Dalapon	2.711	2.709	4110.7E6	1860.0E6	682.500	682.500
2) T	3,5-DICHL...	6.508	6.715	3694.2E6	1104.7E6	697.500	697.500
3) T	4-Nitroph...	7.148	7.301	1058.7E6	1177.8E6	682.500	682.500
5) T	DICAMBA	7.539	7.972	10887.5E6	4629.4E6	705.000	705.000
6) T	MCP	7.721	8.071	692.7E6	159.8E6	70.500	70.500
7) T	MCPA	7.872	8.321	834.2E6	225.9E6	69.750	69.750
8) T	DICHLORPROP	8.253	8.691	2465.2E6	1056.4E6	705.000	705.000
9) T	2,4-D	8.486	9.027	2330.4E6	1143.9E6	705.000	705.000
10) T	Pentachlo...	8.793	9.552	38633.6E6	27600.8E6	712.500	712.500
11) T	2,4,5-TP ...	9.375	9.933	14234.3E6	10200.5E6	712.500	712.500
12) T	2,4,5-T	9.668	10.359	11991.4E6	9758.3E6	712.500	712.500
13) T	2,4-DB	10.247	10.927	1715.3E6	829.1E6	712.500	712.500
14) T	DINOSEB	11.466	11.311	9809.4E6	7563.6E6	705.000	705.000
15) T	Picloram	11.275	12.421	11657.4E6	16873.3E6	712.500	712.500
16) T	DCPA	11.762	12.356	18560.8E6	15913.5E6	720.000	720.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
Data File : PS030741.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2025 12:13
Operator : AR\AJ
Sample : HSTDICC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/19/2025
Supervised By : mohammad ahmed 06/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 13:28:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Wed Jun 18 13:26:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

