

Cover Page

Order ID : Q2487

Project ID : Construction of Shafts 17B-18B - PN 220084

Client : Walsh Construction Company II, LLC

Lab Sample Number

Client Sample Number

Q2487-01	G4(1.5)
Q2487-02	G4(10)
Q2487-03	G3(9)
Q2487-04	G3(3)
Q2487-05	G2(2.5)
Q2487-06	G2(9)
Q2487-07	G1(4.5)
Q2487-08	G1(10)
Q2487-09	G4(0-6)
Q2487-10	G4(6-12)
Q2487-11	G3(0-6)
Q2487-12	G3(6-12)
Q2487-13	G2(0-6)
Q2487-14	G2(6-12)
Q2487-15	G1(0-6)
Q2487-16	G1(6-12)
Q2487-17	G4(1.5)
Q2487-18	G4(10)
Q2487-19	G3(9)
Q2487-20	G3(3)
Q2487-21	G2(2.5)
Q2487-22	G2(9)
Q2487-23	G1(10)
Q2487-24	G1(4.5)

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 7/15/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2487

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: PRADIP PRAJAPATI

Date: 07/15/2025

LAB CHRONICLE

OrderID:	Q2487	OrderDate:	7/2/2025 10:05:00 AM
Client:	Walsh Construction Company II, LLC	Project:	Construction of Shafts 17B-18B - PN 220084
Contact:	Jesse A. Sylvestri	Location:	A22,A53,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2487-09	G4(0-6)	SOIL			07/01/25			07/01/25
			Pesticide-TCL	8081B		07/03/25	07/03/25	
			Diesel Range Organics	8015D		07/08/25	07/08/25	
			Gasoline Range Organics	8015D			07/09/25	
			Herbicide	8151A		07/08/25	07/09/25	
			PCB	8082A		07/03/25	07/07/25	
			EPH_NF	NJEPH		07/03/25	07/07/25	
Q2487-09RE	G4(0-6)RE	SOIL			07/01/25			07/01/25
			Herbicide	8151A		07/08/25	07/10/25	
Q2487-10	G4(6-12)	SOIL			07/01/25			07/01/25
			PCB	8082A		07/03/25	07/04/25	
			Pesticide-TCL	8081B		07/03/25	07/03/25	
			Diesel Range Organics	8015D		07/08/25	07/08/25	
			Gasoline Range Organics	8015D			07/09/25	
			Herbicide	8151A		07/08/25	07/09/25	
			EPH_NF	NJEPH		07/03/25	07/07/25	
Q2487-11	G3(0-6)	SOIL			07/01/25			07/01/25
			PCB	8082A		07/03/25	07/04/25	
			Diesel Range Organics	8015D		07/08/25	07/09/25	
			Gasoline Range Organics	8015D			07/09/25	
			Herbicide	8151A		07/08/25	07/09/25	
			PCB	8082A		07/03/25	07/07/25	
			Pesticide-TCL	8081B		07/03/25	07/07/25	
			EPH_NF	NJEPH		07/03/25	07/07/25	
			EPH_NF	NJEPH		07/03/25	07/08/25	
Q2487-11DL	G3(0-6)DL	Solid			07/01/25			07/01/25
			EPH_NF	NJEPH		07/03/25	07/08/25	

LAB CHRONICLE

Q2487-11RE	G3(0-6)RE	SOIL			07/01/25		07/01/25
			Herbicide	8151A		07/08/25	07/10/25
Q2487-12	G3(6-12)	SOIL			07/01/25		07/01/25
			PCB	8082A		07/03/25	07/04/25
			Diesel Range Organics	8015D		07/08/25	07/08/25
			Gasoline Range Organics	8015D			07/09/25
			Herbicide	8151A		07/08/25	07/09/25
			Pesticide-TCL	8081B		07/03/25	07/07/25
			EPH_NF	NJEPH		07/03/25	07/07/25
Q2487-13	G2(0-6)	SOIL			07/01/25		07/01/25
			PCB	8082A		07/03/25	07/04/25
			Diesel Range Organics	8015D		07/08/25	07/09/25
			Gasoline Range Organics	8015D			07/09/25
			Herbicide	8151A		07/08/25	07/09/25
			Pesticide-TCL	8081B		07/03/25	07/07/25
			EPH_NF	NJEPH		07/03/25	07/07/25
Q2487-13DL	G2(0-6)DL	Solid			07/01/25		07/01/25
			EPH_NF	NJEPH		07/03/25	07/08/25
Q2487-14	G2(6-12)	SOIL			07/01/25		07/01/25
			PCB	8082A		07/03/25	07/04/25
			Diesel Range Organics	8015D		07/08/25	07/09/25
			Gasoline Range Organics	8015D			07/09/25
			Herbicide	8151A		07/08/25	07/10/25
			Pesticide-TCL	8081B		07/03/25	07/07/25
			EPH_NF	NJEPH		07/03/25	07/07/25
Q2487-14DL	G2(6-12)DL	Solid			07/01/25		07/01/25
			EPH_NF	NJEPH		07/03/25	07/08/25
Q2487-15	G1(0-6)	SOIL			07/01/25		07/01/25
			Diesel Range Organics	8015D		07/08/25	07/09/25
			Gasoline Range Organics	8015D			07/10/25
			Herbicide	8151A		07/08/25	07/10/25
			PCB	8082A		07/03/25	07/08/25
			Pesticide-TCL	8081B		07/03/25	07/07/25
			EPH_NF	NJEPH		07/03/25	07/07/25

LAB CHRONICLE

Q2487-15DL	G1(0-6)DL	Solid	EPH_NF	NJEPH		07/03/25	07/08/25	
					07/01/25			07/01/25
Q2487-16	G1(6-12)	SOIL	EPH_NF	NJEPH		07/03/25	07/08/25	
					07/01/25			07/01/25
			Diesel Range Organics	8015D		07/08/25	07/09/25	
			Gasoline Range Organics	8015D			07/10/25	
			Herbicide	8151A		07/08/25	07/10/25	
			PCB	8082A		07/03/25	07/07/25	
			Pesticide-TCL	8081B		07/03/25	07/07/25	
			EPH_NF	NJEPH		07/03/25	07/07/25	
			EPH_NF	NJEPH		07/03/25	07/08/25	
					07/01/25			07/01/25
Q2487-16DL	G1(6-12)DL	Solid	EPH_NF	NJEPH		07/03/25	07/08/25	

Hit Summary Sheet
SW-846

SDG No.: Q2487

Order ID: Q2487

Client: Walsh Construction Company II, LLC

Project ID: Construction of Shafts 17B-18B - PN 2

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :	G4(6-12)							
Q2487-10	G4(6-12)	SOIL	2,4-D	124		12.4	92.1	ug/Kg
Total Concentration:				124.000				
Client ID :	G3(6-12)							
Q2487-12	G3(6-12)	SOIL	2,4-D	72.1 J		12.9	95.3	ug/Kg
Total Concentration:				72.100				



QC SUMMARY

Surrogate Summary

SDG No.: Q2487

Client: Walsh Construction Company II, LLC

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PS030738.D	PIBLK-PS030738.D	2,4-DCAA	1	500	414	83		61	136
		2,4-DCAA	2	500	487	97		61	136
I.BLK-PS030958.D	PIBLK-PS030958.D	2,4-DCAA	1	500	456	91		61	136
		2,4-DCAA	2	500	462	92		61	136
Q2493-01MS	WC-11MS	2,4-DCAA	1	500	487	97		10	141
		2,4-DCAA	2	500	428	86		10	141
Q2493-01MSD	WC-11MSD	2,4-DCAA	1	500	489	98		10	141
		2,4-DCAA	2	500	426	85		10	141
PB168753BL	PB168753BL	2,4-DCAA	1	500	414	83		10	141
		2,4-DCAA	2	500	441	88		10	141
PB168753BS	PB168753BS	2,4-DCAA	1	500	539	108		10	141
		2,4-DCAA	2	500	505	101		10	141
I.BLK-PS030966.D	PIBLK-PS030966.D	2,4-DCAA	1	500	449	90		61	136
		2,4-DCAA	2	500	467	93		61	136
Q2487-09	G4(0-6)	2,4-DCAA	1	500	41.6	8	*	10	141
		2,4-DCAA	2	500	53.1	11		10	141
Q2487-10	G4(6-12)	2,4-DCAA	1	500	309	62		10	141
		2,4-DCAA	2	500	345	69		10	141
Q2487-11	G3(0-6)	2,4-DCAA	1	500	37.4	7	*	10	141
		2,4-DCAA	2	500	52.6	11		10	141
Q2487-12	G3(6-12)	2,4-DCAA	1	500	118	24		10	141
		2,4-DCAA	2	500	115	23		10	141
Q2487-13	G2(0-6)	2,4-DCAA	1	500	54.6	11		10	141
		2,4-DCAA	2	500	59.8	12		10	141
I.BLK-PS030976.D	PIBLK-PS030976.D	2,4-DCAA	1	500	463	93		61	136
		2,4-DCAA	2	500	467	93		61	136
I.BLK-PS030979.D	PIBLK-PS030979.D	2,4-DCAA	1	500	474	95		61	136
		2,4-DCAA	2	500	473	95		61	136
Q2487-09RE	G4(0-6)RE	2,4-DCAA	1	500	42.3	8	*	10	141
		2,4-DCAA	2	500	53.1	11		10	141
Q2487-11RE	G3(0-6)RE	2,4-DCAA	1	500	38.9	8	*	10	141
		2,4-DCAA	2	500	51.5	10		10	141
Q2487-14	G2(6-12)	2,4-DCAA	1	500	128	26		10	141
		2,4-DCAA	2	500	116	23		10	141
Q2487-15	G1(0-6)	2,4-DCAA	1	500	89.4	18		10	141
		2,4-DCAA	2	500	90.7	18		10	141
Q2487-16	G1(6-12)	2,4-DCAA	1	500	56.3	11		10	141
		2,4-DCAA	2	500	56.2	11		10	141
I.BLK-PS030989.D	PIBLK-PS030989.D	2,4-DCAA	1	500	476	95		61	136
		2,4-DCAA	2	500	476	95		61	136

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2487</u>	Analytical Method:	<u>8151A</u>
Client:	<u>Walsh Construction Company II, LLC</u>	DataFile :	<u>PS030961.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD		Low	Limits	
			Result	Result				RPD	Qual		High	RPD
Lab Sample ID:	Q2493-01MS (Column 1)	Client Sample ID:		WC-11MS								
	DICAMBA	187.5	0	137	ug/Kg	73				10	112	
	DICHLORPROP	187.5	0	144	ug/Kg	77				10	113	
	2,4-D	187.5	0	545	ug/Kg	291	*			10	144	
	2,4,5-TP(Silvex)	187.5	0	143	ug/Kg	76				10	114	
	2,4,5-T	187.5	0	155	ug/Kg	83				10	115	
	2,4-DB	187.5	0	134	ug/Kg	71				10	140	
	Dinoseb	187.5	0	0	ug/Kg	0	*			10	118	
Lab Sample ID:	Q2493-01MS (Column 2)	Client Sample ID:		WC-11MS								
	DICAMBA	187.5	0	133	ug/Kg	71				10	112	
	DICHLORPROP	187.5	0	142	ug/Kg	76				10	113	
	2,4-D	187.5	0	124	ug/Kg	66	*			10	144	
	2,4,5-TP(Silvex)	187.5	0	142	ug/Kg	76				10	114	
	2,4,5-T	187.5	0	140	ug/Kg	75				10	115	
	2,4-DB	187.5	0	117	ug/Kg	62				10	140	
	Dinoseb	187.5	0	0	ug/Kg	0	*			10	118	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2487</u>	Analytical Method:	<u>8151A</u>
Client:	<u>Walsh Construction Company II, LLC</u>	DataFile :	<u>PS030962.D</u>

		Sample				Rec		RPD		Limits		
	Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2493-01MSD	Client Sample ID:		WC-11MSD								
	(Column 1)											
	DICAMBA	187.4	0	138	ug/Kg	74		1		10	112	20
	DICHLORPROP	187.4	0	133	ug/Kg	71		8		10	113	20
	2,4-D	187.4	0	562	ug/Kg	300	*	3		10	144	20
	2,4,5-TP(Silvex)	187.4	0	144	ug/Kg	77		1		10	114	20
	2,4,5-T	187.4	0	159	ug/Kg	85		2		10	115	20
	2,4-DB	187.4	0	126	ug/Kg	67		6		10	140	20
	Dinoseb	187.4	0	0	ug/Kg	0	*	0		10	118	20
Lab Sample ID:	Q2493-01MSD	Client Sample ID:		WC-11MSD								
	(Column 2)											
	DICAMBA	187.4	0	133	ug/Kg	71		0		10	112	20
	DICHLORPROP	187.4	0	143	ug/Kg	76		0		10	113	20
	2,4-D	187.4	0	128	ug/Kg	68	*	3		10	144	20
	2,4,5-TP(Silvex)	187.4	0	143	ug/Kg	76		0		10	114	20
	2,4,5-T	187.4	0	142	ug/Kg	76		1		10	115	20
	2,4-DB	187.4	0	119	ug/Kg	64		3		10	140	20
	Dinoseb	187.4	0	0	ug/Kg	0	*	0		10	118	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2487

Analytical Method: 8151A

Client: Walsh Construction Company II, LLC

Datafile : PS030964.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD			Limits	
						RPD	Qual	Qual	Low	High
PB168753BS (Column 1)	DICAMBA	166.6	172	ug/Kg	103				72	129
	DICHLORPROP	166.6	163	ug/Kg	98				77	135
	2,4-D	166.6	170	ug/Kg	102				65	144
	2,4,5-TP(Silvex)	166.6	176	ug/Kg	106				74	146
	2,4,5-T	166.6	185	ug/Kg	111				77	134
	2,4-DB	166.6	193	ug/Kg	116				72	122
	Dinoseb	166.6	168	ug/Kg	101				74	132
PB168753BS (Column 2)	DICAMBA	166.6	162	ug/Kg	97				72	129
	DICHLORPROP	166.6	159	ug/Kg	95				77	135
	2,4-D	166.6	153	ug/Kg	92				65	144
	2,4,5-TP(Silvex)	166.6	167	ug/Kg	100				74	146
	2,4,5-T	166.6	168	ug/Kg	101				77	134
	2,4-DB	166.6	161	ug/Kg	97				72	122
	Dinoseb	166.6	153	ug/Kg	92				74	132

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB168753BL

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Lab Sample ID: PB168753BL

Lab File ID: PS030963.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 07/08/2025

Date Analyzed (1): 07/09/2025

Date Analyzed (2): 07/09/2025

Time Analyzed (1): 13:12

Time Analyzed (2): 13:12

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column (1): RTX-CLP ID: 0.32 (mm)

GC Column (2): RTX-CLP2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
WC-11MS	Q2493-01MS	PS030961.D	07/09/2025	07/09/2025
WC-11MSD	Q2493-01MSD	PS030962.D	07/09/2025	07/09/2025
PB168753BS	PB168753BS	PS030964.D	07/09/2025	07/09/2025
G4 (0-6)	Q2487-09	PS030968.D	07/09/2025	07/09/2025
G4 (6-12)	Q2487-10	PS030969.D	07/09/2025	07/09/2025
G3 (0-6)	Q2487-11	PS030970.D	07/09/2025	07/09/2025
G3 (6-12)	Q2487-12	PS030971.D	07/09/2025	07/09/2025
G2 (0-6)	Q2487-13	PS030972.D	07/09/2025	07/09/2025
G2 (6-12)	Q2487-14	PS030985.D	07/10/2025	07/10/2025
G1 (0-6)	Q2487-15	PS030986.D	07/10/2025	07/10/2025
G1 (6-12)	Q2487-16	PS030987.D	07/10/2025	07/10/2025

COMMENTS: _____



SAMPLE DATA

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/01/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/01/25
Client Sample ID:	G4(0-6)	SDG No.:	Q2487
Lab Sample ID:	Q2487-09	Matrix:	SOIL
Analytical Method:	8151A	% Solid:	87.1
Sample Wt/Vol:	30.07	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	8151A		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030968.D	1	07/08/25 13:30	07/09/25 15:17	PB168753

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1918-00-9	DICAMBA	8.90	U	8.90	76.7	ug/Kg
120-36-5	DICHLORPROP	14.7	U	14.7	76.7	ug/Kg
94-75-7	2,4-D	10.4	U	10.4	76.7	ug/Kg
93-72-1	2,4,5-TP (Silvex)	10.4	U	10.4	76.7	ug/Kg
93-76-5	2,4,5-T	10.0	U	10.0	76.7	ug/Kg
94-82-6	2,4-DB	27.7	U	27.7	76.7	ug/Kg
88-85-7	DINOSEB	12.4	U	12.4	76.7	ug/Kg
SURROGATES						
19719-28-9	2,4-DCAA	53.1		10 - 141	11%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030968.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 15:17
Operator : AR\AJ
Sample : Q2487-09
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G4(0-6)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:43:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.336	7.767	153.0E6	56910442	41.643	53.090 #

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030968.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 15:17
Operator : AR\AJ
Sample : Q2487-09
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G4(0-6)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:43:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.336 min
Delta R.T.: -0.014 min
Response: 153013014
Conc: 41.64 ng/ml

Instrument :
ECD_S
ClientSampleId :
G4(0-6)



#4 2,4-DCAA

R.T.: 7.767 min
Delta R.T.: 0.000 min
Response: 56910442
Conc: 53.09 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/01/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/01/25
Client Sample ID:	G4(0-6)RE	SDG No.:	Q2487
Lab Sample ID:	Q2487-09RE	Matrix:	SOIL
Analytical Method:	8151A	% Solid:	87.1
Sample Wt/Vol:	30.07	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	8151A		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030983.D	1	07/08/25 13:30	07/10/25 11:46	PB168753

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1918-00-9	DICAMBA	8.90	U	8.90	76.7	ug/Kg
120-36-5	DICHLORPROP	14.7	U	14.7	76.7	ug/Kg
94-75-7	2,4-D	10.4	U	10.4	76.7	ug/Kg
93-72-1	2,4,5-TP (Silvex)	10.4	U	10.4	76.7	ug/Kg
93-76-5	2,4,5-T	10.0	U	10.0	76.7	ug/Kg
94-82-6	2,4-DB	27.7	U	27.7	76.7	ug/Kg
88-85-7	DINOSEB	12.4	U	12.4	76.7	ug/Kg
SURROGATES						
19719-28-9	2,4-DCAA	53.1		10 - 141	11%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS030983.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 11:46
Operator : AR\AJ
Sample : Q2487-09RE
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G4(0-6)RE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 22:56:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.336	7.764	155.6E6	56919126	42.337	53.098 #

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS030983.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 11:46
Operator : AR\AJ
Sample : Q2487-09RE
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G4(0-6)RE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 22:56:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.336 min
Delta R.T.: -0.014 min
Response: 155561015
Conc: 42.34 ng/ml

Instrument :
ECD_S
ClientSampleId :
G4(0-6)RE



#4 2,4-DCAA

R.T.: 7.764 min
Delta R.T.: -0.004 min
Response: 56919126
Conc: 53.10 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/01/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/01/25
Client Sample ID:	G4(6-12)	SDG No.:	Q2487
Lab Sample ID:	Q2487-10	Matrix:	SOIL
Analytical Method:	8151A	% Solid:	72.6
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	8151A		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030969.D	1	07/08/25 13:30	07/09/25 15:41	PB168753

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1918-00-9	DICAMBA	10.7	U	10.7	92.1	ug/Kg
120-36-5	DICHLORPROP	17.6	U	17.6	92.1	ug/Kg
94-75-7	2,4-D	124		12.4	92.1	ug/Kg
93-72-1	2,4,5-TP (Silvex)	12.5	U	12.5	92.1	ug/Kg
93-76-5	2,4,5-T	12.0	U	12.0	92.1	ug/Kg
94-82-6	2,4-DB	33.3	U	33.3	92.1	ug/Kg
88-85-7	DINOSEB	14.8	U	14.8	92.1	ug/Kg
SURROGATES						
19719-28-9	2,4-DCAA	345		10 - 141	69%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
 Data File : PS030969.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jul 2025 15:41
 Operator : AR\AJ
 Sample : Q2487-10
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 G4(6-12)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 01:43:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 19 05:11:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.337	7.768	1134.8E6	369.5E6	308.842	344.663
Target Compounds							
9) T	2,4-D	8.490	9.025	834.9E6	461.2E6	242.387	271.398



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030969.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 15:41
Operator : AR\AJ
Sample : Q2487-10
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G4(6-12)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:43:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 7.337 min
Delta R.T.: -0.013 min
Response: 1134806449
Conc: 308.84 ng/ml

Instrument :
ECD_S
ClientSampleId :
G4(6-12)




#4 2,4-DCAA

R.T.: 7.768 min
Delta R.T.: 0.000 min
Response: 369466160
Conc: 344.66 ng/ml



#9 2,4-D

R.T.: 8.490 min
Delta R.T.: 0.003 min
Response: 834880847
Conc: 242.39 ng/ml



#9 2,4-D

R.T.: 9.025 min
Delta R.T.: 0.000 min
Response: 461215668
Conc: 271.40 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/01/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/01/25
Client Sample ID:	G3(0-6)	SDG No.:	Q2487
Lab Sample ID:	Q2487-11	Matrix:	SOIL
Analytical Method:	8151A	% Solid:	91.2
Sample Wt/Vol:	30.08	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	8151A		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030970.D	1	07/08/25 13:30	07/09/25 16:05	PB168753

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1918-00-9	DICAMBA	8.50	U	8.50	73.3	ug/Kg
120-36-5	DICHLORPROP	14.0	U	14.0	73.3	ug/Kg
94-75-7	2,4-D	9.90	U	9.90	73.3	ug/Kg
93-72-1	2,4,5-TP (Silvex)	9.90	U	9.90	73.3	ug/Kg
93-76-5	2,4,5-T	9.50	U	9.50	73.3	ug/Kg
94-82-6	2,4-DB	26.5	U	26.5	73.3	ug/Kg
88-85-7	DINOSEB	11.8	U	11.8	73.3	ug/Kg
SURROGATES						
19719-28-9	2,4-DCAA	52.6		10 - 141	11%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030970.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 16:05
Operator : AR\AJ
Sample : Q2487-11
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G3(0-6)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:43:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.335	7.766	137.5E6	56430026	37.424	52.642 #

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030970.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 16:05
Operator : AR\AJ
Sample : Q2487-11
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G3(0-6)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:43:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.335 min
Delta R.T.: -0.014 min
Response: 137509966
Conc: 37.42 ng/ml

Instrument :
ECD_S
ClientSampleId :
G3(0-6)



#4 2,4-DCAA

R.T.: 7.766 min
Delta R.T.: -0.002 min
Response: 56430026
Conc: 52.64 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/01/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/01/25
Client Sample ID:	G3(0-6)RE	SDG No.:	Q2487
Lab Sample ID:	Q2487-11RE	Matrix:	SOIL
Analytical Method:	8151A	% Solid:	91.2
Sample Wt/Vol:	30.08	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	8151A		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030984.D	1	07/08/25 13:30	07/10/25 12:10	PB168753

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1918-00-9	DICAMBA	8.50	U	8.50	73.3	ug/Kg
120-36-5	DICHLORPROP	14.0	U	14.0	73.3	ug/Kg
94-75-7	2,4-D	9.90	U	9.90	73.3	ug/Kg
93-72-1	2,4,5-TP (Silvex)	9.90	U	9.90	73.3	ug/Kg
93-76-5	2,4,5-T	9.50	U	9.50	73.3	ug/Kg
94-82-6	2,4-DB	26.5	U	26.5	73.3	ug/Kg
88-85-7	DINOSEB	11.8	U	11.8	73.3	ug/Kg
SURROGATES						
19719-28-9	2,4-DCAA	51.5		10 - 141	10%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS030984.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 12:10
Operator : AR\AJ
Sample : Q2487-11RE
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G3(0-6)RE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 22:56:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.334	7.765	143.1E6	55151265	38.943	51.449 #

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS030984.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 12:10
Operator : AR\AJ
Sample : Q2487-11RE
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G3(0-6)RE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 22:56:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.334 min
Delta R.T.: -0.016 min
Response: 143090517
Conc: 38.94 ng/ml

Instrument :
ECD_S
ClientSampleId :
G3(0-6)RE



#4 2,4-DCAA

R.T.: 7.765 min
Delta R.T.: -0.003 min
Response: 55151265
Conc: 51.45 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/01/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/01/25
Client Sample ID:	G3(6-12)	SDG No.:	Q2487
Lab Sample ID:	Q2487-12	Matrix:	SOIL
Analytical Method:	8151A	% Solid:	70.2
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	PH :	
Prep Method :	8151A	Decanted:	
		Final Vol:	10000
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030971.D	1	07/08/25 13:30	07/09/25 16:29	PB168753

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1918-00-9	DICAMBA	11.0	U	11.0	95.3	ug/Kg
120-36-5	DICHLORPROP	18.2	U	18.2	95.3	ug/Kg
94-75-7	2,4-D	72.1	J	12.9	95.3	ug/Kg
93-72-1	2,4,5-TP (Silvex)	12.9	U	12.9	95.3	ug/Kg
93-76-5	2,4,5-T	12.4	U	12.4	95.3	ug/Kg
94-82-6	2,4-DB	34.4	U	34.4	95.3	ug/Kg
88-85-7	DINOSEB	15.4	U	15.4	95.3	ug/Kg
SURROGATES						
19719-28-9	2,4-DCAA	118		10 - 141	24%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
 Data File : PS030971.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jul 2025 16:29
 Operator : AR\AJ
 Sample : Q2487-12
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 G3(6-12)

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 07/10/2025
 Supervised By : mohammad ahmed 07/11/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 01:43:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 19 05:11:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.336	7.769	433.5E6	122.8E6	117.986	114.555
Target Compounds						
9) T 2,4-D	8.490	9.026	496.5E6	258.1E6	144.133m	151.893

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030971.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 16:29
Operator : AR\AJ
Sample : Q2487-12
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_S
ClientSampleId :
G3(6-12)


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/10/2025
Supervised By :mohammad ahmed 07/11/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:43:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.336 min
Delta R.T.: -0.013 min
Response: 433525614
Conc: 117.99 ng/ml

Instrument :
ECD_S
ClientSampleId :
G3(6-12)


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/10/2025
Supervised By :mohammad ahmed 07/11/2025




#4 2,4-DCAA

R.T.: 7.769 min
Delta R.T.: 0.001 min
Response: 122799070
Conc: 114.56 ng/ml



#9 2,4-D

R.T.: 8.490 min
Delta R.T.: 0.004 min
Response: 496453550
Conc: 144.13 ng/ml m



#9 2,4-D

R.T.: 9.026 min
Delta R.T.: 0.000 min
Response: 258127852
Conc: 151.89 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/01/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/01/25
Client Sample ID:	G2(0-6)	SDG No.:	Q2487
Lab Sample ID:	Q2487-13	Matrix:	SOIL
Analytical Method:	8151A	% Solid:	89.8
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	PH :	
Prep Method :	8151A	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030972.D	1	07/08/25 13:30	07/09/25 16:53	PB168753

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1918-00-9	DICAMBA	8.60	U	8.60	74.6	ug/Kg
120-36-5	DICHLORPROP	14.2	U	14.2	74.6	ug/Kg
94-75-7	2,4-D	10.1	U	10.1	74.6	ug/Kg
93-72-1	2,4,5-TP (Silvex)	10.1	U	10.1	74.6	ug/Kg
93-76-5	2,4,5-T	9.70	U	9.70	74.6	ug/Kg
94-82-6	2,4-DB	26.9	U	26.9	74.6	ug/Kg
88-85-7	DINOSEB	12.0	U	12.0	74.6	ug/Kg
SURROGATES						
19719-28-9	2,4-DCAA	59.8		10 - 141	12%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030972.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 16:53
Operator : AR\AJ
Sample : Q2487-13
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G2(0-6)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:43:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.335	7.769	200.7E6	64094437	54.608	59.792

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030972.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 16:53
Operator : AR\AJ
Sample : Q2487-13
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G2(0-6)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:43:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.335 min
Delta R.T.: -0.014 min
Response: 200650604
Conc: 54.61 ng/ml

Instrument :
ECD_S
ClientSampleId :
G2(0-6)



#4 2,4-DCAA

R.T.: 7.769 min
Delta R.T.: 0.000 min
Response: 64094437
Conc: 59.79 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/01/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/01/25
Client Sample ID:	G2(6-12)	SDG No.:	Q2487
Lab Sample ID:	Q2487-14	Matrix:	SOIL
Analytical Method:	8151A	% Solid:	67.7
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	PH :	
Prep Method :	8151A	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030985.D	1	07/08/25 13:30	07/10/25 13:03	PB168753

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1918-00-9	DICAMBA	11.4	U	11.4	98.9	ug/Kg
120-36-5	DICHLORPROP	18.9	U	18.9	98.9	ug/Kg
94-75-7	2,4-D	13.3	U	13.3	98.9	ug/Kg
93-72-1	2,4,5-TP (Silvex)	13.4	U	13.4	98.9	ug/Kg
93-76-5	2,4,5-T	12.8	U	12.8	98.9	ug/Kg
94-82-6	2,4-DB	35.7	U	35.7	98.9	ug/Kg
88-85-7	DINOSEB	15.9	U	15.9	98.9	ug/Kg
SURROGATES						
19719-28-9	2,4-DCAA	128		10 - 141	26%	SPK: 500

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS030985.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 13:03
Operator : AR\AJ
Sample : Q2487-14
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G2(6-12)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 22:57:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.336	7.767	469.8E6	124.5E6	127.851	116.126

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS030985.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 13:03
Operator : AR\AJ
Sample : Q2487-14
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G2(6-12)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 22:57:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.336 min
Delta R.T.: -0.014 min
Response: 469776720
Conc: 127.85 ng/ml

Instrument :
ECD_S
ClientSampleId :
G2(6-12)



#4 2,4-DCAA

R.T.: 7.767 min
Delta R.T.: -0.001 min
Response: 124482910
Conc: 116.13 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/01/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/01/25
Client Sample ID:	G1(0-6)	SDG No.:	Q2487
Lab Sample ID:	Q2487-15	Matrix:	SOIL
Analytical Method:	8151A	% Solid:	80.7
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	PH :	
Prep Method :	8151A	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030986.D	1	07/08/25 13:30	07/10/25 14:20	PB168753

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1918-00-9	DICAMBA	9.60	U	9.60	82.9	ug/Kg
120-36-5	DICHLORPROP	15.8	U	15.8	82.9	ug/Kg
94-75-7	2,4-D	11.2	U	11.2	82.9	ug/Kg
93-72-1	2,4,5-TP (Silvex)	11.2	U	11.2	82.9	ug/Kg
93-76-5	2,4,5-T	10.8	U	10.8	82.9	ug/Kg
94-82-6	2,4-DB	29.9	U	29.9	82.9	ug/Kg
88-85-7	DINOSEB	13.4	U	13.4	82.9	ug/Kg
SURROGATES						
19719-28-9	2,4-DCAA	90.7		10 - 141	18%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS030986.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 14:20
Operator : AR\AJ
Sample : Q2487-15
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G1(0-6)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 22:57:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.336	7.764	328.5E6	97258974	89.399	90.730

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS030986.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 14:20
Operator : AR\AJ
Sample : Q2487-15
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G1(0-6)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 22:57:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.336 min
Delta R.T.: -0.014 min
Response: 328487168
Conc: 89.40 ng/ml

Instrument :
ECD_S
ClientSampleId :
G1(0-6)



#4 2,4-DCAA

R.T.: 7.764 min
Delta R.T.: -0.003 min
Response: 97258974
Conc: 90.73 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/01/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/01/25
Client Sample ID:	G1(6-12)	SDG No.:	Q2487
Lab Sample ID:	Q2487-16	Matrix:	SOIL
Analytical Method:	8151A	% Solid:	93.3
Sample Wt/Vol:	30.05	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	8151A		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030987.D	1	07/08/25 13:30	07/10/25 14:44	PB168753

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1918-00-9	DICAMBA	8.30	U	8.30	71.7	ug/Kg
120-36-5	DICHLORPROP	13.7	U	13.7	71.7	ug/Kg
94-75-7	2,4-D	9.70	U	9.70	71.7	ug/Kg
93-72-1	2,4,5-TP (Silvex)	9.70	U	9.70	71.7	ug/Kg
93-76-5	2,4,5-T	9.30	U	9.30	71.7	ug/Kg
94-82-6	2,4-DB	25.9	U	25.9	71.7	ug/Kg
88-85-7	DINOSEB	11.6	U	11.6	71.7	ug/Kg
SURROGATES						
19719-28-9	2,4-DCAA	56.3		10 - 141	11%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
 Data File : PS030987.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jul 2025 14:44
 Operator : AR\AJ
 Sample : Q2487-16
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 G1(6-12)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 22:58:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 19 05:11:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.335	7.767	206.9E6	60199970	56.301	56.159

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS030987.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 14:44
Operator : AR\AJ
Sample : Q2487-16
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
G1(6-12)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 22:58:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.335 min
Delta R.T.: -0.014 min
Response: 206871290
Conc: 56.30 ng/ml

Instrument :
ECD_S
ClientSampleId :
G1(6-12)



#4 2,4-DCAA

R.T.: 7.767 min
Delta R.T.: -0.001 min
Response: 60199970
Conc: 56.16 ng/ml



CALIBRATION SUMMARY

Calibration Times: 11:25 13:29

LAB FILE ID:	RT 200 =	<u>PS030739.D</u>	RT 500 =	<u>PS030740.D</u>
	RT 750 =	PS030741.D	RT 1000 =	PS030742.D
			RT 1500 =	PS030743.D

[illegible]

Calibration Times: **11:25** **13:29**

LAB FILE ID:	RT 200 =	<u>PS030739.D</u>	RT 500 =	<u>PS030740.D</u>
	RT 750 =	PS030741.D	RT 1000 =	PS030742.D
			RT 1500 =	PS030743.D

[illegible]



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Instrument ID: ECD_S

Calibration Date(s): 06/18/2025 06/18/2025

Calibration Times: 11:25 13:29

GC Column: RTX-CLP

ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS030739.D</u>	CF 500 =	<u>PS030740.D</u>		
CF 750 =		<u>PS030741.D</u>	CF 1000 =	<u>PS030742.D</u>	CF 1500 =	<u>PS030743.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-T	17256700000	16680000000	16830000000	16187900000	15311000000	16453200000	5
2,4,5-TP(Silvex)	22282500000	20395400000	19977900000	18863900000	18632300000	20030400000	7
2,4-D	3868070000	3398090000	3305570000	3150720000	3499600000	3444410000	8
2,4-DB	2481600000	2363020000	2407410000	2376040000	1839440000	2293500000	11
2,4-DCAA	4451200000	3788090000	3593380000	3449860000	3089450000	3674390000	14
DICAMBA	17792700000	15914300000	15443300000	14605900000	12319200000	15215100000	13
DICHLORPROP	4637630000	3720970000	3496760000	3281660000	3591790000	3745760000	14
Dinoseb	15211800000	14001800000	13914000000	13244000000	12994300000	13873200000	6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Instrument ID: ECD_S

Calibration Date(s): 06/18/2025 06/18/2025

Calibration Times: 11:25 13:29

GC Column: RTX-CLP2

ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS030739.D</u>	CF 500 =	<u>PS030740.D</u>		
CF 750 =		<u>PS030741.D</u>	CF 1000 =	<u>PS030742.D</u>	CF 1500 =	<u>PS030743.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-T	15648900000	14018900000	13695900000	12933300000	12026100000	13664600000	10
2,4,5-TP(Silvex)	16493700000	14706300000	14316500000	13463300000	13163600000	14428700000	9
2,4-D	1965540000	1671750000	1622570000	1523690000	1713470000	1699410000	10
2,4-DB	1373320000	1199380000	1163600000	1108750000	1069470000	1182900000	10
2,4-DCAA	1283250000	1084080000	1050700000	993038000	948747000	1071960000	12
DICAMBA	7304970000	6653080000	6566460000	6277600000	5963960000	6553210000	8
DICHLORPROP	1855180000	1562920000	1498450000	1412800000	1533120000	1572500000	11
Dinoseb	12099400000	10862400000	10728600000	10218500000	9997410000	10781200000	8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
 Data File : PS030739.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jun 2025 11:25
 Operator : AR\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

HSTDICC200

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 06/19/2025

Supervised By :mohammad ahmed 06/20/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 13:36:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 18 13:31:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.350	7.769	890.2E6	256.7E6	238.920	237.129
Target Compounds							
1) T	Dalapon	2.708	2.710	1216.6E6	586.5E6	209.922m	209.695
2) T	3,5-DICHL...	6.508	6.716	1200.4E6	355.0E6	220.928	217.928
3) T	4-Nitroph...	7.150	7.302	330.2E6	350.3E6	207.247	199.990
5) T	DICAMBA	7.540	7.972	3345.0E6	1373.3E6	213.508	207.020
6) T	MCP	7.718	8.067	159.0E6	35228767	16.559	16.286
7) T	MCPA	7.868	8.315	212.7E6	56624493	17.882	17.775
8) T	DICHLORPROP	8.254	8.691	871.9E6	348.8E6	236.185	225.527
9) T	2,4-D	8.487	9.027	727.2E6	369.5E6	214.344	221.836
10) T	Pentachlo...	8.794	9.551	12128.2E6	8330.8E6	234.117	217.854
11) T	2,4,5-TP ...	9.375	9.933	4233.7E6	3133.8E6	210.774	216.695
12) T	2,4,5-T	9.670	10.359	3278.8E6	2973.3E6	196.020	214.899
13) T	2,4-DB	10.249	10.927	471.5E6	260.9E6	193.094	218.864
14) T	DINOSEB	11.468	11.312	2859.8E6	2274.7E6	204.681	209.757
15) T	Picloram	11.278	12.421	3022.6E6	4586.9E6	185.564	196.104
16) T	DCPA	11.763	12.355	5510.6E6	4823.7E6	213.250	217.126

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
Data File : PS030739.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2025 11:25
Operator : AR\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

HSTDICC200

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 06/19/2025

Supervised By :mohammad ahmed 06/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 13:36:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Wed Jun 18 13:31:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
 Data File : PS030740.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jun 2025 11:49
 Operator : AR\AJ
 Sample : HSTDICC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 13:34:18 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 18 13:31:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.349	7.769	1894.0E6	542.0E6	534.313	525.189
Target Compounds							
1) T	Dalapon	2.708	2.709	2862.2E6	1279.7E6	484.339	475.652
2) T	3,5-DICHL...	6.508	6.715	2569.3E6	760.8E6	496.180	488.050
3) T	4-Nitroph...	7.149	7.302	729.5E6	790.6E6	474.299	462.739
5) T	DICAMBA	7.540	7.972	7479.7E6	3126.9E6	494.182	483.595
6) T	MCPD	7.720	8.069	436.8E6	102.3E6	44.172	45.746
7) T	MCPA	7.870	8.318	540.3E6	147.9E6	45.001	45.908
8) T	DICHLORPROP	8.254	8.692	1748.9E6	734.6E6	506.186	499.946
9) T	2,4-D	8.487	9.027	1597.1E6	785.7E6	487.842	493.919
10) T	Pentachlo...	8.795	9.552	26722.4E6	18957.2E6	547.623	514.600
11) T	2,4,5-TP ...	9.376	9.933	9687.8E6	6985.5E6	495.865	500.612
12) T	2,4,5-T	9.670	10.359	7923.0E6	6659.0E6	477.455	497.589
13) T	2,4-DB	10.247	10.927	1122.4E6	569.7E6	461.548	496.724
14) T	DINOSEB	11.468	11.312	6580.8E6	5105.3E6	481.683	484.806
15) T	Picloram	11.277	12.420	7497.3E6	11149.1E6	457.602	480.512
16) T	DCPA	11.762	12.356	12752.8E6	10935.3E6	507.557	508.871



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
Data File : PS030740.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2025 11:49
Operator : AR\AJ
Sample : HSTDICC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 13:34:18 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Wed Jun 18 13:31:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
 Data File : PS030741.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jun 2025 12:13
 Operator : AR\AJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

HSTDICC750

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 06/19/2025

Supervised By :mohammad ahmed 06/20/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 13:28:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 18 13:26:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.349	7.770	2695.0E6	788.0E6	735.040m	750.000
Target Compounds							
1) T	Dalapon	2.711	2.709	4110.7E6	1860.0E6	682.500	682.500
2) T	3,5-DICHL...	6.508	6.715	3694.2E6	1104.7E6	697.500	697.500
3) T	4-Nitroph...	7.148	7.301	1058.7E6	1177.8E6	682.500	682.500
5) T	DICAMBA	7.539	7.972	10887.5E6	4629.4E6	705.000	705.000
6) T	MCP	7.721	8.071	692.7E6	159.8E6	70.500	70.500
7) T	MCPA	7.872	8.321	834.2E6	225.9E6	69.750	69.750
8) T	DICHLORPROP	8.253	8.691	2465.2E6	1056.4E6	705.000	705.000
9) T	2,4-D	8.486	9.027	2330.4E6	1143.9E6	705.000	705.000
10) T	Pentachlo...	8.793	9.552	38633.6E6	27600.8E6	712.500	712.500
11) T	2,4,5-TP ...	9.375	9.933	14234.3E6	10200.5E6	712.500	712.500
12) T	2,4,5-T	9.668	10.359	11991.4E6	9758.3E6	712.500	712.500
13) T	2,4-DB	10.247	10.927	1715.3E6	829.1E6	712.500	712.500
14) T	DINOSEB	11.466	11.311	9809.4E6	7563.6E6	705.000	705.000
15) T	Picloram	11.275	12.421	11657.4E6	16873.3E6	712.500	712.500
16) T	DCPA	11.762	12.356	18560.8E6	15913.5E6	720.000	720.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
Data File : PS030741.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2025 12:13
Operator : AR\AJ
Sample : HSTDICC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

HSTDICC750

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 06/19/2025

Supervised By :mohammad ahmed 06/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 13:28:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Wed Jun 18 13:26:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
 Data File : PS030742.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jun 2025 12:37
 Operator : AR\AJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

HSTDICC1000

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 06/19/2025

Supervised By :mohammad ahmed 06/20/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 13:32:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 18 13:31:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

4) S	2,4-DCAA	7.349	7.770	3449.9E6	993.0E6	1007.482m	978.598
------	----------	-------	-------	----------	---------	-----------	---------

Target Compounds

1) T	Dalapon	2.711	2.710	5150.3E6	2361.2E6	890.676	891.083
2) T	3,5-DICHL...	6.508	6.716	4617.8E6	1397.4E6	912.175	911.477
3) T	4-Nitroph...	7.149	7.302	1346.0E6	1506.7E6	887.706	886.926
5) T	DICAMBA	7.540	7.973	13729.6E6	5900.9E6	922.935	921.488
6) T	MCP	7.723	8.073	915.4E6	207.0E6	90.761	91.801
7) T	MCPA	7.873	8.323	1089.6E6	295.3E6	89.789	91.284
8) T	DICHLORPROP	8.254	8.692	3084.8E6	1328.0E6	916.366	923.462
9) T	2,4-D	8.486	9.027	2961.7E6	1432.3E6	916.252	915.890
10) T	Pentachlo...	8.799	9.553	44933.5E6	34544.9E6	970.273	964.535
11) T	2,4,5-TP ...	9.375	9.933	17920.7E6	12790.2E6	930.890	933.380
12) T	2,4,5-T	9.669	10.359	15378.5E6	12286.7E6	928.339	932.902
13) T	2,4-DB	10.248	10.927	2257.2E6	1053.3E6	919.502	932.595
14) T	DINOSEB	11.468	11.313	12449.4E6	9605.3E6	918.844	921.811
15) T	Picloram	11.275	12.421	15192.8E6	21456.6E6	916.121	928.345
16) T	DCPA	11.763	12.357	23325.9E6	19939.4E6	946.473	946.862

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
Data File : PS030742.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2025 12:37
Operator : AR\AJ
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/19/2025
Supervised By : mohammad ahmed 06/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 13:32:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Wed Jun 18 13:31:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
 Data File : PS030743.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jun 2025 13:29
 Operator : AR\AJ
 Sample : HSTDICC1500
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

HSTDICC1500

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 06/19/2025

Supervised By :mohammad ahmed 06/20/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 13:51:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 18 13:50:13 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.349	7.768	4634.2E6	1423.1E6	1255.782m	1327.583
Target Compounds						
1) T Dalapon	2.708	2.707	7057.1E6	3212.8E6	1238.824	1171.806
2) T 3,5-DICHL...	6.509	6.713	6482.8E6	2001.5E6	1205.529	1240.719
3) T 4-Nitroph...	7.149	7.299	1960.3E6	2216.2E6	1243.237	1278.444
5) T DICAMBA	7.540	7.971	17370.1E6	8409.2E6	1141.638	1283.215
6) T MCPP	7.727	8.076	1357.6E6	277.5E6	144.610	132.298
7) T MCPA	7.878	8.325	1663.5E6	431.1E6	141.809	137.032
8) T DICHLORPROP	8.254	8.690	5064.4E6	2161.7E6	1352.041	1374.694
9) T 2,4-D	8.487	9.026	4934.4E6	2416.0E6	1432.593	1421.668
10) T Pentachlo...	8.783	9.553	61297.3E6	45248.4E6	1166.760m	1186.933
11) T 2,4,5-TP ...	9.377	9.932	26551.1E6	18758.2E6	1325.538	1300.061
12) T 2,4,5-T	9.670	10.358	21818.2E6	17137.2E6	1326.082	1254.129
13) T 2,4-DB	10.250	10.925	2621.2E6	1524.0E6	1142.883	1288.354
14) T DINOSEB	11.468	11.311	18321.9E6	14096.3E6	1320.671	1307.487
15) T Picloram	11.276	12.419	23276.8E6	31814.5E6	1447.914	1368.873
16) T DCPA	11.765	12.356	32738.8E6	27988.3E6	1278.048	1269.724

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
Data File : PS030743.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2025 13:29
Operator : AR\AJ
Sample : HSTDICC1500
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_S
ClientSampleId :
HSTDICC1500

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/19/2025
Supervised By : mohammad ahmed 06/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 13:51:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Wed Jun 18 13:50:13 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
 Data File : PS030744.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jun 2025 14:18
 Operator : AR\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

ICVPS061825

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 06/19/2025

Supervised By :mohammad ahmed 06/20/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 05:14:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 19 05:11:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.348	7.770	2655.3E6	781.4E6	722.658m	728.902
Target Compounds							
1) T	Dalapon	2.710	2.710	4056.7E6	1841.2E6	680.013	671.536
2) T	3,5-DICHL...	6.508	6.715	3653.0E6	1096.7E6	679.304	679.863
3) T	4-Nitroph...	7.149	7.302	1067.2E6	1168.5E6	676.796	674.057
5) T	DICAMBA	7.539	7.972	10807.3E6	4588.7E6	710.304	700.216
6) T	MCPD	7.721	8.072	688.5E6	156.4E6	73.332	74.580
7) T	MCPA	7.871	8.321	828.7E6	223.9E6	70.648	71.187
8) T	DICHLORPROP	8.253	8.691	2460.1E6	1042.7E6	656.766	663.094
9) T	2,4-D	8.486	9.027	2352.8E6	1124.1E6	683.086	661.445
10) T	Pentachlo...	8.794	9.552	38388.3E6	27302.7E6	725.329	716.192
11) T	2,4,5-TP ...	9.375	9.934	14291.4E6	10089.0E6	713.485	699.233
12) T	2,4,5-T	9.669	10.359	12186.1E6	9656.1E6	740.654	706.651
13) T	2,4-DB	10.248	10.927	1788.9E6	824.5E6	779.995	697.026
14) T	DINOSEB	11.467	11.312	9884.9E6	7472.6E6	712.518	693.112
15) T	Picloram	11.276	12.421	11908.4E6	16569.9E6	740.750	712.948
16) T	DCPA	11.763	12.357	18641.3E6	15691.5E6	727.713	711.865

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
Data File : PS030744.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2025 14:18
Operator : AR\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

ICVPS061825

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 06/19/2025

Supervised By :mohammad ahmed 06/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 05:14:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/09/2025

Initial Calibration Date(s): 06/18/2025 06/18/2025

Continuing Calib Time: 11:36

Initial Calibration Time(s): 11:25 13:29

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.53	7.54	7.44	7.64	0.01
2,4-DCAA	7.34	7.35	7.25	7.45	0.01
DICHLORPROP	8.24	8.25	8.15	8.35	0.01
2,4-D	8.47	8.49	8.39	8.59	0.02
2,4,5-TP(Silvex)	9.36	9.38	9.28	9.48	0.02
2,4,5-T	9.66	9.67	9.57	9.77	0.01
2,4-DB	10.23	10.25	10.15	10.35	0.02
Dinoseb	11.45	11.47	11.37	11.57	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/09/2025

Initial Calibration Date(s): 06/18/2025

06/18/2025

Continuing Calib Time: 11:36

Initial Calibration Time(s): 11:25

13:29

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.97	7.97	7.87	8.07	0.00
2,4-DCAA	7.76	7.77	7.67	7.87	0.01
DICHLORPROP	8.69	8.69	8.59	8.79	0.00
2,4-D	9.02	9.03	8.93	9.13	0.01
2,4,5-TP(Silvex)	9.93	9.93	9.83	10.03	0.00
2,4,5-T	10.36	10.36	10.26	10.46	0.00
2,4-DB	10.93	10.93	10.83	11.03	0.00
Dinoseb	11.31	11.31	11.21	11.41	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/09/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030959.D **Time Analyzed:** 11:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.656	9.568	9.768	725.530	712.500	1.8
2,4,5-TP(Silvex)	9.361	9.275	9.475	678.460	712.500	-4.8
2,4-D	8.474	8.386	8.586	643.260	705.000	-8.8
2,4-DB	10.233	10.147	10.347	758.610	712.500	6.5
2,4-DCAA	7.338	7.249	7.449	683.770	750.000	-8.8
DICAMBA	7.528	7.439	7.639	643.540	705.000	-8.7
DICHLORPROP	8.241	8.153	8.353	643.260	705.000	-8.8
Dinoseb	11.451	11.366	11.566	661.940	705.000	-6.1



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance Contract: WALS01
 Lab Code: ACE SDG NO.: Q2487
 GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL01 Date Analyzed: 07/09/2025
 Lab Sample No.: HSTDCCC750 Data File : PS030959.D Time Analyzed: 11:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.357	10.259	10.459	629.350	712.500	-11.7
2,4,5-TP(Silvex)	9.931	9.833	10.033	622.060	712.500	-12.7
2,4-D	9.024	8.927	9.127	614.410	705.000	-12.8
2,4-DB	10.925	10.827	11.027	607.740	712.500	-14.7
2,4-DCAA	7.764	7.670	7.870	694.400	750.000	-7.4
DICAMBA	7.967	7.872	8.072	602.990	705.000	-14.5
DICHLORPROP	8.687	8.591	8.791	588.230	705.000	-16.6
Dinoseb	11.311	11.211	11.411	586.380	705.000	-16.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
 Data File : PS030959.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jul 2025 11:36
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 14:14:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 19 05:11:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.338	7.764	2512.5E6	744.4E6	683.773m	694.396m
Target Compounds							
1) T	Dalapon	2.699	2.702	3575.2E6	1539.9E6	599.301	561.646
2) T	3,5-DICHL...	6.498	6.710	3405.2E6	937.6E6	633.222	581.206
3) T	4-Nitroph...	7.138	7.296	973.1E6	1004.5E6	617.127	579.447
5) T	DICAMBA	7.528	7.967	9791.5E6	3951.5E6	643.541m	602.987
6) T	MCPD	7.709	8.066	614.0E6	132.9E6	65.397m	63.351
7) T	MCPA	7.860	8.315	734.3E6	186.2E6	62.601	59.179
8) T	DICHLORPROP	8.241	8.687	2409.5E6	925.0E6	643.264	588.232
9) T	2,4-D	8.474	9.024	2215.7E6	1044.1E6	643.264m	614.408m
10) T	Pentachlo...	8.781	9.548	35706.6E6	24300.1E6	674.659	637.429
11) T	2,4,5-TP ...	9.361	9.931	13589.8E6	8975.5E6	678.457	622.060
12) T	2,4,5-T	9.656	10.357	11937.3E6	8599.9E6	725.533	629.353
13) T	2,4-DB	10.233	10.925	1739.9E6	718.9E6	758.608m	607.740
14) T	DINOSEB	11.451	11.311	9183.1E6	6321.9E6	661.935m	586.376
15) T	Picloram	11.263	12.422	12079.2E6	14739.1E6	751.377m	634.174
16) T	DCPA	11.749	12.356	18024.8E6	14058.3E6	703.648	637.772



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030959.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 11:36
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 14:14:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/09/2025

Initial Calibration Date(s): 06/18/2025 06/18/2025

Continuing Calib Time: 14:52

Initial Calibration Time(s): 11:25 13:29

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.53	7.54	7.44	7.64	0.01
2,4-DCAA	7.34	7.35	7.25	7.45	0.01
DICHLORPROP	8.24	8.25	8.15	8.35	0.01
2,4-D	8.47	8.49	8.39	8.59	0.02
2,4,5-TP(Silvex)	9.36	9.38	9.28	9.48	0.02
2,4,5-T	9.65	9.67	9.57	9.77	0.02
2,4-DB	10.23	10.25	10.15	10.35	0.02
Dinoseb	11.45	11.47	11.37	11.57	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/09/2025

Initial Calibration Date(s): 06/18/2025

06/18/2025

Continuing Calib Time: 14:52

Initial Calibration Time(s): 11:25

13:29

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.97	7.97	7.87	8.07	0.00
2,4-DCAA	7.77	7.77	7.67	7.87	0.00
DICHLORPROP	8.69	8.69	8.59	8.79	0.00
2,4-D	9.03	9.03	8.93	9.13	0.00
2,4,5-TP(Silvex)	9.93	9.93	9.83	10.03	0.00
2,4,5-T	10.36	10.36	10.26	10.46	0.00
2,4-DB	10.93	10.93	10.83	11.03	0.00
Dinoseb	11.31	11.31	11.21	11.41	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/09/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030967.D **Time Analyzed:** 14:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.653	9.568	9.768	709.820	712.500	-0.4
2,4,5-TP(Silvex)	9.359	9.275	9.475	668.070	712.500	-6.2
2,4-D	8.471	8.386	8.586	643.940	705.000	-8.7
2,4-DB	10.231	10.147	10.347	759.090	712.500	6.5
2,4-DCAA	7.337	7.249	7.449	668.070	750.000	-10.9
DICAMBA	7.526	7.439	7.639	648.820	705.000	-8.0
DICHLORPROP	8.240	8.153	8.353	607.260	705.000	-13.9
Dinoseb	11.447	11.366	11.566	644.130	705.000	-8.6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/09/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030967.D **Time Analyzed:** 14:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.361	10.259	10.459	635.110	712.500	-10.9
2,4,5-TP(Silvex)	9.934	9.833	10.033	629.760	712.500	-11.6
2,4-D	9.028	8.927	9.127	626.500	705.000	-11.1
2,4-DB	10.929	10.827	11.027	617.890	712.500	-13.3
2,4-DCAA	7.769	7.670	7.870	636.310	750.000	-15.2
DICAMBA	7.972	7.872	8.072	613.280	705.000	-13.0
DICHLORPROP	8.692	8.591	8.791	588.860	705.000	-16.5
Dinoseb	11.313	11.211	11.411	588.950	705.000	-16.5

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
 Data File : PS030967.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jul 2025 14:52
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 01:42:46 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 19 05:11:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.337	7.769	2454.7E6	682.1E6	668.065	636.314m
Target Compounds							
1) T	Dalapon	2.701	2.711	3411.1E6	1538.6E6	571.789	561.169
2) T	3,5-DICHL...	6.497	6.715	3365.9E6	941.2E6	625.908	583.482
3) T	4-Nitroph...	7.136	7.302	967.6E6	1003.1E6	613.686	578.618
5) T	DICAMBA	7.526	7.972	9871.9E6	4018.9E6	648.824m	613.275
6) T	MCPD	7.707	8.070	629.9E6	134.4E6	67.093m	64.081
7) T	MCPA	7.858	8.320	749.4E6	200.2E6	63.883	63.636
8) T	DICHLORPROP	8.240	8.692	2274.7E6	926.0E6	607.263	588.860m
9) T	2,4-D	8.471	9.028	2218.0E6	1064.7E6	643.935	626.498m
10) T	Pentachlo...	8.779	9.553	35073.3E6	24565.8E6	662.692	644.398
11) T	2,4,5-TP ...	9.359	9.934	13381.7E6	9086.6E6	668.068	629.756
12) T	2,4,5-T	9.653	10.361	11678.8E6	8678.5E6	709.822	635.105
13) T	2,4-DB	10.231	10.929	1741.0E6	730.9E6	759.092	617.893
14) T	DINOSEB	11.447	11.313	8936.1E6	6349.6E6	644.126m	588.947m
15) T	Picloram	11.259	12.424	11665.6E6	14630.9E6	725.649m	629.516
16) T	DCPA	11.744	12.358	17913.9E6	14185.7E6	699.319	643.551



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030967.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 14:52
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:42:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/10/2025

Initial Calibration Date(s): 06/18/2025 06/18/2025

Continuing Calib Time: 02:37

Initial Calibration Time(s): 11:25 13:29

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.53	7.54	7.44	7.64	0.01
2,4-DCAA	7.34	7.35	7.25	7.45	0.01
DICHLORPROP	8.24	8.25	8.15	8.35	0.01
2,4-D	8.47	8.49	8.39	8.59	0.02
2,4,5-TP(Silvex)	9.36	9.38	9.28	9.48	0.02
2,4,5-T	9.65	9.67	9.57	9.77	0.02
2,4-DB	10.23	10.25	10.15	10.35	0.02
Dinoseb	11.45	11.47	11.37	11.57	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/10/2025

Initial Calibration Date(s): 06/18/2025

06/18/2025

Continuing Calib Time: 02:37

Initial Calibration Time(s): 11:25

13:29

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.97	7.97	7.87	8.07	0.00
2,4-DCAA	7.77	7.77	7.67	7.87	0.00
DICHLORPROP	8.69	8.69	8.59	8.79	0.00
2,4-D	9.03	9.03	8.93	9.13	0.00
2,4,5-TP(Silvex)	9.93	9.93	9.83	10.03	0.00
2,4,5-T	10.36	10.36	10.26	10.46	0.00
2,4-DB	10.93	10.93	10.83	11.03	0.00
Dinoseb	11.31	11.31	11.21	11.41	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/10/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030977.D **Time Analyzed:** 02:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.651	9.568	9.768	750.960	712.500	5.4
2,4,5-TP(Silvex)	9.357	9.275	9.475	694.030	712.500	-2.6
2,4-D	8.470	8.386	8.586	674.330	705.000	-4.4
2,4-DB	10.229	10.147	10.347	816.110	712.500	14.5
2,4-DCAA	7.335	7.249	7.449	690.790	750.000	-7.9
DICAMBA	7.525	7.439	7.639	664.390	705.000	-5.8
DICHLORPROP	8.238	8.153	8.353	628.000	705.000	-10.9
Dinoseb	11.446	11.366	11.566	669.630	705.000	-5.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/10/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030977.D **Time Analyzed:** 02:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.359	10.259	10.459	652.640	712.500	-8.4
2,4,5-TP(Silvex)	9.933	9.833	10.033	640.710	712.500	-10.1
2,4-D	9.027	8.927	9.127	639.890	705.000	-9.2
2,4-DB	10.927	10.827	11.027	634.340	712.500	-11.0
2,4-DCAA	7.768	7.670	7.870	647.960	750.000	-13.6
DICAMBA	7.971	7.872	8.072	629.820	705.000	-10.7
DICHLORPROP	8.691	8.591	8.791	613.040	705.000	-13.0
Dinoseb	11.312	11.211	11.411	611.410	705.000	-13.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
 Data File : PS030977.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jul 2025 02:37
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 03:21:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 19 05:11:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.335	7.768	2538.3E6	694.6E6	690.794	647.957
Target Compounds							
1) T	Dalapon	2.702	2.708	3608.5E6	1595.3E6	604.890m	581.849m
2) T	3,5-DICHL...	6.496	6.713	3461.0E6	964.6E6	643.602	597.955m
3) T	4-Nitroph...	7.134	7.301	1018.4E6	1030.0E6	645.898m	594.188
5) T	DICAMBA	7.525	7.971	10108.8E6	4127.3E6	664.392m	629.816
6) T	MCPD	7.706	8.070	664.3E6	144.2E6	70.754m	68.741
7) T	MCPA	7.856	8.319	780.4E6	197.5E6	66.523	62.772
8) T	DICHLORPROP	8.238	8.691	2352.3E6	964.0E6	628.002	613.038
9) T	2,4-D	8.470	9.027	2322.7E6	1087.4E6	674.334	639.894m
10) T	Pentachlo...	8.777	9.552	36596.9E6	25148.3E6	691.480m	659.679
11) T	2,4,5-TP ...	9.357	9.933	13901.8E6	9244.6E6	694.035	640.710
12) T	2,4,5-T	9.651	10.359	12355.7E6	8918.1E6	750.963	652.641
13) T	2,4-DB	10.229	10.927	1871.7E6	750.4E6	816.107	634.342m
14) T	DINOSEB	11.446	11.312	9289.9E6	6591.8E6	669.630m	611.409m
15) T	Picloram	11.257	12.423	12859.7E6	15283.2E6	799.925m	657.584
16) T	DCPA	11.743	12.356	18328.6E6	14593.9E6	715.508	662.070m



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030977.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 02:37
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 03:21:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/10/2025

Initial Calibration Date(s): 06/18/2025

06/18/2025

Continuing Calib Time: 09:48

Initial Calibration Time(s): 11:25

13:29

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.53	7.54	7.44	7.64	0.01
2,4-DCAA	7.34	7.35	7.25	7.45	0.01
DICHLORPROP	8.24	8.25	8.15	8.35	0.01
2,4-D	8.47	8.49	8.39	8.59	0.02
2,4,5-TP(Silvex)	9.36	9.38	9.28	9.48	0.02
2,4,5-T	9.65	9.67	9.57	9.77	0.02
2,4-DB	10.23	10.25	10.15	10.35	0.02
Dinoseb	11.45	11.47	11.37	11.57	0.03



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/10/2025

Initial Calibration Date(s): 06/18/2025

06/18/2025

Continuing Calib Time: 09:48

Initial Calibration Time(s): 11:25

13:29

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.97	7.97	7.87	8.07	0.00
2,4-DCAA	7.77	7.77	7.67	7.87	0.00
DICHLORPROP	8.69	8.69	8.59	8.79	0.00
2,4-D	9.03	9.03	8.93	9.13	0.00
2,4,5-TP(Silvex)	9.93	9.93	9.83	10.03	0.00
2,4,5-T	10.36	10.36	10.26	10.46	0.00
2,4-DB	10.93	10.93	10.83	11.03	0.00
Dinoseb	11.31	11.31	11.21	11.41	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/10/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030980.D **Time Analyzed:** 09:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.652	9.568	9.768	767.030	712.500	7.7
2,4,5-TP(Silvex)	9.356	9.275	9.475	706.790	712.500	-0.8
2,4-D	8.470	8.386	8.586	689.810	705.000	-2.2
2,4-DB	10.229	10.147	10.347	835.670	712.500	17.3
2,4-DCAA	7.336	7.249	7.449	701.710	750.000	-6.4
DICAMBA	7.525	7.439	7.639	677.750	705.000	-3.9
DICHLORPROP	8.238	8.153	8.353	639.810	705.000	-9.2
Dinoseb	11.445	11.366	11.566	678.890	705.000	-3.7



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/10/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030980.D **Time Analyzed:** 09:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.360	10.259	10.459	657.200	712.500	-7.8
2,4,5-TP(Silvex)	9.934	9.833	10.033	652.700	712.500	-8.4
2,4-D	9.027	8.927	9.127	643.630	705.000	-8.7
2,4-DB	10.928	10.827	11.027	639.290	712.500	-10.3
2,4-DCAA	7.769	7.670	7.870	654.250	750.000	-12.8
DICAMBA	7.972	7.872	8.072	635.160	705.000	-9.9
DICHLORPROP	8.691	8.591	8.791	616.740	705.000	-12.5
Dinoseb	11.313	11.211	11.411	618.480	705.000	-12.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
 Data File : PS030980.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jul 2025 09:48
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 07/11/2025
 Supervised By :mohammad ahmed 07/11/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 22:55:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 19 05:11:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.336	7.769	2578.4E6	701.3E6	701.710	654.252
Target Compounds							
1) T	Dalapon	2.701	2.709	3456.7E6	1541.2E6	579.428	562.111
2) T	3,5-DICHL...	6.496	6.715	3524.2E6	970.3E6	655.350	601.487
3) T	4-Nitroph...	7.136	7.301	1046.4E6	1048.0E6	663.655	604.560
5) T	DICAMBA	7.525	7.972	10312.0E6	4162.4E6	677.749m	635.164
6) T	MCPD	7.707	8.071	658.4E6	143.2E6	70.132m	68.263
7) T	MCPA	7.857	8.320	795.2E6	206.6E6	67.787	65.690
8) T	DICHLORPROP	8.238	8.691	2396.6E6	969.8E6	639.813	616.744
9) T	2,4-D	8.470	9.027	2376.0E6	1093.8E6	689.815	643.625m
10) T	Pentachlo...	8.777	9.552	37334.0E6	25412.5E6	705.408m	666.609
11) T	2,4,5-TP ...	9.356	9.934	14157.3E6	9417.7E6	706.792	652.703
12) T	2,4,5-T	9.652	10.360	12620.1E6	8980.3E6	767.032	657.195
13) T	2,4-DB	10.229	10.928	1916.6E6	756.2E6	835.673	639.291
14) T	DINOSEB	11.445	11.313	9418.4E6	6668.0E6	678.890m	618.478
15) T	Picloram	11.257	12.423	12932.0E6	15438.9E6	804.426m	664.283
16) T	DCPA	11.743	12.357	18731.1E6	14635.5E6	731.219	663.961

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS030980.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 09:48
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

HSTDCCC750

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 07/11/2025

Supervised By :mohammad ahmed 07/11/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 22:55:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/10/2025

Initial Calibration Date(s): 06/18/2025 06/18/2025

Continuing Calib Time: 16:59

Initial Calibration Time(s): 11:25 13:29

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.52	7.54	7.44	7.64	0.02
2,4-DCAA	7.33	7.35	7.25	7.45	0.02
DICHLORPROP	8.24	8.25	8.15	8.35	0.01
2,4-D	8.47	8.49	8.39	8.59	0.02
2,4,5-TP(Silvex)	9.36	9.38	9.28	9.48	0.03
2,4,5-T	9.65	9.67	9.57	9.77	0.02
2,4-DB	10.23	10.25	10.15	10.35	0.02
Dinoseb	11.44	11.47	11.37	11.57	0.03



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/10/2025

Initial Calibration Date(s): 06/18/2025

06/18/2025

Continuing Calib Time: 16:59

Initial Calibration Time(s): 11:25

13:29

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.97	7.97	7.87	8.07	0.00
2,4-DCAA	7.77	7.77	7.67	7.87	0.00
DICHLORPROP	8.69	8.69	8.59	8.79	0.00
2,4-D	9.03	9.03	8.93	9.13	0.00
2,4,5-TP(Silvex)	9.93	9.93	9.83	10.03	0.00
2,4,5-T	10.36	10.36	10.26	10.46	0.00
2,4-DB	10.93	10.93	10.83	11.03	0.00
Dinoseb	11.31	11.31	11.21	11.41	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/10/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030990.D **Time Analyzed:** 16:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.650	9.568	9.768	782.090	712.500	9.8
2,4,5-TP(Silvex)	9.355	9.275	9.475	720.380	712.500	1.1
2,4-D	8.469	8.386	8.586	702.820	705.000	-0.3
2,4-DB	10.227	10.147	10.347	848.520	712.500	19.1
2,4-DCAA	7.334	7.249	7.449	718.940	750.000	-4.1
DICAMBA	7.524	7.439	7.639	694.160	705.000	-1.5
DICHLORPROP	8.237	8.153	8.353	649.970	705.000	-7.8
Dinoseb	11.443	11.366	11.566	705.470	705.000	0.1



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/10/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030990.D **Time Analyzed:** 16:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.359	10.259	10.459	678.370	712.500	-4.8
2,4,5-TP(Silvex)	9.932	9.833	10.033	671.710	712.500	-5.7
2,4-D	9.027	8.927	9.127	622.580	705.000	-11.7
2,4-DB	10.927	10.827	11.027	653.050	712.500	-8.3
2,4-DCAA	7.768	7.670	7.870	672.370	750.000	-10.4
DICAMBA	7.971	7.872	8.072	654.990	705.000	-7.1
DICHLORPROP	8.691	8.591	8.791	633.040	705.000	-10.2
Dinoseb	11.312	11.211	11.411	643.930	705.000	-8.7

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
 Data File : PS030990.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jul 2025 16:59
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 07/11/2025
 Supervised By :mohammad ahmed 07/11/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 11 07:16:06 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 19 05:11:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.334	7.768	2641.7E6	720.8E6	718.943	672.370
Target Compounds							
1) T	Dalapon	2.698	2.709	3496.5E6	1569.3E6	586.107	572.362
2) T	3,5-DICHL...	6.495	6.714	3605.4E6	986.8E6	670.447	611.728
3) T	4-Nitroph...	7.134	7.300	1070.4E6	1090.5E6	678.828	629.051
5) T	DICAMBA	7.524	7.971	10561.6E6	4292.3E6	694.155m	654.986
6) T	MCPD	7.705	8.069	678.2E6	143.4E6	72.238m	68.373
7) T	MCPA	7.855	8.318	810.4E6	200.5E6	69.082	63.728
8) T	DICHLORPROP	8.237	8.691	2434.6E6	995.4E6	649.971	633.036
9) T	2,4-D	8.469	9.027	2420.8E6	1058.0E6	702.823	622.578
10) T	Pentachlo...	8.776	9.551	38201.2E6	26213.8E6	721.794m	687.627
11) T	2,4,5-TP ...	9.355	9.932	14429.5E6	9692.0E6	720.381	671.713
12) T	2,4,5-T	9.650	10.359	12867.9E6	9269.7E6	782.095	678.373
13) T	2,4-DB	10.227	10.927	1946.1E6	772.5E6	848.523	653.046
14) T	DINOSEB	11.443	11.312	9787.1E6	6942.4E6	705.472m	643.929
15) T	Picloram	11.255	12.422	13142.7E6	15262.8E6	817.527m	656.706
16) T	DCPA	11.741	12.356	19012.0E6	15126.3E6	742.186	686.222m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS030990.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 16:59
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

HSTDCCC750

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 07/11/2025

Supervised By :mohammad ahmed 07/11/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 11 07:16:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Analytical Sequence

Client: Walsh Construction Company II, LLC	SDG No.: Q2487
Project: Construction of Shafts 17B-18B - PN 220084	Instrument ID: ECD_S
GC Column: RTX-CLP	ID: 0.32 (mm) Inst. Calib. Date(s): 06/18/2025 06/18/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCAA RT #	RT #
IBLK	IBLK	06/18/2025	11:01	PS030738.D	7.35	0.00
HSTDICC200	HSTDICC200	06/18/2025	11:25	PS030739.D	7.35	0.00
HSTDICC500	HSTDICC500	06/18/2025	11:49	PS030740.D	7.35	0.00
HSTDICC750	HSTDICC750	06/18/2025	12:13	PS030741.D	7.35	0.00
HSTDICC1000	HSTDICC1000	06/18/2025	12:37	PS030742.D	7.35	0.00
HSTDICC1500	HSTDICC1500	06/18/2025	13:29	PS030743.D	7.35	0.00
IBLK	IBLK	07/09/2025	09:32	PS030958.D	7.34	0.00
HSTDCCC750	HSTDCCC750	07/09/2025	11:36	PS030959.D	7.34	0.00
WC-11MS	Q2493-01MS	07/09/2025	12:24	PS030961.D	7.34	0.00
WC-11MSD	Q2493-01MSD	07/09/2025	12:48	PS030962.D	7.34	0.00
PB168753BL	PB168753BL	07/09/2025	13:12	PS030963.D	7.34	0.00
PB168753BS	PB168753BS	07/09/2025	13:36	PS030964.D	7.34	0.00
IBLK	IBLK	07/09/2025	14:28	PS030966.D	7.34	0.00
HSTDCCC750	HSTDCCC750	07/09/2025	14:52	PS030967.D	7.34	0.00
G4(0-6)	Q2487-09	07/09/2025	15:17	PS030968.D	7.34	0.00
G4(6-12)	Q2487-10	07/09/2025	15:41	PS030969.D	7.34	0.00
G3(0-6)	Q2487-11	07/09/2025	16:05	PS030970.D	7.34	0.00
G3(6-12)	Q2487-12	07/09/2025	16:29	PS030971.D	7.34	0.00
G2(0-6)	Q2487-13	07/09/2025	16:53	PS030972.D	7.34	0.00
IBLK	IBLK	07/09/2025	18:30	PS030976.D	7.34	0.00
HSTDCCC750	HSTDCCC750	07/10/2025	02:37	PS030977.D	7.34	0.00
IBLK	IBLK	07/10/2025	09:24	PS030979.D	7.34	0.00
HSTDCCC750	HSTDCCC750	07/10/2025	09:48	PS030980.D	7.34	0.00
G4(0-6)RE	Q2487-09RE	07/10/2025	11:46	PS030983.D	7.34	0.00
G3(0-6)RE	Q2487-11RE	07/10/2025	12:10	PS030984.D	7.33	0.00
G2(6-12)	Q2487-14	07/10/2025	13:03	PS030985.D	7.34	0.00
G1(0-6)	Q2487-15	07/10/2025	14:20	PS030986.D	7.34	0.00
G1(6-12)	Q2487-16	07/10/2025	14:44	PS030987.D	7.34	0.00
IBLK	IBLK	07/10/2025	16:34	PS030989.D	7.34	0.00
HSTDCCC750	HSTDCCC750	07/10/2025	16:59	PS030990.D	7.33	0.00

Analytical Sequence

Client: Walsh Construction Company II, LLC	SDG No.: Q2487
Project: Construction of Shafts 17B-18B - PN 220084	Instrument ID: ECD_S
GC Column: RTX-CLP2	ID: 0.32 (mm) Inst. Calib. Date(s): 06/18/2025 06/18/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCAA RT #	RT #
IBLK	IBLK	06/18/2025	11:01	PS030738.D	7.77	0.00
HSTDICC200	HSTDICC200	06/18/2025	11:25	PS030739.D	7.77	0.00
HSTDICC500	HSTDICC500	06/18/2025	11:49	PS030740.D	7.77	0.00
HSTDICC750	HSTDICC750	06/18/2025	12:13	PS030741.D	7.77	0.00
HSTDICC1000	HSTDICC1000	06/18/2025	12:37	PS030742.D	7.77	0.00
HSTDICC1500	HSTDICC1500	06/18/2025	13:29	PS030743.D	7.77	0.00
IBLK	IBLK	07/09/2025	09:32	PS030958.D	7.77	0.00
HSTDCCC750	HSTDCCC750	07/09/2025	11:36	PS030959.D	7.76	0.00
WC-11MS	Q2493-01MS	07/09/2025	12:24	PS030961.D	7.77	0.00
WC-11MSD	Q2493-01MSD	07/09/2025	12:48	PS030962.D	7.77	0.00
PB168753BL	PB168753BL	07/09/2025	13:12	PS030963.D	7.77	0.00
PB168753BS	PB168753BS	07/09/2025	13:36	PS030964.D	7.77	0.00
IBLK	IBLK	07/09/2025	14:28	PS030966.D	7.77	0.00
HSTDCCC750	HSTDCCC750	07/09/2025	14:52	PS030967.D	7.77	0.00
G4(0-6)	Q2487-09	07/09/2025	15:17	PS030968.D	7.77	0.00
G4(6-12)	Q2487-10	07/09/2025	15:41	PS030969.D	7.77	0.00
G3(0-6)	Q2487-11	07/09/2025	16:05	PS030970.D	7.77	0.00
G3(6-12)	Q2487-12	07/09/2025	16:29	PS030971.D	7.77	0.00
G2(0-6)	Q2487-13	07/09/2025	16:53	PS030972.D	7.77	0.00
IBLK	IBLK	07/09/2025	18:30	PS030976.D	7.77	0.00
HSTDCCC750	HSTDCCC750	07/10/2025	02:37	PS030977.D	7.77	0.00
IBLK	IBLK	07/10/2025	09:24	PS030979.D	7.77	0.00
HSTDCCC750	HSTDCCC750	07/10/2025	09:48	PS030980.D	7.77	0.00
G4(0-6)RE	Q2487-09RE	07/10/2025	11:46	PS030983.D	7.76	0.00
G3(0-6)RE	Q2487-11RE	07/10/2025	12:10	PS030984.D	7.77	0.00
G2(6-12)	Q2487-14	07/10/2025	13:03	PS030985.D	7.77	0.00
G1(0-6)	Q2487-15	07/10/2025	14:20	PS030986.D	7.76	0.00
G1(6-12)	Q2487-16	07/10/2025	14:44	PS030987.D	7.77	0.00
IBLK	IBLK	07/10/2025	16:34	PS030989.D	7.77	0.00
HSTDCCC750	HSTDCCC750	07/10/2025	16:59	PS030990.D	7.77	0.00

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

G3(6-12)

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Lab Sample ID: Q2487-12

Date(s) Analyzed: 07/09/2025 07/09/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP ID: 0.32 (mm)

GC Column:(2): RTX-CLP2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.49	8.44	8.54	68.4	5.3
	2	9.03	8.98	9.08	72.1	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

G4(6-12)

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Lab Sample ID: Q2487-10

Date(s) Analyzed: 07/09/2025 07/09/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.49	8.44	8.54	111	11.1
	2	9.03	8.98	9.08	124	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB168753BS

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Lab Sample ID: PB168753BS

Date(s) Analyzed: 07/09/2025 07/09/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
DICHLORPROP	1	8.24	8.19	8.29	163	2.5
	2	8.69	8.64	8.74	159	
2,4-D	1	8.47	8.42	8.52	170	10.5
	2	9.03	8.98	9.08	153	
2,4,5-TP(Silvex)	1	9.36	9.31	9.41	176	5.2
	2	9.93	9.88	9.98	167	
2,4,5-T	1	9.65	9.60	9.70	185	9.6
	2	10.36	10.31	10.41	168	
2,4-DB	1	10.23	10.18	10.28	193	18.1
	2	10.93	10.88	10.98	161	
Dinoseb	1	11.45	11.40	11.50	168	9.3
	2	11.31	11.26	11.36	153	
DICAMBA	1	7.53	7.48	7.58	172	6
	2	7.97	7.92	8.02	162	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

WC-11MS

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Lab Sample ID: Q2493-01MS

Date(s) Analyzed: 07/09/2025 07/09/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
DICHLORPROP	1	8.24	8.19	8.29	144	1.4
	2	8.69	8.64	8.74	142	
2,4-D	1	8.46	8.41	8.51	545	125.9
	2	9.03	8.98	9.08	124	
2,4,5-TP(Silvex)	1	9.36	9.31	9.41	143	0.7
	2	9.93	9.88	9.98	142	
2,4,5-T	1	9.65	9.60	9.70	155	10.2
	2	10.36	10.31	10.41	140	
2,4-DB	1	10.23	10.18	10.28	134	13.5
	2	10.93	10.88	10.98	117	
DICAMBA	1	7.53	7.48	7.58	137	3
	2	7.97	7.92	8.02	133	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

WC-11MSD

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Lab Sample ID: Q2493-01MSD

Date(s) Analyzed: 07/09/2025 07/09/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
DICAMBA	1	7.53	7.48	7.58	138	3.7
	2	7.97	7.92	8.02	133	
DICHLORPROP	1	8.24	8.19	8.29	133	7.2
	2	8.69	8.64	8.74	143	
2,4-D	1	8.46	8.41	8.51	562	125.8
	2	9.03	8.98	9.08	128	
2,4,5-TP(Silvex)	1	9.36	9.31	9.41	144	0.7
	2	9.93	9.88	9.98	143	
2,4,5-T	1	9.65	9.60	9.70	159	11.3
	2	10.36	10.31	10.41	142	
2,4-DB	1	10.23	10.18	10.28	126	5.7
	2	10.93	10.88	10.98	119	



QC SAMPLE DATA

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	
Client Sample ID:	PB168753BL	SDG No.:	Q2487
Lab Sample ID:	PB168753BL	Matrix:	SOIL
Analytical Method:	8151A	% Solid:	100 Decanted:
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Herbicide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	8151A		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030963.D	1	07/08/25 13:30	07/09/25 13:12	PB168753

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1918-00-9	DICAMBA	7.70	U	7.70	67.0	ug/Kg
120-36-5	DICHLORPROP	12.8	U	12.8	67.0	ug/Kg
94-75-7	2,4-D	9.00	U	9.00	67.0	ug/Kg
93-72-1	2,4,5-TP (Silvex)	9.10	U	9.10	67.0	ug/Kg
93-76-5	2,4,5-T	8.70	U	8.70	67.0	ug/Kg
94-82-6	2,4-DB	24.2	U	24.2	67.0	ug/Kg
88-85-7	DINOSEB	10.8	U	10.8	67.0	ug/Kg
SURROGATES						
19719-28-9	2,4-DCAA	441		10 - 141	88%	SPK: 500

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030963.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 13:12
Operator : AR\AJ
Sample : PB168753BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB168753BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:41:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.337	7.769	1521.2E6	472.9E6	414.006	441.174

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030963.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 13:12
Operator : AR\AJ
Sample : PB168753BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB168753BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:41:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.337 min
Delta R.T.: -0.013 min
Response: 1521220253
Conc: 414.01 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB168753BL



#4 2,4-DCAA

R.T.: 7.769 min
Delta R.T.: 0.001 min
Response: 472922456
Conc: 441.17 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	06/18/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	06/18/25
Client Sample ID:	PIBLK-PS030738.D	SDG No.:	Q2487
Lab Sample ID:	I.BLK-PS030738.D	Matrix:	WATER
Analytical Method:	8151A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	PH :	
Prep Method :	SW3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030738.D	1		06/18/25	PS061825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1918-00-9	DICAMBA	0.65	U	0.65	2.00	ug/L
120-36-5	DICHLORPROP	0.76	U	0.76	2.00	ug/L
94-75-7	2,4-D	0.92	U	0.92	2.00	ug/L
93-72-1	2,4,5-TP (Silvex)	0.78	U	0.78	2.00	ug/L
93-76-5	2,4,5-T	0.71	U	0.71	2.00	ug/L
94-82-6	2,4-DB	0.65	U	0.65	2.00	ug/L
88-85-7	DINOSEB	0.89	U	0.89	2.00	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	487		61 - 136	97%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
Data File : PS030738.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2025 11:01
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 05:14:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.349	7.770	1521.1E6	522.5E6	413.971	487.438

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
Data File : PS030738.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2025 11:01
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 05:14:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm





#4 2,4-DCAA

R.T.: 7.349 min

Delta R.T.: 0.000 min

Response: 1521092313


Conc: 413.97 ng/ml

Instrument :

ECD_S

ClientSampleId :

I.BLK



#4 2,4-DCAA

R.T.: 7.770 min

Delta R.T.: 0.002 min

Response: 522516165

Conc: 487.44 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/09/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/09/25
Client Sample ID:	PIBLK-PS030958.D	SDG No.:	Q2487
Lab Sample ID:	I.BLK-PS030958.D	Matrix:	WATER
Analytical Method:	8151A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	PH :	
Prep Method :	SW3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030958.D	1		07/09/25	PS070925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1918-00-9	DICAMBA	0.65	U	0.65	2.00	ug/L
120-36-5	DICHLORPROP	0.76	U	0.76	2.00	ug/L
94-75-7	2,4-D	0.92	U	0.92	2.00	ug/L
93-72-1	2,4,5-TP (Silvex)	0.78	U	0.78	2.00	ug/L
93-76-5	2,4,5-T	0.71	U	0.71	2.00	ug/L
94-82-6	2,4-DB	0.65	U	0.65	2.00	ug/L
88-85-7	DINOSEB	0.89	U	0.89	2.00	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	462		61 - 136	92%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030958.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 09:32
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 14:13:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.337	7.769	1674.8E6	495.3E6	455.812	462.043

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030958.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 09:32
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 14:13:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.337 min
Delta R.T.: -0.012 min
Response: 1674833902
Conc: 455.81 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK



#4 2,4-DCAA

R.T.: 7.769 min
Delta R.T.: 0.001 min
Response: 495293715
Conc: 462.04 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/09/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/09/25
Client Sample ID:	PIBLK-PS030966.D	SDG No.:	Q2487
Lab Sample ID:	I.BLK-PS030966.D	Matrix:	WATER
Analytical Method:	8151A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	PH :	
Prep Method :	SW3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030966.D	1		07/09/25	ps070925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1918-00-9	DICAMBA	0.65	U	0.65	2.00	ug/L
120-36-5	DICHLORPROP	0.76	U	0.76	2.00	ug/L
94-75-7	2,4-D	0.92	U	0.92	2.00	ug/L
93-72-1	2,4,5-TP (Silvex)	0.78	U	0.78	2.00	ug/L
93-76-5	2,4,5-T	0.71	U	0.71	2.00	ug/L
94-82-6	2,4-DB	0.65	U	0.65	2.00	ug/L
88-85-7	DINOSEB	0.89	U	0.89	2.00	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	467		61 - 136	93%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030966.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 14:28
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:42:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.337	7.769	1649.5E6	500.8E6	448.914	467.135

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030966.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 14:28
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:42:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.337 min
Delta R.T.: -0.012 min
Response: 1649487038
Conc: 448.91 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK



#4 2,4-DCAA

R.T.: 7.769 min
Delta R.T.: 0.002 min
Response: 500751864
Conc: 467.14 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/09/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/09/25
Client Sample ID:	PIBLK-PS030976.D	SDG No.:	Q2487
Lab Sample ID:	I.BLK-PS030976.D	Matrix:	WATER
Analytical Method:	8151A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	PH :	
Prep Method :	SW3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030976.D	1		07/09/25	ps070925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1918-00-9	DICAMBA	0.65	U	0.65	2.00	ug/L
120-36-5	DICHLORPROP	0.76	U	0.76	2.00	ug/L
94-75-7	2,4-D	0.92	U	0.92	2.00	ug/L
93-72-1	2,4,5-TP (Silvex)	0.78	U	0.78	2.00	ug/L
93-76-5	2,4,5-T	0.71	U	0.71	2.00	ug/L
94-82-6	2,4-DB	0.65	U	0.65	2.00	ug/L
88-85-7	DINOSEB	0.89	U	0.89	2.00	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	467		61 - 136	93%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030976.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 18:30
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:44:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.336	7.769	1701.6E6	500.9E6	463.089	467.263

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030976.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 18:30
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:44:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.336 min
Delta R.T.: -0.014 min
Response: 1701572123
Conc: 463.09 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK



#4 2,4-DCAA

R.T.: 7.769 min
Delta R.T.: 0.001 min
Response: 500889283
Conc: 467.26 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/10/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/10/25
Client Sample ID:	PIBLK-PS030979.D	SDG No.:	Q2487
Lab Sample ID:	I.BLK-PS030979.D	Matrix:	WATER
Analytical Method:	8151A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	PH :	
Prep Method :	SW3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030979.D	1		07/10/25	ps071025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1918-00-9	DICAMBA	0.65	U	0.65	2.00	ug/L
120-36-5	DICHLORPROP	0.76	U	0.76	2.00	ug/L
94-75-7	2,4-D	0.92	U	0.92	2.00	ug/L
93-72-1	2,4,5-TP (Silvex)	0.78	U	0.78	2.00	ug/L
93-76-5	2,4,5-T	0.71	U	0.71	2.00	ug/L
94-82-6	2,4-DB	0.65	U	0.65	2.00	ug/L
88-85-7	DINOSEB	0.89	U	0.89	2.00	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	474		61 - 136	95%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS030979.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 09:24
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 22:54:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.336	7.768	1742.3E6	507.1E6	474.166	473.051

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS030979.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 09:24
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 22:54:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.336 min
Delta R.T.: -0.014 min
Response: 1742272074
Conc: 474.17 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK



#4 2,4-DCAA

R.T.: 7.768 min
Delta R.T.: 0.000 min
Response: 507093052
Conc: 473.05 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/10/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/10/25
Client Sample ID:	PIBLK-PS030989.D	SDG No.:	Q2487
Lab Sample ID:	I.BLK-PS030989.D	Matrix:	WATER
Analytical Method:	8151A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	PH :	
Prep Method :	SW3510C	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030989.D	1		07/10/25	ps071025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1918-00-9	DICAMBA	0.65	U	0.65	2.00	ug/L
120-36-5	DICHLORPROP	0.76	U	0.76	2.00	ug/L
94-75-7	2,4-D	0.92	U	0.92	2.00	ug/L
93-72-1	2,4,5-TP (Silvex)	0.78	U	0.78	2.00	ug/L
93-76-5	2,4,5-T	0.71	U	0.71	2.00	ug/L
94-82-6	2,4-DB	0.65	U	0.65	2.00	ug/L
88-85-7	DINOSEB	0.89	U	0.89	2.00	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	476		61 - 136	95%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS030989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 16:34
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 11 06:46:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.336	7.767	1748.0E6	510.2E6	475.728	475.916

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071025\
Data File : PS030989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2025 16:34
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 11 06:46:50 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 7.336 min
Delta R.T.: -0.014 min
Response: 1748013305
Conc: 475.73 ng/ml

Instrument :
ECD_S
ClientSampleId :
I.BLK



#4 2,4-DCAA

R.T.: 7.767 min
Delta R.T.: 0.000 min
Response: 510164777
Conc: 475.92 ng/ml

Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	
Client Sample ID:	PB168753BS	SDG No.:	Q2487
Lab Sample ID:	PB168753BS	Matrix:	SOIL
Analytical Method:	8151A	% Solid:	100 Decanted:
Sample Wt/Vol:	30.01 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	Herbicide
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	8151A		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030964.D	1	07/08/25 13:30	07/09/25 13:36	PB168753

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1918-00-9	DICAMBA	172		7.70	67.0	ug/Kg
120-36-5	DICHLORPROP	163		12.8	67.0	ug/Kg
94-75-7	2,4-D	170		9.00	67.0	ug/Kg
93-72-1	2,4,5-TP (Silvex)	176		9.10	67.0	ug/Kg
93-76-5	2,4,5-T	185		8.70	67.0	ug/Kg
94-82-6	2,4-DB	193		24.2	67.0	ug/Kg
88-85-7	DINOSEB	168		10.8	67.0	ug/Kg
SURROGATES						
19719-28-9	2,4-DCAA	539		10 - 141	108%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
 Data File : PS030964.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jul 2025 13:36
 Operator : AR\AJ
 Sample : PB168753BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB168753BS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/10/2025
 Supervised By :mohammad ahmed 07/11/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 01:41:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 19 05:11:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.337	7.769	1979.3E6	541.7E6	538.681	505.330
Target Compounds							
1) T	Dalapon	2.706	2.712	2791.6E6	1267.6E6	467.942	462.347
2) T	3,5-DICHL...	6.497	6.715	2706.6E6	756.4E6	503.315	468.900
3) T	4-Nitroph...	7.136	7.302	775.4E6	794.5E6	491.750	458.297
5) T	DICAMBA	7.525	7.972	7871.1E6	3184.2E6	517.319m	485.899
6) T	MCPD	7.706	8.069	480.0E6	102.3E6	51.125m	48.767
7) T	MCPA	7.856	8.318	575.0E6	147.5E6	49.018	46.887
8) T	DICHLORPROP	8.239	8.691	1831.6E6	749.6E6	488.981	476.699
9) T	2,4-D	8.471	9.028	1752.7E6	779.4E6	508.857	458.655
10) T	Pentachlo...	8.778	9.552	28134.7E6	19538.4E6	531.591m	512.522
11) T	2,4,5-TP ...	9.358	9.934	10601.3E6	7247.9E6	529.259	502.327
12) T	2,4,5-T	9.652	10.360	9139.1E6	6876.5E6	555.460	503.233
13) T	2,4-DB	10.230	10.929	1326.8E6	572.9E6	578.488	484.344
14) T	DINOSEB	11.447	11.314	6999.2E6	4961.2E6	504.516m	460.166
15) T	Picloram	11.258	12.424	8600.8E6	11199.0E6	535.002m	481.854m
16) T	DCPA	11.742	12.357	14068.6E6	11338.4E6	549.205m	514.383m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030964.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 13:36
Operator : AR\AJ
Sample : PB168753BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

PB168753BS

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 07/10/2025

Supervised By :mohammad ahmed 07/11/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:41:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/02/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/02/25
Client Sample ID:	WC-11MS	SDG No.:	Q2487
Lab Sample ID:	Q2493-01MS	Matrix:	SOIL
Analytical Method:	8151A	% Solid:	88.7
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	8151A		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030961.D	1	07/08/25 13:30	07/09/25 12:24	PB168753

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1918-00-9	DICAMBA	137		8.70	75.4	ug/Kg
120-36-5	DICHLORPROP	144		14.4	75.4	ug/Kg
94-75-7	2,4-D	545	P	10.2	75.4	ug/Kg
93-72-1	2,4,5-TP (Silvex)	143		10.2	75.4	ug/Kg
93-76-5	2,4,5-T	155		9.80	75.4	ug/Kg
94-82-6	2,4-DB	134		27.2	75.4	ug/Kg
88-85-7	DINOSEB	12.2	U	12.2	75.4	ug/Kg
SURROGATES						
19719-28-9	2,4-DCAA	487		10 - 141	97%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
 Data File : PS030961.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jul 2025 12:24
 Operator : AR\AJ
 Sample : Q2493-01MS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 WC-11MS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 07/10/2025
 Supervised By :mohammad ahmed 07/11/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 01:41:18 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 19 05:11:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.336	7.769	1788.4E6	459.1E6	486.722	428.288
Target Compounds							
1) T	Dalapon	2.664f	2.680f	2699.5E6	2190.1E6	452.503m	798.808m#
2) T	3,5-DICHL...	6.498	6.715	2051.7E6	616.2E6	381.530	381.981
5) T	DICAMBA	7.526	7.972	5559.3E6	2323.9E6	365.378m	354.614
6) T	MCP	7.706	8.071	347.6E6	79842312	37.027m	38.069
7) T	MCPA	7.856	8.316	514.2E6	89761954	43.833m	28.534 #
8) T	DICHLORPROP	8.240	8.691	1435.2E6	595.6E6	383.146	378.741
9) T	2,4-D	8.459f	9.027	5006.7E6	563.7E6	1453.578	331.677 #
10) T	Pentachlo...	8.778	9.552	8532.7E6	6161.8E6	161.221	161.633
11) T	2,4,5-TP ...	9.359	9.933	7631.3E6	5447.4E6	380.987	377.542m
12) T	2,4,5-T	9.653	10.360	6821.1E6	5090.8E6	414.579	372.551m
13) T	2,4-DB	10.232	10.927	820.7E6	368.6E6	357.829m	311.608
15) T	Picloram	11.257	12.423	6065.5E6	6626.8E6	377.298m	285.127m
16) T	DCPA	11.742	12.358	8471.0E6	8405.5E6	330.688	381.325

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030961.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 12:24
Operator : AR\AJ
Sample : Q2493-01MS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

WC-11MS

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 07/10/2025

Supervised By :mohammad ahmed 07/11/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:41:18 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	Walsh Construction Company II, LLC	Date Collected:	07/02/25
Project:	Construction of Shafts 17B-18B - PN 220084	Date Received:	07/02/25
Client Sample ID:	WC-11MSD	SDG No.:	Q2487
Lab Sample ID:	Q2493-01MSD	Matrix:	SOIL
Analytical Method:	8151A	% Solid:	88.7
Sample Wt/Vol:	30.08	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	8151A		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030962.D	1	07/08/25 13:30	07/09/25 12:48	PB168753

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1918-00-9	DICAMBA	138		8.70	75.3	ug/Kg
120-36-5	DICHLORPROP	143		14.4	75.3	ug/Kg
94-75-7	2,4-D	562	P	10.2	75.3	ug/Kg
93-72-1	2,4,5-TP (Silvex)	144		10.2	75.3	ug/Kg
93-76-5	2,4,5-T	159		9.80	75.3	ug/Kg
94-82-6	2,4-DB	126		27.2	75.3	ug/Kg
88-85-7	DINOSEB	12.1	U	12.1	75.3	ug/Kg
SURROGATES						
19719-28-9	2,4-DCAA	489		10 - 141	98%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
 Data File : PS030962.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jul 2025 12:48
 Operator : AR\AJ
 Sample : Q2493-01MSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 WC-11MSD

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 07/10/2025
 Supervised By :mohammad ahmed 07/11/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 01:41:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 19 05:11:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.337	7.770	1798.3E6	456.4E6	489.414	425.726
Target Compounds							
1) T	Dalapon	2.668f	2.685	2323.0E6	1850.4E6	389.392m	674.887m#
2) T	3,5-DICHL...	6.498	6.715	2034.9E6	622.9E6	378.403	386.157
5) T	DICAMBA	7.526	7.972	5612.9E6	2333.9E6	368.900m	356.146
6) T	MCPD	7.707	8.073	345.4E6	78973867	36.790m	37.655
7) T	MCPA	7.857	8.317	527.3E6	92112621	44.951m	29.282 #
8) T	DICHLORPROP	8.240	8.692	1330.4E6	600.8E6	355.184m	382.086
9) T	2,4-D	8.459f	9.028	5165.4E6	581.0E6	1499.645m	341.906 #
10) T	Pentachlo...	8.778	9.552	8566.2E6	5754.5E6	161.854m	150.948m
11) T	2,4,5-TP ...	9.359	9.934	7694.9E6	5494.1E6	384.159	380.778m
12) T	2,4,5-T	9.653	10.360	6974.3E6	5173.3E6	423.890	378.590m
13) T	2,4-DB	10.231	10.928	770.8E6	375.9E6	336.068m	317.757
15) T	Picloram	11.258	12.423	6321.9E6	6910.1E6	393.249m	297.317m
16) T	DCPA	11.742	12.358	8511.9E6	8724.7E6	332.286	395.807m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS070925\
Data File : PS030962.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jul 2025 12:48
Operator : AR\AJ
Sample : Q2493-01MSD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

WC-11MSD

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 07/10/2025

Supervised By :mohammad ahmed 07/11/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 01:41:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Thu Jun 19 05:11:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:

PS061825

Instrument

ECD_s

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
HSTDICC200	PS030739.D	Dalapon	Abdul	6/19/2025 8:24:10 AM	mohammad	6/20/2025 8:27:44	Peak Integrated by Software
HSTDICC750	PS030741.D	2,4-DCAA	Abdul	6/19/2025 8:24:14 AM	mohammad	6/20/2025 8:27:44	Peak Integrated by Software
HSTDICC1000	PS030742.D	2,4-DCAA	Abdul	6/19/2025 8:24:17 AM	mohammad	6/20/2025 8:27:44	Peak Integrated by Software
HSTDICC1500	PS030743.D	2,4-DCAA	Abdul	6/19/2025 8:24:22 AM	mohammad	6/20/2025 8:27:44	Peak Integrated by Software
HSTDICC1500	PS030743.D	Pentachlorophenol	Abdul	6/19/2025 8:24:22 AM	mohammad	6/20/2025 8:27:44	Peak Integrated by Software
HSTDICV750	PS030744.D	2,4-DCAA	Abdul	6/19/2025 8:24:26 AM	mohammad	6/20/2025 8:27:44	Peak Integrated by Software
HSTDCCC750	PS030746.D	2,4-DCAA	Abdul	6/20/2025 8:24:58 AM	mohammad	6/20/2025 8:27:44	Peak Integrated by Software
HSTDCCC750	PS030746.D	DCPA #2	Abdul	6/20/2025 8:24:58 AM	mohammad	6/20/2025 8:27:44	Peak Integrated by Software
HSTDCCC750	PS030750.D	2,4-D #2	Abdul	6/20/2025 8:25:02 AM	mohammad	6/20/2025 8:27:44	Peak Integrated by Software
HSTDCCC750	PS030750.D	DCPA #2	Abdul	6/20/2025 8:25:02 AM	mohammad	6/20/2025 8:27:44	Peak Integrated by Software

Manual Integration Report

Sequence:	PS070925	Instrument	ECD_s
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
HSTDCCC750	PS030959.D	2,4-D	yogesh	7/15/2025 10:52:47 AM	 	 	Peak Integrated by Software
HSTDCCC750	PS030959.D	2,4-D #2	yogesh	7/15/2025 10:52:47 AM	 	 	Peak Integrated by Software
HSTDCCC750	PS030959.D	2,4-DB	yogesh	7/15/2025 10:52:47 AM	 	 	Peak Integrated by Software
HSTDCCC750	PS030959.D	2,4-DCAA	yogesh	7/15/2025 10:52:47 AM	 	 	Peak Integrated by Software
HSTDCCC750	PS030959.D	2,4-DCAA #2	yogesh	7/15/2025 10:52:47 AM	 	 	Peak Integrated by Software
HSTDCCC750	PS030959.D	DICAMBA	yogesh	7/15/2025 10:52:47 AM	 	 	Peak Integrated by Software
HSTDCCC750	PS030959.D	DINOSEB	yogesh	7/15/2025 10:52:47 AM	 	 	Peak Integrated by Software
HSTDCCC750	PS030959.D	MCPP	yogesh	7/15/2025 10:52:47 AM	 	 	Peak Integrated by Software
HSTDCCC750	PS030959.D	Picloram	yogesh	7/15/2025 10:52:47 AM	 	 	Peak Integrated by Software
Q2493-01MS	PS030961.D	2,4,5-T #2	Abdul	7/10/2025 8:19:27 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MS	PS030961.D	2,4,5-TP (SILVEX) #2	Abdul	7/10/2025 8:19:27 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MS	PS030961.D	2,4-DB	Abdul	7/10/2025 8:19:27 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MS	PS030961.D	Dalapon	Abdul	7/10/2025 8:19:27 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software

Manual Integration Report

Sequence:	PS070925	Instrument	ECD_s
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q2493-01MS	PS030961.D	Dalapon #2	Abdul	7/10/2025 8:19:27 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MS	PS030961.D	DICAMBA	Abdul	7/10/2025 8:19:27 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MS	PS030961.D	MCPA	Abdul	7/10/2025 8:19:27 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MS	PS030961.D	MCPP	Abdul	7/10/2025 8:19:27 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MS	PS030961.D	Picloram	Abdul	7/10/2025 8:19:27 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MS	PS030961.D	Picloram #2	Abdul	7/10/2025 8:19:27 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MSD	PS030962.D	2,4,5-T #2	Abdul	7/10/2025 8:21:46 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MSD	PS030962.D	2,4,5-TP (SILVEX) #2	Abdul	7/10/2025 8:21:46 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MSD	PS030962.D	2,4-D	Abdul	7/10/2025 8:21:46 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MSD	PS030962.D	2,4-DB	Abdul	7/10/2025 8:21:46 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MSD	PS030962.D	Dalapon	Abdul	7/10/2025 8:21:46 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MSD	PS030962.D	Dalapon #2	Abdul	7/10/2025 8:21:46 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MSD	PS030962.D	DCPA #2	Abdul	7/10/2025 8:21:46 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software

Manual Integration Report

Sequence:	PS070925	Instrument	ECD_s
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q2493-01MSD	PS030962.D	DICAMBA	Abdul	7/10/2025 8:21:46 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MSD	PS030962.D	DICHLORPROP	Abdul	7/10/2025 8:21:46 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MSD	PS030962.D	MCPA	Abdul	7/10/2025 8:21:46 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MSD	PS030962.D	MCPP	Abdul	7/10/2025 8:21:46 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MSD	PS030962.D	Pentachlorophenol	Abdul	7/10/2025 8:21:46 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MSD	PS030962.D	Pentachlorophenol #2	Abdul	7/10/2025 8:21:46 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MSD	PS030962.D	Picloram	Abdul	7/10/2025 8:21:46 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
Q2493-01MSD	PS030962.D	Picloram #2	Abdul	7/10/2025 8:21:46 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
PB168753BS	PS030964.D	DCPA	Abdul	7/10/2025 8:21:49 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
PB168753BS	PS030964.D	DCPA #2	Abdul	7/10/2025 8:21:49 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
PB168753BS	PS030964.D	DICAMBA	Abdul	7/10/2025 8:21:49 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
PB168753BS	PS030964.D	DINOSEB	Abdul	7/10/2025 8:21:49 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
PB168753BS	PS030964.D	MCPP	Abdul	7/10/2025 8:21:49 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software

Manual Integration Report

Sequence:	PS070925	Instrument	ECD_s
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PB168753BS	PS030964.D	Pentachlorophenol	Abdul	7/10/2025 8:21:49 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
PB168753BS	PS030964.D	Picloram	Abdul	7/10/2025 8:21:49 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
PB168753BS	PS030964.D	Picloram #2	Abdul	7/10/2025 8:21:49 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software
HSTDCCC750	PS030967.D	2,4-D #2	yogesh	7/15/2025 10:52:49 AM	 	 	Peak Integrated by Software
HSTDCCC750	PS030967.D	2,4-DCAA #2	yogesh	7/15/2025 10:52:49 AM	 	 	Peak Integrated by Software
HSTDCCC750	PS030967.D	DICAMBA	yogesh	7/15/2025 10:52:49 AM	 	 	Peak Integrated by Software
HSTDCCC750	PS030967.D	DICHLORPROP #2	yogesh	7/15/2025 10:52:49 AM	 	 	Peak Integrated by Software
HSTDCCC750	PS030967.D	DINOSEB	yogesh	7/15/2025 10:52:49 AM	 	 	Peak Integrated by Software
HSTDCCC750	PS030967.D	DINOSEB #2	yogesh	7/15/2025 10:52:49 AM	 	 	Peak Integrated by Software
HSTDCCC750	PS030967.D	MCPPP	yogesh	7/15/2025 10:52:49 AM	 	 	Peak Integrated by Software
HSTDCCC750	PS030967.D	Picloram	yogesh	7/15/2025 10:52:49 AM	 	 	Peak Integrated by Software
Q2487-12	PS030971.D	2,4-D	Abdul	7/10/2025 8:21:58 AM	mohammad	7/14/2025 4:17:20	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

PS070925

Instrument

ECD_s

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:	ps071025	Instrument	ECD_s
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
HSTDCCC750	PS030980.D	2,4-D #2	Abdul	7/11/2025 8:21:33 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS030980.D	DICAMBA	Abdul	7/11/2025 8:21:33 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS030980.D	DINOSEB	Abdul	7/11/2025 8:21:33 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS030980.D	MCPP	Abdul	7/11/2025 8:21:33 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS030980.D	Pentachlorophenol	Abdul	7/11/2025 8:21:33 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS030980.D	Picloram	Abdul	7/11/2025 8:21:33 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS030990.D	DCPA #2	Abdul	7/11/2025 8:21:42 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS030990.D	DICAMBA	Abdul	7/11/2025 8:21:42 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS030990.D	DINOSEB	Abdul	7/11/2025 8:21:42 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS030990.D	MCPP	Abdul	7/11/2025 8:21:42 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS030990.D	Pentachlorophenol	Abdul	7/11/2025 8:21:42 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS030990.D	Picloram	Abdul	7/11/2025 8:21:42 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS031003.D	2,4-DCAA	Abdul	7/11/2025 8:21:57 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software

Manual Integration Report

Sequence:

ps071025

Instrument

ECD_s

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
HSTDCCC750	PS031003.D	DICAMBA	Abdul	7/11/2025 8:21:57 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS031003.D	DINOSEB	Abdul	7/11/2025 8:21:57 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS031003.D	MCPP	Abdul	7/11/2025 8:21:57 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS031003.D	Pentachlorophenol	Abdul	7/11/2025 8:21:57 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software
HSTDCCC750	PS031003.D	Picloram	Abdul	7/11/2025 8:21:57 AM	mohammad	7/11/2025 9:09:38	Peak Integrated by Software

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS061825

Review By	Abdul	Review On	6/19/2025 8:25:06 AM
Supervise By	mohammad	Supervise On	6/20/2025 8:27:44 AM
SubDirectory	PS061825	HP Acquire Method	HP Processing Method PS061825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS030737.D	18 Jun 2025 10:37	AR\AJ	Ok
2	I.BLK	PS030738.D	18 Jun 2025 11:01	AR\AJ	Ok
3	HSTDICC200	PS030739.D	18 Jun 2025 11:25	AR\AJ	Ok,M
4	HSTDICC500	PS030740.D	18 Jun 2025 11:49	AR\AJ	Ok
5	HSTDICC750	PS030741.D	18 Jun 2025 12:13	AR\AJ	Ok,M
6	HSTDICC1000	PS030742.D	18 Jun 2025 12:37	AR\AJ	Ok,M
7	HSTDICC1500	PS030743.D	18 Jun 2025 13:29	AR\AJ	Ok,M
8	HSTDICV750	PS030744.D	18 Jun 2025 14:18	AR\AJ	Ok,M
9	I.BLK	PS030745.D	18 Jun 2025 14:44	AR\AJ	Ok
10	HSTDCCC750	PS030746.D	18 Jun 2025 15:08	AR\AJ	Ok,M
11	Q2312-01MS	PS030747.D	18 Jun 2025 17:07	AR\AJ	Ok,M
12	Q2312-01MSD	PS030748.D	18 Jun 2025 17:31	AR\AJ	Ok,M
13	I.BLK	PS030749.D	18 Jun 2025 17:55	AR\AJ	Ok
14	HSTDCCC750	PS030750.D	18 Jun 2025 18:19	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS070925

Review By	Abdul	Review On	7/10/2025 8:22:37 AM
Supervise By	mohammad	Supervise On	7/14/2025 4:17:20 AM
SubDirectory	PS070925	HP Acquire Method	HP Processing Method ps061825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS030957.D	09 Jul 2025 09:08	AR\AJ	Ok
2	I.BLK	PS030958.D	09 Jul 2025 09:32	AR\AJ	Ok
3	HSTDCCC750	PS030959.D	09 Jul 2025 11:36	AR\AJ	Ok,NS
4	Q2493-01	PS030960.D	09 Jul 2025 12:00	AR\AJ	Ok
5	Q2493-01MS	PS030961.D	09 Jul 2025 12:24	AR\AJ	Ok,M
6	Q2493-01MSD	PS030962.D	09 Jul 2025 12:48	AR\AJ	Ok,M
7	PB168753BL	PS030963.D	09 Jul 2025 13:12	AR\AJ	Ok
8	PB168753BS	PS030964.D	09 Jul 2025 13:36	AR\AJ	Ok,M
9	PB168706BS	PS030965.D	09 Jul 2025 14:04	AR\AJ	Ok,M
10	I.BLK	PS030966.D	09 Jul 2025 14:28	AR\AJ	Ok
11	HSTDCCC750	PS030967.D	09 Jul 2025 14:52	AR\AJ	Ok,NS
12	Q2487-09	PS030968.D	09 Jul 2025 15:17	AR\AJ	ReRun
13	Q2487-10	PS030969.D	09 Jul 2025 15:41	AR\AJ	Ok
14	Q2487-11	PS030970.D	09 Jul 2025 16:05	AR\AJ	ReRun
15	Q2487-12	PS030971.D	09 Jul 2025 16:29	AR\AJ	Ok,M
16	Q2487-13	PS030972.D	09 Jul 2025 16:53	AR\AJ	Ok
17	Q2487-14	PS030973.D	09 Jul 2025 17:17	AR\AJ	Not Ok
18	Q2487-15	PS030974.D	09 Jul 2025 17:42	AR\AJ	Not Ok
19	Q2487-16	PS030975.D	09 Jul 2025 18:06	AR\AJ	Not Ok
20	I.BLK	PS030976.D	09 Jul 2025 18:30	AR\AJ	Ok
21	HSTDCCC750	PS030977.D	10 Jul 2025 02:37	AR\AJ	Ok,NR

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS071025

Review By	Abdul	Review On	7/11/2025 8:22:34 AM
Supervise By	mohammad	Supervise On	7/11/2025 9:09:38 AM
SubDirectory	PS071025	HP Acquire Method	HP Processing Method ps061825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS030978.D	10 Jul 2025 09:00	AR\AJ	Ok
2	I.BLK	PS030979.D	10 Jul 2025 09:24	AR\AJ	Ok
3	HSTDCCC750	PS030980.D	10 Jul 2025 09:48	AR\AJ	Ok,M
4	Q2493-01MS	PS030981.D	10 Jul 2025 10:36	AR\AJ	Not Ok
5	Q2493-01MSD	PS030982.D	10 Jul 2025 11:15	AR\AJ	Not Ok
6	Q2487-09RE	PS030983.D	10 Jul 2025 11:46	AR\AJ	Confirms
7	Q2487-11RE	PS030984.D	10 Jul 2025 12:10	AR\AJ	Confirms
8	Q2487-14	PS030985.D	10 Jul 2025 13:03	AR\AJ	Ok
9	Q2487-15	PS030986.D	10 Jul 2025 14:20	AR\AJ	Ok
10	Q2487-16	PS030987.D	10 Jul 2025 14:44	AR\AJ	Ok
11	Q2484-03	PS030988.D	10 Jul 2025 16:06	AR\AJ	Ok
12	I.BLK	PS030989.D	10 Jul 2025 16:34	AR\AJ	Ok
13	HSTDCCC750	PS030990.D	10 Jul 2025 16:59	AR\AJ	Ok,M
14	Q2517-01	PS030991.D	10 Jul 2025 17:47	AR\AJ	ReRun
15	Q2514-01	PS030992.D	10 Jul 2025 18:11	AR\AJ	Ok
16	Q2514-02	PS030993.D	10 Jul 2025 18:35	AR\AJ	Ok
17	Q2514-03	PS030994.D	10 Jul 2025 19:00	AR\AJ	Ok
18	Q2514-04	PS030995.D	10 Jul 2025 19:24	AR\AJ	Ok
19	Q2514-05	PS030996.D	10 Jul 2025 19:48	AR\AJ	Ok
20	Q2514-06	PS030997.D	10 Jul 2025 20:12	AR\AJ	Ok
21	Q2514-07	PS030998.D	10 Jul 2025 20:36	AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QCBatch ID # PS071025

Review By	Abdul	Review On	7/11/2025 8:22:34 AM
Supervise By	mohammad	Supervise On	7/11/2025 9:09:38 AM
SubDirectory	PS071025	HP Acquire Method	HP Processing Method ps061825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

22	Q2514-08	PS030999.D	10 Jul 2025 21:00	AR\AJ	Ok
23	Q2514-09	PS031000.D	10 Jul 2025 21:25	AR\AJ	Ok,M
24	Q2514-10	PS031001.D	10 Jul 2025 21:49	AR\AJ	ReRun
25	I.BLK	PS031002.D	10 Jul 2025 22:13	AR\AJ	Ok
26	HSTDCCC750	PS031003.D	10 Jul 2025 23:00	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS061825

Review By	Abdul	Review On	6/19/2025 8:25:06 AM
Supervise By	mohammad	Supervise On	6/20/2025 8:27:44 AM
SubDirectory	PS061825	HP Acquire Method	HP Processing Method PS061825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC	PP24559		
Internal Standard/PEM			
ICV/I.BLK	PP24562		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS030737.D	18 Jun 2025 10:37		AR\AJ	Ok
2	I.BLK	I.BLK	PS030738.D	18 Jun 2025 11:01		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS030739.D	18 Jun 2025 11:25		AR\AJ	Ok,M
4	HSTDICC500	HSTDICC500	PS030740.D	18 Jun 2025 11:49		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS030741.D	18 Jun 2025 12:13		AR\AJ	Ok,M
6	HSTDICC1000	HSTDICC1000	PS030742.D	18 Jun 2025 12:37		AR\AJ	Ok,M
7	HSTDICC1500	HSTDICC1500	PS030743.D	18 Jun 2025 13:29		AR\AJ	Ok,M
8	HSTDICV750	ICVPS061825	PS030744.D	18 Jun 2025 14:18		AR\AJ	Ok,M
9	I.BLK	I.BLK	PS030745.D	18 Jun 2025 14:44		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS030746.D	18 Jun 2025 15:08		AR\AJ	Ok,M
11	Q2312-01MS	TP-1MS	PS030747.D	18 Jun 2025 17:07	some compounds fail for recovery	AR\AJ	Ok,M
12	Q2312-01MSD	TP-1MSD	PS030748.D	18 Jun 2025 17:31	some compounds fail for recovery & RPD	AR\AJ	Ok,M
13	I.BLK	I.BLK	PS030749.D	18 Jun 2025 17:55		AR\AJ	Ok
14	HSTDCCC750	HSTDCCC750	PS030750.D	18 Jun 2025 18:19		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS070925

Review By	Abdul	Review On	7/10/2025 8:22:37 AM
Supervise By	mohammad	Supervise On	7/14/2025 4:17:20 AM
SubDirectory	PS070925	HP Acquire Method	HP Processing Method ps061825 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24562
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS030957.D	09 Jul 2025 09:08		AR\AJ	Ok
2	I.BLK	I.BLK	PS030958.D	09 Jul 2025 09:32		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS030959.D	09 Jul 2025 11:36		AR\AJ	Ok,NS
4	Q2493-01	WC-11	PS030960.D	09 Jul 2025 12:00		AR\AJ	Ok
5	Q2493-01MS	WC-11MS	PS030961.D	09 Jul 2025 12:24		AR\AJ	Ok,M
6	Q2493-01MSD	WC-11MSD	PS030962.D	09 Jul 2025 12:48		AR\AJ	Ok,M
7	PB168753BL	PB168753BL	PS030963.D	09 Jul 2025 13:12		AR\AJ	Ok
8	PB168753BS	PB168753BS	PS030964.D	09 Jul 2025 13:36		AR\AJ	Ok,M
9	PB168706BS	PB168706BS	PS030965.D	09 Jul 2025 14:04		AR\AJ	Ok,M
10	I.BLK	I.BLK	PS030966.D	09 Jul 2025 14:28		AR\AJ	Ok
11	HSTDCCC750	HSTDCCC750	PS030967.D	09 Jul 2025 14:52		AR\AJ	Ok,NS
12	Q2487-09	G4(0-6)	PS030968.D	09 Jul 2025 15:17	Surrogate low in 1st column	AR\AJ	ReRun
13	Q2487-10	G4(6-12)	PS030969.D	09 Jul 2025 15:41		AR\AJ	Ok
14	Q2487-11	G3(0-6)	PS030970.D	09 Jul 2025 16:05	Surrogate low in 1st column	AR\AJ	ReRun
15	Q2487-12	G3(6-12)	PS030971.D	09 Jul 2025 16:29		AR\AJ	Ok,M
16	Q2487-13	G2(0-6)	PS030972.D	09 Jul 2025 16:53		AR\AJ	Ok
17	Q2487-14	G2(6-12)	PS030973.D	09 Jul 2025 17:17	surrogate fail in both column	AR\AJ	Not Ok
18	Q2487-15	G1(0-6)	PS030974.D	09 Jul 2025 17:42	Surrogate low in both column	AR\AJ	Not Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS070925

Review By	Abdul	Review On	7/10/2025 8:22:37 AM
Supervise By	mohammad	Supervise On	7/14/2025 4:17:20 AM
SubDirectory	PS070925	HP Acquire Method	HP Processing Method ps061825 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC Internal Standard/PEM	PP24559
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562

19	Q2487-16	G1(6-12)	PS030975.D	09 Jul 2025 18:06	Surrogate low in both column F flag in surrogate in 2nd column , no proper surrogate hit	AR\AJ	Not Ok
20	I.BLK	I.BLK	PS030976.D	09 Jul 2025 18:30		AR\AJ	Ok
21	HSTDCCC750	HSTDCCC750	PS030977.D	10 Jul 2025 02:37		AR\AJ	Ok,NR

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS071025

Review By	Abdul	Review On	7/11/2025 8:22:34 AM
Supervise By	mohammad	Supervise On	7/11/2025 9:09:38 AM
SubDirectory	PS071025	HP Acquire Method	HP Processing Method ps061825 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24562
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS030978.D	10 Jul 2025 09:00		AR\AJ	Ok
2	I.BLK	I.BLK	PS030979.D	10 Jul 2025 09:24		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS030980.D	10 Jul 2025 09:48		AR\AJ	Ok,M
4	Q2493-01MS	WC-11MS	PS030981.D	10 Jul 2025 10:36	Comp#14 not detected , F Flag in Comp#1.9.16 . some compound recovery fail	AR\AJ	Not Ok
5	Q2493-01MSD	WC-11MSD	PS030982.D	10 Jul 2025 11:15	Comp#14 not detected , F Flag in Comp#1.16 . some compound recovery fail	AR\AJ	Not Ok
6	Q2487-09RE	G4(0-6)RE	PS030983.D	10 Jul 2025 11:46	Surrogate low in 1st column	AR\AJ	Confirms
7	Q2487-11RE	G3(0-6)RE	PS030984.D	10 Jul 2025 12:10	Surrogate low in 1st column	AR\AJ	Confirms
8	Q2487-14	G2(6-12)	PS030985.D	10 Jul 2025 13:03		AR\AJ	Ok
9	Q2487-15	G1(0-6)	PS030986.D	10 Jul 2025 14:20		AR\AJ	Ok
10	Q2487-16	G1(6-12)	PS030987.D	10 Jul 2025 14:44		AR\AJ	Ok
11	Q2484-03	TP-64	PS030988.D	10 Jul 2025 16:06		AR\AJ	Ok
12	I.BLK	I.BLK	PS030989.D	10 Jul 2025 16:34		AR\AJ	Ok
13	HSTDCCC750	HSTDCCC750	PS030990.D	10 Jul 2025 16:59		AR\AJ	Ok,M
14	Q2517-01	TP-14	PS030991.D	10 Jul 2025 17:47	Surrogate low in 2nd column	AR\AJ	ReRun
15	Q2514-01	TP-92	PS030992.D	10 Jul 2025 18:11		AR\AJ	Ok
16	Q2514-02	TP-93	PS030993.D	10 Jul 2025 18:35		AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS071025

Review By	Abdul	Review On	7/11/2025 8:22:34 AM
Supervise By	mohammad	Supervise On	7/11/2025 9:09:38 AM
SubDirectory	PS071025	HP Acquire Method	HP Processing Method ps061825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC	PP24559		
Internal Standard/PEM	PP24562		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

17	Q2514-03	TP-94	PS030994.D	10 Jul 2025 19:00		AR\AJ	Ok
18	Q2514-04	TP-96	PS030995.D	10 Jul 2025 19:24		AR\AJ	Ok
19	Q2514-05	TP-97	PS030996.D	10 Jul 2025 19:48		AR\AJ	Ok
20	Q2514-06	TP-103	PS030997.D	10 Jul 2025 20:12		AR\AJ	Ok
21	Q2514-07	TP-36	PS030998.D	10 Jul 2025 20:36		AR\AJ	Ok
22	Q2514-08	TP-78	PS030999.D	10 Jul 2025 21:00		AR\AJ	Ok
23	Q2514-09	TP-81	PS031000.D	10 Jul 2025 21:25		AR\AJ	Ok,M
24	Q2514-10	TP-90	PS031001.D	10 Jul 2025 21:49	Surrogate low in 1st column	AR\AJ	ReRun
25	I.BLK	I.BLK	PS031002.D	10 Jul 2025 22:13		AR\AJ	Ok
26	HSTDCCC750	HSTDCCC750	PS031003.D	10 Jul 2025 23:00		AR\AJ	Ok,M

M : Manual Integration



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 7/7/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 07/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:37
Out Date: 07/04/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB136368

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2487-01	G4 (1.5)	1	1.15	10.21	11.36	10.14	88.1	
Q2487-02	G4 (10)	2	1.19	10.38	11.57	8.76	72.9	
Q2487-03	G3 (9)	3	1.15	10.66	11.81	7.38	58.4	
Q2487-04	G3 (3)	4	1.15	10.80	11.95	10.8	89.4	
Q2487-05	G2 (2.5)	5	1.17	10.00	11.17	10.48	93.1	
Q2487-06	G2 (9)	6	1.16	10.26	11.42	8.00	66.7	
Q2487-07	G1 (4.5)	7	1.19	10.11	11.3	7.8	65.4	
Q2487-08	G1 (10)	8	1.16	9.96	11.12	6.65	55.1	
Q2487-09	G4 (0-6)	9	1.18	10.26	11.44	10.12	87.1	
Q2487-10	G4 (6-12)	10	1.16	10.14	11.3	8.52	72.6	
Q2487-11	G3 (0-6)	11	1.18	10.42	11.6	10.68	91.2	
Q2487-12	G3 (6-12)	12	1.18	10.41	11.59	8.49	70.2	
Q2487-13	G2 (0-6)	13	1.17	9.95	11.12	10.11	89.8	
Q2487-14	G2 (6-12)	14	1.16	10.68	11.84	8.39	67.7	
Q2487-15	G1 (0-6)	15	1.17	10.55	11.72	9.68	80.7	
Q2487-16	G1 (6-12)	16	1.13	10.24	11.37	10.68	93.3	
Q2501-05	SVOC-GPC-BLANK	17	1.00	1.00	2.00	2.00	100.0	
Q2501-06	PEST-GPC-BLANK	18	1.00	1.00	2.00	2.00	100.0	
Q2501-07	PEST-GPC-BLANK-SPIKE	19	1.00	1.00	2.00	2.00	100.0	
Q2501-08	PCB-GPC-BLANK	20	1.00	1.00	2.00	2.00	100.0	
Q2501-09	PCB-GPC-BLANK-SPIKE	21	1.00	1.00	2.00	2.00	100.0	
Q2501-10	SVOC-GPC-BLANK	22	1.00	1.00	2.00	2.00	100.0	
Q2501-11	PEST-GPC-BLANK	23	1.00	1.00	2.00	2.00	100.0	
Q2501-12	PEST-GPC-BLANK-SPIKE	24	1.00	1.00	2.00	2.00	100.0	
Q2501-13	PCB-GPC2-BLANK	25	1.00	1.00	2.00	2.00	100.0	
Q2501-14	PCB-GPC2-BLANK-SPIKE	26	1.00	1.00	2.00	2.00	100.0	
Q2503-03	GCAP2	27	1.18	10.65	11.83	8.76	71.2	
Q2503-04	GCAP3	28	1.18	10.64	11.82	8.16	65.6	

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 7/7/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 07/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:37
Out Date: 07/04/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB136368

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2503-05	GCAP2A	29	1.18	10.30	11.48	9.11	77.0	
Q2504-01	WASTE	30	1.14	10.73	11.87	10.16	84.1	
Q2504-02	VOC	31	1.18	10.81	11.99	10.45	85.8	
Q2504-03	1	32	1.14	10.84	11.98	10.36	85.1	
Q2504-04	2	33	1.19	10.41	11.6	9.91	83.8	
Q2504-05	3	34	1.13	10.75	11.88	10.22	84.6	
Q2504-06	4	35	1.14	10.52	11.66	10.03	84.5	
Q2504-07	5	36	1.12	10.87	11.99	10.16	83.2	
Q2505-01	#62825	37	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2505-02	#62525	38	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2505-03	#2008	39	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2507-01	SU-04-7.3-2025	40	1.17	10.26	11.43	10.1	87.0	
Q2507-02	SU-04-7.3-2025-EPH	41	1.18	10.30	11.48	9.99	85.5	
Q2507-03	SU-04-7.3-2025-VOC	42	1.13	10.45	11.58	9.95	84.4	
Q2508-01	AUD-25-0105	43	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2508-02	AUD-25-0106	44	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2508-03	AUD-25-0107	45	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q2509-02	AUD-25-0112	46	1.00	1.00	2.00	2.00	100.0	oilly-debris
Q2510-01	#63025-A	47	1.15	11.63	12.78	10.79	82.9	
Q2510-02	#63025-A-VOC	48	1.18	10.19	11.37	9.86	85.2	
Q2513-01	HR-2-070325	49	1.18	9.89	11.07	10.51	94.3	
Q2513-02	HR-2-070325-E2	50	1.18	10.25	11.43	10.68	92.7	
Q2513-03	HR-3-070325	51	1.18	10.27	11.45	10.88	94.4	
Q2513-04	HR-3-070325-E2	52	1.14	10.67	11.81	11.23	94.6	
Q2514-01	TP-92	53	1.14	10.78	11.92	10.56	87.4	
Q2514-02	TP-93	54	1.19	10.71	11.9	10.58	87.7	
Q2514-03	TP-94	55	1.13	10.86	11.99	10.7	88.1	
Q2514-04	TP-96	56	1.14	11.10	12.24	10.66	85.8	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 7/7/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 07/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:37
Out Date: 07/04/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB136368

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2514-05	TP-97	57	1.16	10.02	11.18	9.68	85.0	
Q2514-06	TP-103	58	1.18	10.22	11.4	10.02	86.5	
Q2514-07	TP-36	59	1.15	10.78	11.93	10.88	90.3	
Q2514-08	TP-78	60	1.13	9.99	11.12	9.75	86.3	
Q2514-09	TP-81	61	1.16	10.50	11.66	10.22	86.3	
Q2514-10	TP-90	62	1.18	10.43	11.61	10.73	91.6	
Q2515-01	wc-1	63	1.15	10.31	11.46	10.17	87.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-070325

WorkList ID : 190533

Department : Wet-Chemistry

Date : 07-03-2025 08:34:29

196368

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2487-01	G4(1.5)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-02	G4(10)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-03	G3(9)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-04	G3(3)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-05	G2(2.5)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-06	G2(9)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-07	G1(4.5)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-08	G1(10)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-09	G4(0-6)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-10	G4(6-12)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-11	G3(0-6)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-12	G3(6-12)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-13	G2(0-6)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-14	G2(6-12)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-15	G1(0-6)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2487-16	G1(6-12)	Solid	Percent Solids	Cool 4 deg C	WALS01	A22	07/01/2025	Chemtech -SO
Q2501-05	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-06	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-07	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-08	PCB-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-09	PCB-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO

Date/Time 07/03/25 15:15

Raw Sample Received by: JDC

Raw Sample Relinquished by: JDC

Date/Time 07/03/25

Raw Sample Received by: JDC

Raw Sample Relinquished by: JDC

WORKLIST(Hardcopy Internal Chain)

136368

WorkList Name : %1-070325

WorkList ID : 190533

Department : Wet-Chemistry

Date : 07-03-2025 08:34:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2501-10	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-11	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-12	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-13	PCB-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2501-14	PCB-GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	D31	06/27/2025	Chemtech -SO
Q2503-03	GCAP2	Solid	Percent Solids	Cool 4 deg C	GENV01	O11	06/27/2025	Chemtech -SO
Q2503-04	GCAP3	Solid	Percent Solids	Cool 4 deg C	GENV01	O11	07/02/2025	Chemtech -SO
Q2503-05	GCAP2A	Solid	Percent Solids	Cool 4 deg C	GENV01	O11	07/02/2025	Chemtech -SO
Q2504-01	WASTE	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-02	VOC	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-03	1	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-04	2	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-05	3	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-06	4	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2504-07	5	Solid	Percent Solids	Cool 4 deg C	SCIA01	O12	07/02/2025	Chemtech -SO
Q2505-01	#62825	Solid	Percent Solids	Cool 4 deg C	PSEG03	O13	07/03/2025	Chemtech -SO
Q2505-02	#62525	Solid	Percent Solids	Cool 4 deg C	PSEG03	O13	07/03/2025	Chemtech -SO
Q2505-03	#2008	Solid	Percent Solids	Cool 4 deg C	PSEG03	O13	07/03/2025	Chemtech -SO
Q2507-01	SU-04-7.3-2025	Solid	Percent Solids	Cool 4 deg C	PSEG03	O12	07/04/2025	Chemtech -SO
Q2507-02	SU-04-7.3-2025-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	O12	07/04/2025	Chemtech -SO
Q2507-03	SU-04-7.3-2025-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	O12	07/04/2025	Chemtech -SO

Date/Time 07/03/25 15:15
Raw Sample Received by: WWC
Raw Sample Relinquished by: JDCSM

Date/Time 07/03/25 17:15
Raw Sample Received by: JDCSM
Raw Sample Relinquished by: WWC

WORKLIST(Hardcopy Internal Chain)

136368

WorkList Name : %1-070325

WorkList ID : 190533

Department : Wet-Chemistry

Date : 07-03-2025 08:34:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2508-01	AUD-25-0105	Solid	Percent Solids	Cool 4 deg C	PSEG03	O11	07/03/2025	Chemtech -SO
Q2508-02	AUD-25-0106	Solid	Percent Solids	Cool 4 deg C	PSEG03	O11	07/03/2025	Chemtech -SO
Q2508-03	AUD-25-0107	Solid	Percent Solids	Cool 4 deg C	PSEG03	O11	07/03/2025	Chemtech -SO
Q2509-02	AUD-25-0112	Solid	Percent Solids	Cool 4 deg C	PSEG03	O22	07/03/2025	Chemtech -SO
Q2510-01	#63025-A	Solid	Percent Solids	Cool 4 deg C	PSEG05	O22	07/03/2025	Chemtech -SO
Q2510-02	#63025-A-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG05	O22	07/03/2025	Chemtech -SO
Q2513-01	HR-2-070325	Solid	Percent Solids	Cool 4 deg C	PSEG05	O21	07/03/2025	Chemtech -SO
Q2513-02	HR-2-070325-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	O21	07/03/2025	Chemtech -SO
Q2513-03	HR-3-070325	Solid	Percent Solids	Cool 4 deg C	PSEG05	O21	07/03/2025	Chemtech -SO
Q2513-04	HR-3-070325-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	O21	07/03/2025	Chemtech -SO
Q2514-01	TP-92	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-02	TP-93	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-03	TP-94	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-04	TP-96	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-05	TP-97	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-06	TP-103	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-07	TP-36	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/02/2025	Chemtech -SO
Q2514-08	TP-78	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/03/2025	Chemtech -SO
Q2514-09	TP-81	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/03/2025	Chemtech -SO
Q2514-10	TP-90	Solid	Percent Solids	Cool 4 deg C	CAMP02	O21	07/03/2025	Chemtech -SO
Q2515-01	wc-1	Solid	Percent Solids	Cool 4 deg C	ENVO01	O23	07/03/2025	Chemtech -SO

Date/Time 07/03/25 1515
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time

Raw Sample Received by:

SOP ID: M8151A-Herbicide-23

Clean Up SOP #: N/A **Extraction Start Date :** 07/08/2025

Matrix : Solid **Extraction Start Time :** 13:30

Weigh By: EH **Extraction By:** RJ **Extraction End Date :** 07/09/2025

Balance check: RJ **Filter By:** RJ **Extraction End Time :** 11:25

Balance ID: EX-SC-2 **pH Meter ID:** N/A **Concentration By:** EH

pH Strip Lot#: E3880 **Hood ID:** 3,4,5,7 **Supervisor By :** RUPESH

Extraction Method: ☐ Separatory Funnel ☐ Continous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5/500 PPM	PP24654
Surrogate	1.0ML	5000 PPB	PP24653
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
MeCl2/Acetone/1:1	N/A	EP2612
Acidified Na2SO4	N/A	EP2621
Sand	N/A	E3951
HCL	N/A	M6151
DI WATER	N/A	N/A
37% KOH	N/A	EP2616
Methylene Chloride	N/A	E3943
1:3 SULPHURIC ACID	N/A	EP2598
Ether	N/A	E3952
ISO OCTANE	N/A	E3554
METHANOL	N/A	V14622
Diazomethane	N/A	EP2618
Hexane	N/A	E3950
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

pH adjusted with HCL <2 for soil Extraction, PH adjusted with 1:3 H2SO4 <2 after Hydrolysis, Derivatization procedure is completed and samples are ready to Analyze, 40ML Vial Lot # 03-40 BTS723.

KD Bath ID: N/A **Envap ID:** NEVAP-02

KD Bath Temperature: N/A **Envap Temperature:** 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
07/9/25	RS (Ext Lab)	RS (Ext Lab)
11:30	Preparation Group	Analysis Group

Analytical Method: M8151A-Herbicide-23

Concentration Date: 07/09/2025

Sample ID	Client Sample ID	Test	g/ mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB168753BL	HBLK753	Herbicide	30.02	N/A	ritesh	Evelyn	10			U1-1
PB168753BS	HLCS753	Herbicide	30.01	N/A	ritesh	Evelyn	10			2
Q2487-09	G4(0-6)	Herbicide	30.07	N/A	ritesh	Evelyn	10	D		3
Q2487-10	G4(6-12)	Herbicide	30.06	N/A	ritesh	Evelyn	10	D		4
Q2487-11	G3(0-6)	Herbicide	30.08	N/A	ritesh	Evelyn	10	D		5
Q2487-12	G3(6-12)	Herbicide	30.03	N/A	ritesh	Evelyn	10	D		6
Q2487-13	G2(0-6)	Herbicide	30.01	N/A	ritesh	Evelyn	10	D		U2-1
Q2487-14	G2(6-12)	Herbicide	30.02	N/A	ritesh	Evelyn	10	D		2
Q2487-15	G1(0-6)	Herbicide	30.04	N/A	ritesh	Evelyn	10	D		3
Q2487-16	G1(6-12)	Herbicide	30.05	N/A	ritesh	Evelyn	10	D		4
Q2493-01	WC-11	Herbicide	30.07	N/A	ritesh	Evelyn	10	D		5
Q2493-01MS	WC-11MS	Herbicide	30.06	N/A	ritesh	Evelyn	10	D		6
Q2493-01MS D	WC-11MSD	Herbicide	30.08	N/A	ritesh	Evelyn	10	D		U3-1
Q2514-01	TP-92	Herbicide	30.05	N/A	ritesh	Evelyn	10	E		2
Q2514-02	TP-93	Herbicide	30.02	N/A	ritesh	Evelyn	10	E		3
Q2514-03	TP-94	Herbicide	30.05	N/A	ritesh	Evelyn	10	E		4
Q2514-04	TP-96	Herbicide	30.01	N/A	ritesh	Evelyn	10	E		5
Q2514-05	TP-97	Herbicide	30.04	N/A	ritesh	Evelyn	10	E		6
Q2514-06	TP-103	Herbicide	30.09	N/A	ritesh	Evelyn	10	E		U6-1
Q2514-07	TP-36	Herbicide	30.07	N/A	ritesh	Evelyn	10	E		2
Q2514-08	TP-78	Herbicide	30.02	N/A	ritesh	Evelyn	10	E		3
Q2514-09	TP-81	Herbicide	30.06	N/A	ritesh	Evelyn	10	E		4
Q2514-10	TP-90	Herbicide	30.03	N/A	ritesh	Evelyn	10	E		5
Q2517-01	TP-14	Herbicide	30.05	N/A	ritesh	Evelyn	10	E		6

16802
123

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q2493H

WorkList ID : 190578

Department : Extraction

Date : 07-08-2025 08:25:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2487-09	G4(0-6)	Solid	Herbicide	Cool 4 deg C	WALS01	A22	07/01/2025	8151A
Q2487-10	G4(6-12)	Solid	Herbicide	Cool 4 deg C	WALS01	A22	07/01/2025	8151A
Q2487-11	G3(0-6)	Solid	Herbicide	Cool 4 deg C	WALS01	A22	07/01/2025	8151A
Q2487-12	G3(6-12)	Solid	Herbicide	Cool 4 deg C	WALS01	A22	07/01/2025	8151A
Q2487-13	G2(0-6)	Solid	Herbicide	Cool 4 deg C	WALS01	A22	07/01/2025	8151A
Q2487-14	G2(6-12)	Solid	Herbicide	Cool 4 deg C	WALS01	A22	07/01/2025	8151A
Q2487-15	G1(0-6)	Solid	Herbicide	Cool 4 deg C	WALS01	A22	07/01/2025	8151A
Q2487-16	G1(6-12)	Solid	Herbicide	Cool 4 deg C	WALS01	A22	07/01/2025	8151A
Q2493-01	WC-11	Solid	Herbicide	Cool 4 deg C	WALS01	A22	07/01/2025	8151A
Q2514-01	TP-92	Solid	Herbicide	Cool 4 deg C	PSEG03	A43	07/02/2025	8151A
Q2514-02	TP-93	Solid	Herbicide	Cool 4 deg C	CAMP02	O21	07/02/2025	8151A
Q2514-03	TP-94	Solid	Herbicide	Cool 4 deg C	CAMP02	O21	07/02/2025	8151A
Q2514-04	TP-96	Solid	Herbicide	Cool 4 deg C	CAMP02	O21	07/02/2025	8151A
Q2514-05	TP-97	Solid	Herbicide	Cool 4 deg C	CAMP02	O21	07/02/2025	8151A
Q2514-06	TP-103	Solid	Herbicide	Cool 4 deg C	CAMP02	O21	07/02/2025	8151A
Q2514-07	TP-36	Solid	Herbicide	Cool 4 deg C	CAMP02	O21	07/02/2025	8151A
Q2514-08	TP-78	Solid	Herbicide	Cool 4 deg C	CAMP02	O21	07/03/2025	8151A
Q2514-09	TP-81	Solid	Herbicide	Cool 4 deg C	CAMP02	O21	07/03/2025	8151A
Q2514-10	TP-90	Solid	Herbicide	Cool 4 deg C	CAMP02	O21	07/03/2025	8151A
Q2517-01	TP-14	Solid	Herbicide	Cool 4 deg C	PSEG03	O23	07/07/2025	8151A

Date/Time 07/08/25 13:25

Raw Sample Received by: RJCA166

Raw Sample Relinquished by: [Signature]

Date/Time 07/08/25 14:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: RJCA166



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Walsh Construction
ADDRESS: 150 Clare Rd, 11th Floor
CITY Little Falls STATE: NJ ZIP: 07424
ATTENTION: Benie Dion Gokan
PHONE: 646-285-7234 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Construction of Shaft 17B488
PROJECT NO.: 220084 LOCATION: Queens, NY
PROJECT MANAGER: Jesse Sylvestri
e-mail: j.sylvestri@walshgroup.com
PHONE: 201-681-9740 FAX:

CLIENT BILLING INFORMATION

BILL TO: Walsh Construction PO#:
ADDRESS: 150 Clare Rd, 11th Floor
CITY Little Falls STATE: NJ ZIP: 07424
ATTENTION: Jesse Sylvestri PHONE: 201-681-9740

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: STANDARD TAT DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☒ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT _____

ANALYSIS

1. TCL VOCs + TVS
2. TCL SVOCs + TVS
3. TCL VOCs + TVS
4. TCL VOCs + TVS
5. Rest Herb PCB
6. Rest Herb PCB
7. Rest Herb PCB
8. Rest Herb PCB
9. Rest Herb PCB
See add'l analyses in comments

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		F+E	E	E	E	E	E	E	E	E	← Specify Preservatives	
																	A-HCl B-HNO3 C-H2SO4	D-NaOH E-ICE F-OTHER (methanol)
1.	G4(1.5)	Soil		X	7/1/25	1025	6	X										3x vials (temperature set)
2.	G4(10)	↓		X		1045	↓	X										+ 2x enclosures +
3.	G3(9)			X		1150	↓	X										1x plastic
4.	G3(3)			X		1200	↓	X										
5.	G2(2.5)			X		1325	↓	X										↓
6.	G2(9)	↓		X		1330	↓	X										
7.	G1(4.5)			X		1430	↓	X										
8.	G1(10)			X		1440	↓	X										3x vials (temperature set) 1x 802 + 1x plastic + 2x enclosures
9.	G4(0-6)	↓	X			1100	7		X	X	X	X	X	X	X	X	X	7x 802 jars
10.	G4(6-12)	↓	X			1110	↓		X	X	X	X	X	X	X	X	X	↓

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP _____ °C
1. <u>[Signature]</u>	7/1/25 1600	1. <u>Benie Dion Gokan</u>	Comments: Full analyte list in B. Gokan email on 6/26/25 to J. Hedust
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	See Bottle order # B2506068 & B2506069
2. <u>Benie Dion Gokan</u>	7-1-25 1600	2. <u>[Signature]</u>	Add'l analyses - Paint Filler, Organic Content by LOI, TS, TVS,
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Ammonia & Nitrogen, COD, O.I. & Grease
3. <u>[Signature]</u>	7-1-25	3. <u>[Signature]</u>	

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Walsh Construction
ADDRESS: 150 Clure Rd, 11th Floor
CITY: Little Falls STATE: NJ ZIP: 07424
ATTENTION: Benue Dion Gordon
PHONE: 646-285-7234 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Construction of Shaft 17B-H873
PROJECT NO.: 220084 LOCATION: Queens, NY
PROJECT MANAGER: Jesse Sylvestri
e-mail: jssylvestri@walshgroup.com
PHONE: 201-681-9740 FAX:

CLIENT BILLING INFORMATION

BILL TO: Walsh Construction PO#:
ADDRESS: 150 Clure Rd, 11th Floor
CITY: Little Falls STATE: NJ ZIP: 07424
ATTENTION: Jesse Sylvestri PHONE: 201-681-9740

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: STANDARD TAT DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☒ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT _____

1. TC-VOCs + TICs
2. TC-Sulfates + TICs
3. TC-Nitrates + TICs
4. TC-Phenols + TICs
5. TC-Hydrocarbons + TICs
6. TC-Pesticides + TICs
7. TC-PCBs + TICs
8. TC-DDTs + TICs
9. TC-PAHs + TICs
10. TC-Organic Acids + TICs
11. TC-Inorganic Acids + TICs
12. TC-Ammonia + TICs
13. TC-Nitrogen + TICs
14. TC-Phosphorus + TICs
15. TC-Sulfur + TICs
16. TC-Halogenated Compounds + TICs
17. TC-Trace Metals + TICs
18. TC-Heavy Metals + TICs
19. TC-Low Level Metals + TICs
20. TC-Trace Organic Compounds + TICs
21. TC-Trace Inorganic Compounds + TICs
22. TC-Trace Gases + TICs
23. TC-Trace Volatiles + TICs
24. TC-Trace Semi-Volatiles + TICs
25. TC-Trace Non-Volatiles + TICs
26. TC-Trace Solids + TICs
27. TC-Trace Liquids + TICs
28. TC-Trace Gels + TICs
29. TC-Trace Powders + TICs
30. TC-Trace Crystals + TICs
31. TC-Trace Fibers + TICs
32. TC-Trace Particles + TICs
33. TC-Trace Dusts + TICs
34. TC-Trace Aerosols + TICs
35. TC-Trace Vapors + TICs
36. TC-Trace Gases + TICs
37. TC-Trace Liquids + TICs
38. TC-Trace Solids + TICs
39. TC-Trace Gels + TICs
40. TC-Trace Powders + TICs
41. TC-Trace Crystals + TICs
42. TC-Trace Fibers + TICs
43. TC-Trace Particles + TICs
44. TC-Trace Dusts + TICs
45. TC-Trace Aerosols + TICs
46. TC-Trace Vapors + TICs
47. TC-Trace Gases + TICs
48. TC-Trace Liquids + TICs
49. TC-Trace Solids + TICs
50. TC-Trace Gels + TICs
51. TC-Trace Powders + TICs
52. TC-Trace Crystals + TICs
53. TC-Trace Fibers + TICs
54. TC-Trace Particles + TICs
55. TC-Trace Dusts + TICs
56. TC-Trace Aerosols + TICs
57. TC-Trace Vapors + TICs
58. TC-Trace Gases + TICs
59. TC-Trace Liquids + TICs
60. TC-Trace Solids + TICs
61. TC-Trace Gels + TICs
62. TC-Trace Powders + TICs
63. TC-Trace Crystals + TICs
64. TC-Trace Fibers + TICs
65. TC-Trace Particles + TICs
66. TC-Trace Dusts + TICs
67. TC-Trace Aerosols + TICs
68. TC-Trace Vapors + TICs
69. TC-Trace Gases + TICs
70. TC-Trace Liquids + TICs
71. TC-Trace Solids + TICs
72. TC-Trace Gels + TICs
73. TC-Trace Powders + TICs
74. TC-Trace Crystals + TICs
75. TC-Trace Fibers + TICs
76. TC-Trace Particles + TICs
77. TC-Trace Dusts + TICs
78. TC-Trace Aerosols + TICs
79. TC-Trace Vapors + TICs
80. TC-Trace Gases + TICs
81. TC-Trace Liquids + TICs
82. TC-Trace Solids + TICs
83. TC-Trace Gels + TICs
84. TC-Trace Powders + TICs
85. TC-Trace Crystals + TICs
86. TC-Trace Fibers + TICs
87. TC-Trace Particles + TICs
88. TC-Trace Dusts + TICs
89. TC-Trace Aerosols + TICs
90. TC-Trace Vapors + TICs
91. TC-Trace Gases + TICs
92. TC-Trace Liquids + TICs
93. TC-Trace Solids + TICs
94. TC-Trace Gels + TICs
95. TC-Trace Powders + TICs
96. TC-Trace Crystals + TICs
97. TC-Trace Fibers + TICs
98. TC-Trace Particles + TICs
99. TC-Trace Dusts + TICs
100. TC-Trace Aerosols + TICs
101. TC-Trace Vapors + TICs
102. TC-Trace Gases + TICs
103. TC-Trace Liquids + TICs
104. TC-Trace Solids + TICs
105. TC-Trace Gels + TICs
106. TC-Trace Powders + TICs
107. TC-Trace Crystals + TICs
108. TC-Trace Fibers + TICs
109. TC-Trace Particles + TICs
110. TC-Trace Dusts + TICs
111. TC-Trace Aerosols + TICs
112. TC-Trace Vapors + TICs
113. TC-Trace Gases + TICs
114. TC-Trace Liquids + TICs
115. TC-Trace Solids + TICs
116. TC-Trace Gels + TICs
117. TC-Trace Powders + TICs
118. TC-Trace Crystals + TICs
119. TC-Trace Fibers + TICs
120. TC-Trace Particles + TICs
121. TC-Trace Dusts + TICs
122. TC-Trace Aerosols + TICs
123. TC-Trace Vapors + TICs
124. TC-Trace Gases + TICs
125. TC-Trace Liquids + TICs
126. TC-Trace Solids + TICs
127. TC-Trace Gels + TICs
128. TC-Trace Powders + TICs
129. TC-Trace Crystals + TICs
130. TC-Trace Fibers + TICs
131. TC-Trace Particles + TICs
132. TC-Trace Dusts + TICs
133. TC-Trace Aerosols + TICs
134. TC-Trace Vapors + TICs
135. TC-Trace Gases + TICs
136. TC-Trace Liquids + TICs
137. TC-Trace Solids + TICs
138. TC-Trace Gels + TICs
139. TC-Trace Powders + TICs
140. TC-Trace Crystals + TICs
141. TC-Trace Fibers + TICs
142. TC-Trace Particles + TICs
143. TC-Trace Dusts + TICs
144. TC-Trace Aerosols + TICs
145. TC-Trace Vapors + TICs
146. TC-Trace Gases + TICs
147. TC-Trace Liquids + TICs
148. TC-Trace Solids + TICs
149. TC-Trace Gels + TICs
150. TC-Trace Powders + TICs
151. TC-Trace Crystals + TICs
152. TC-Trace Fibers + TICs
153. TC-Trace Particles + TICs
154. TC-Trace Dusts + TICs
155. TC-Trace Aerosols + TICs
156. TC-Trace Vapors + TICs
157. TC-Trace Gases + TICs
158. TC-Trace Liquids + TICs
159. TC-Trace Solids + TICs
160. TC-Trace Gels + TICs
161. TC-Trace Powders + TICs
162. TC-Trace Crystals + TICs
163. TC-Trace Fibers + TICs
164. TC-Trace Particles + TICs
165. TC-Trace Dusts + TICs
166. TC-Trace Aerosols + TICs
167. TC-Trace Vapors + TICs
168. TC-Trace Gases + TICs
169. TC-Trace Liquids + TICs
170. TC-Trace Solids + TICs
171. TC-Trace Gels + TICs
172. TC-Trace Powders + TICs
173. TC-Trace Crystals + TICs
174. TC-Trace Fibers + TICs
175. TC-Trace Particles + TICs
176. TC-Trace Dusts + TICs
177. TC-Trace Aerosols + TICs
178. TC-Trace Vapors + TICs
179. TC-Trace Gases + TICs
180. TC-Trace Liquids + TICs
181. TC-Trace Solids + TICs
182. TC-Trace Gels + TICs
183. TC-Trace Powders + TICs
184. TC-Trace Crystals + TICs
185. TC-Trace Fibers + TICs
186. TC-Trace Particles + TICs
187. TC-Trace Dusts + TICs
188. TC-Trace Aerosols + TICs
189. TC-Trace Vapors + TICs
190. TC-Trace Gases + TICs
191. TC-Trace Liquids + TICs
192. TC-Trace Solids + TICs
193. TC-Trace Gels + TICs
194. TC-Trace Powders + TICs
195. TC-Trace Crystals + TICs
196. TC-Trace Fibers + TICs
197. TC-Trace Particles + TICs
198. TC-Trace Dusts + TICs
199. TC-Trace Aerosols + TICs
200. TC-Trace Vapors + TICs
201. TC-Trace Gases + TICs
202. TC-Trace Liquids + TICs
203. TC-Trace Solids + TICs
204. TC-Trace Gels + TICs
205. TC-Trace Powders + TICs
206. TC-Trace Crystals + TICs
207. TC-Trace Fibers + TICs
208. TC-Trace Particles + TICs
209. TC-Trace Dusts + TICs
210. TC-Trace Aerosols + TICs
211. TC-Trace Vapors + TICs
212. TC-Trace Gases + TICs
213. TC-Trace Liquids + TICs
214. TC-Trace Solids + TICs
215. TC-Trace Gels + TICs
216. TC-Trace Powders + TICs
217. TC-Trace Crystals + TICs
218. TC-Trace Fibers + TICs
219. TC-Trace Particles + TICs
220. TC-Trace Dusts + TICs
221. TC-Trace Aerosols + TICs
222. TC-Trace Vapors + TICs
223. TC-Trace Gases + TICs
224. TC-Trace Liquids + TICs
225. TC-Trace Solids + TICs
226. TC-Trace Gels + TICs
227. TC-Trace Powders + TICs
228. TC-Trace Crystals + TICs
229. TC-Trace Fibers + TICs
230. TC-Trace Particles + TICs
231. TC-Trace Dusts + TICs
232. TC-Trace Aerosols + TICs
233. TC-Trace Vapors + TICs
234. TC-Trace Gases + TICs
235. TC-Trace Liquids + TICs
236. TC-Trace Solids + TICs
237. TC-Trace Gels + TICs
238. TC-Trace Powders + TICs
239. TC-Trace Crystals + TICs
240. TC-Trace Fibers + TICs
241. TC-Trace Particles + TICs
242. TC-Trace Dusts + TICs
243. TC-Trace Aerosols + TICs
244. TC-Trace Vapors + TICs
245. TC-Trace Gases + TICs
246. TC-Trace Liquids + TICs
247. TC-Trace Solids + TICs
248. TC-Trace Gels + TICs
249. TC-Trace Powders + TICs
250. TC-Trace Crystals + TICs
251. TC-Trace Fibers + TICs
252. TC-Trace Particles + TICs
253. TC-Trace Dusts + TICs
254. TC-Trace Aerosols + TICs
255. TC-Trace Vapors + TICs
256. TC-Trace Gases + TICs
257. TC-Trace Liquids + TICs
258. TC-Trace Solids + TICs
259. TC-Trace Gels + TICs
260. TC-Trace Powders + TICs
261. TC-Trace Crystals + TICs
262. TC-Trace Fibers + TICs
263. TC-Trace Particles + TICs
264. TC-Trace Dusts + TICs
265. TC-Trace Aerosols + TICs
266. TC-Trace Vapors + TICs
267. TC-Trace Gases + TICs
268. TC-Trace Liquids + TICs
269. TC-Trace Solids + TICs
270. TC-Trace Gels + TICs
271. TC-Trace Powders + TICs
272. TC-Trace Crystals + TICs
273. TC-Trace Fibers + TICs
274. TC-Trace Particles + TICs
275. TC-Trace Dusts + TICs
276. TC-Trace Aerosols + TICs
277. TC-Trace Vapors + TICs
278. TC-Trace Gases + TICs
279. TC-Trace Liquids + TICs
280. TC-Trace Solids + TICs
281. TC-Trace Gels + TICs
282. TC-Trace Powders + TICs
283. TC-Trace Crystals + TICs
284. TC-Trace Fibers + TICs
285. TC-Trace Particles + TICs
286. TC-Trace Dusts + TICs
287. TC-Trace Aerosols + TICs
288. TC-Trace Vapors + TICs
289. TC-Trace Gases + TICs
290. TC-Trace Liquids + TICs
291. TC-Trace Solids + TICs
292. TC-Trace Gels + TICs
293. TC-Trace Powders + TICs
294. TC-Trace Crystals + TICs
295. TC-Trace Fibers + TICs
296. TC-Trace Particles + TICs
297. TC-Trace Dusts + TICs
298. TC-Trace Aerosols + TICs
299. TC-Trace Vapors + TICs
300. TC-Trace Gases + TICs
301. TC-Trace Liquids + TICs
302. TC-Trace Solids + TICs
303. TC-Trace Gels + TICs
304. TC-Trace Powders + TICs
305. TC-Trace Crystals + TICs
306. TC-Trace Fibers + TICs
307. TC-Trace Particles + TICs
308. TC-Trace Dusts + TICs
309. TC-Trace Aerosols + TICs
310. TC-Trace Vapors + TICs
311. TC-Trace Gases + TICs
312. TC-Trace Liquids + TICs
313. TC-Trace Solids + TICs
314. TC-Trace Gels + TICs
315. TC-Trace Powders + TICs
316. TC-Trace Crystals + TICs
317. TC-Trace Fibers + TICs
318. TC-Trace Particles + TICs
319. TC-Trace Dusts + TICs
320. TC-Trace Aerosols + TICs
321. TC-Trace Vapors + TICs
322. TC-Trace Gases + TICs
323. TC-Trace Liquids + TICs
324. TC-Trace Solids + TICs
325. TC-Trace Gels + TICs
326. TC-Trace Powders + TICs
327. TC-Trace Crystals + TICs
328. TC-Trace Fibers + TICs
329. TC-Trace Particles + TICs
330. TC-Trace Dusts + TICs
331. TC-Trace Aerosols + TICs
332. TC-Trace Vapors + TICs
333. TC-Trace Gases + TICs
334. TC-Trace Liquids + TICs
335. TC-Trace Solids + TICs
336. TC-Trace Gels + TICs
337. TC-Trace Powders + TICs
338. TC-Trace Crystals + TICs
339. TC-Trace Fibers + TICs
340. TC-Trace Particles + TICs
341. TC-Trace Dusts + TICs
342. TC-Trace Aerosols + TICs
343. TC-Trace Vapors + TICs
344. TC-Trace Gases + TICs
345. TC-Trace Liquids + TICs
346. TC-Trace Solids + TICs
347. TC-Trace Gels + TICs
348. TC-Trace Powders + TICs
349. TC-Trace Crystals + TICs
350. TC-Trace Fibers + TICs
351. TC-Trace Particles + TICs
352. TC-Trace Dusts + TICs
353. TC-Trace Aerosols + TICs
354. TC-Trace Vapors + TICs
355. TC-Trace Gases + TICs
356. TC-Trace Liquids + TICs
357. TC-Trace Solids + TICs
358. TC-Trace Gels + TICs
359. TC-Trace Powders + TICs
360. TC-Trace Crystals + TICs
361. TC-Trace Fibers + TICs
362. TC-Trace Particles + TICs
363. TC-Trace Dusts + TICs
364. TC-Trace Aerosols + TICs
365. TC-Trace Vapors + TICs
366. TC-Trace Gases + TICs
367. TC-Trace Liquids + TICs
368. TC-Trace Solids + TICs
369. TC-Trace Gels + TICs
370. TC-Trace Powders + TICs
371. TC-Trace Crystals + TICs
372. TC-Trace Fibers + TICs
373. TC-Trace Particles + TICs
374. TC-Trace Dusts + TICs
375. TC-Trace Aerosols + TICs
376. TC-Trace Vapors + TICs
377. TC-Trace Gases + TICs
378. TC-Trace Liquids + TICs
379. TC-Trace Solids + TICs
380. TC-Trace Gels + TICs
381. TC-Trace Powders + TICs
382. TC-Trace Crystals + TICs
383. TC-Trace Fibers + TICs
384. TC-Trace Particles + TICs
385. TC-Trace Dusts + TICs
386. TC-Trace Aerosols + TICs
387. TC-Trace Vapors + TICs
388. TC-Trace Gases + TICs
389. TC-Trace Liquids + TICs
390. TC-Trace Solids + TICs
391. TC-Trace Gels + TICs
392. TC-Trace Powders + TICs
393. TC-Trace Crystals + TICs
394. TC-Trace Fibers + TICs
395. TC-Trace Particles + TICs
396. TC-Trace Dusts + TICs
397. TC-Trace Aerosols + TICs
398. TC-Trace Vapors + TICs
399. TC-Trace Gases + TICs
400. TC-Trace Liquids + TICs
401. TC-Trace Solids + TICs
402. TC-Trace Gels + TICs
403. TC-Trace Powders + TICs
404. TC-Trace Crystals + TICs
405. TC-Trace Fibers + TICs
406. TC-Trace Particles + TICs
407. TC-Trace Dusts + TICs
408. TC-Trace Aerosols + TICs
409. TC-Trace Vapors + TICs
410. TC-Trace Gases + TICs
411. TC-Trace Liquids + TICs
412. TC-Trace Solids + TICs
413. TC-Trace Gels + TICs
414. TC-Trace Powders + TICs
415. TC-Trace Crystals + TICs
416. TC-Trace Fibers + TICs
417. TC-Trace Particles + TICs
418. TC-Trace Dusts + TICs
419. TC-Trace Aerosols + TICs
420. TC-Trace Vapors + TICs
421. TC-Trace Gases + TICs
422. TC-Trace Liquids + TICs
423. TC-Trace Solids + TICs
424. TC-Trace Gels + TICs
425. TC-Trace Powders + TICs
426. TC-Trace Crystals + TICs
427. TC-Trace Fibers + TICs
428. TC-Trace Particles + TICs
429. TC-Trace Dusts + TICs
430. TC-Trace Aerosols + TICs
431. TC-Trace Vapors + TICs
432. TC-Trace Gases + TICs
433. TC-Trace Liquids + TICs
434. TC-Trace Solids + TICs
435. TC-Trace Gels + TICs
436. TC-Trace Powders + TICs
437. TC-Trace Crystals + TICs
438. TC-Trace Fibers + TICs
439. TC-Trace Particles + TICs
440. TC-Trace Dusts + TICs
441. TC-Trace Aerosols + TICs
442. TC-Trace Vapors + TICs
443. TC-Trace Gases + TICs
444. TC-Trace Liquids + TICs
445. TC-Trace Solids + TICs
446. TC-Trace Gels + TICs
447. TC-Trace Powders + TICs
448. TC-Trace Crystals + TICs
449. TC-Trace Fibers + TICs
450. TC-Trace Particles + TICs
451. TC-Trace Dusts + TICs
452. TC-Trace Aerosols + TICs
453. TC-Trace Vapors + TICs
454. TC-Trace Gases + TICs
455. TC-Trace Liquids + TICs
456. TC-Trace Solids + TICs
457. TC-Trace Gels + TICs
458. TC-Trace Powders + TICs
459. TC-Trace Crystals + TICs
460. TC-Trace Fibers + TICs
461. TC-Trace Particles + TICs
462. TC-Trace Dusts + TICs
463. TC-Trace Aerosols + TICs
464. TC-Trace Vapors + TICs
465. TC-Trace Gases + TICs
466. TC-Trace Liquids + TICs
467. TC-Trace Solids + TICs
468. TC-Trace Gels + TICs
469. TC-Trace Powders + TICs
470. TC-Trace Crystals + TICs
471. TC-Trace Fibers + TICs
472. TC-Trace Particles + TICs
473. TC-Trace Dusts + TICs
474. TC-Trace Aerosols + TICs
475. TC-Trace Vapors + TICs
476. TC-Trace Gases + TICs
477. TC-Trace Liquids + TICs
478. TC-Trace Solids + TICs
479. TC-Trace Gels + TICs
480. TC-Trace Powders + TICs
481. TC-Trace Crystals + TICs
482. TC-Trace Fibers + TICs
483. TC-Trace Particles + TICs
484. TC-Trace Dusts + TICs
485. TC-Trace Aerosols + TICs
486. TC-Trace Vapors + TICs
487. TC-Trace Gases + TICs
488. TC-Trace Liquids + TICs
489. TC-Trace Solids + TICs
490. TC-Trace Gels + TICs
491. TC-Trace Powders + TICs
492. TC-Trace Crystals + TICs
493. TC-Trace Fibers + TICs
494. TC-Trace Particles + TICs
495. TC-Trace Dusts + TICs
496. TC-Trace Aerosols + TICs
497. TC-Trace Vapors + TICs
498. TC-Trace Gases + TICs
499. TC-Trace Liquids + TICs
500. TC-Trace Solids + TICs
501. TC-Trace Gels + TICs
502. TC-Trace Powders + TICs
503. TC-Trace Crystals + TICs
504. TC-Trace Fibers + TICs
505. TC-Trace Particles + TICs
506. TC-Trace Dusts + TICs
507. TC-Trace Aerosols + TICs
508. TC-Trace Vapors + TICs
509. TC-Trace Gases + TICs
510. TC-Trace Liquids + TICs
511. TC-Trace Solids + TICs
512. TC-Trace Gels + TICs
513. TC-Trace Powders + TICs
514. TC-Trace Crystals + TICs
515. TC-Trace Fibers + TICs
516. TC-Trace Particles + TICs
517. TC-Trace Dusts + TICs
518. TC-Trace Aerosols + TICs
519. TC-Trace Vapors + TICs
520. TC-Trace Gases + TICs
521. TC-Trace Liquids + TICs
522. TC-Trace Solids + TICs
523. TC-Trace Gels + TICs
524. TC-Trace Powders + TICs
525. TC-Trace Crystals + TICs
526. TC-Trace Fibers + TICs
527. TC-Trace Particles + TICs
528. TC-Trace Dusts + TICs
529. TC-Trace Aerosols + TICs
530. TC-Trace Vapors + TICs
531. TC-Trace Gases + TICs
532. TC-Trace Liquids + TICs
533. TC-Trace Solids + TICs
534. TC-Trace Gels + TICs
535. TC-Trace Powders + TICs
536. TC-Trace Crystals + TICs
537. TC-Trace Fibers + TICs
538. TC-Trace Particles + TICs
539. TC-Trace Dusts + TICs
540. TC-Trace Aerosols + TICs
541. TC-Trace Vapors + TICs
542. TC-Trace Gases + TICs
543. TC-Trace Liquids + TICs
544. TC-Trace Solids + TICs
545. TC-Trace Gels + TICs
546. TC-Trace Powders + TICs
547. TC-Trace Crystals + TICs
548. TC-Trace Fibers + TICs
549. TC-Trace Particles + TICs
550. TC-Trace Dusts + TICs
551. TC-Trace Aerosols + TICs
552. TC-Trace Vapors + TICs
553. TC-Trace Gases + TICs
554. TC-Trace Liquids + TICs
555. TC-Trace Solids + TICs
556. TC-Trace Gels + TICs
557. TC-Trace Powders + TICs
558. TC-Trace Crystals + TICs
559. TC-Trace Fibers + TICs
560. TC-Trace Particles + TICs
561

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2487 **WALS01**
Client Name : Walsh Construction Compa
Client Contact : Jesse A. Sylvestri
Invoice Name : Walsh Construction Compa
Invoice Contact : Jesse A. Sylvestri

Order Date : 7/2/2025 10:05:00 AM
Project Name : Construction of Shafts 17B-
Receive DateTime : 7/1/2025 4:00:00 PM
Purchase Order : 18:18

Project Mgr :
Report Type : Level 2
EDD Type : Excel NY
Hard Copy Date :
Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2487-01	G4(1.5)	Solid	07/01/2025	10:25	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2487-02	G4(10)	Solid	07/01/2025	10:25	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2487-03	G3(9)	Solid	07/01/2025	10:25	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2487-04	G3(3)	Solid	07/01/2025	10:25	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2487-05	G2(2.5)	Solid	07/01/2025	10:25	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2487-06	G2(9)	Solid	07/01/2025	10:25	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2487-07	G1(4.5)	Solid	07/01/2025	10:25	VOC-TCLVOA-10		8260D	10 Bus. Days	
Q2487-08	G1(10)	Solid	07/01/2025	10:25	VOC-TCLVOA-10		8260D	10 Bus. Days	



284 Sheffield Street, MountainSide, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2487 WALS01
Client Name : Walsh Construction Company
Client Contact : Jesse A. Sylvestri
Invoice Name : Walsh Construction Company
Invoice Contact : Jesse A. Sylvestri

Order Date : 7/2/2025 10:05:00 AM
Project Name : Construction of Shafts 17B-
Receive Date/Time : 7/1/2025 4:00:00 PM
Purchase Order : 18-1821

Project Mgr :
Report Type : Level 2
EDD Type : Excel NY
Hard Copy Date :
Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
					VOC-TCLVOA-10		8260D		10 Bus. Days

Relinquished By : CD

Date / Time : 7/2/25 15:45

Received By : Zam

Date / Time : 07/02/25 15:45

Storage Area : VOA Refridgerator Room