

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2487</u>	Analytical Method:	<u>8151A</u>
Client:	<u>Walsh Construction Company II, LLC</u>	DataFile :	<u>PS030961.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD		Low	Limits	
			Result	Result				RPD	Qual		High	RPD
Lab Sample ID:	Q2493-01MS (Column 1)	Client Sample ID:		WC-11MS								
	DICAMBA	187.5	0	137	ug/Kg	73				10	112	
	DICHLORPROP	187.5	0	144	ug/Kg	77				10	113	
	2,4-D	187.5	0	545	ug/Kg	291	*			10	144	
	2,4,5-TP(Silvex)	187.5	0	143	ug/Kg	76				10	114	
	2,4,5-T	187.5	0	155	ug/Kg	83				10	115	
	2,4-DB	187.5	0	134	ug/Kg	71				10	140	
	Dinoseb	187.5	0	0	ug/Kg	0	*			10	118	
Lab Sample ID:	Q2493-01MS (Column 2)	Client Sample ID:		WC-11MS								
	DICAMBA	187.5	0	133	ug/Kg	71				10	112	
	DICHLORPROP	187.5	0	142	ug/Kg	76				10	113	
	2,4-D	187.5	0	124	ug/Kg	66	*			10	144	
	2,4,5-TP(Silvex)	187.5	0	142	ug/Kg	76				10	114	
	2,4,5-T	187.5	0	140	ug/Kg	75				10	115	
	2,4-DB	187.5	0	117	ug/Kg	62				10	140	
	Dinoseb	187.5	0	0	ug/Kg	0	*			10	118	