

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2487</u>	Analytical Method:	<u>8151A</u>
Client:	<u>Walsh Construction Company II, LLC</u>	DataFile :	<u>PS030962.D</u>

		Sample				Rec		RPD		Limits		
	Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2493-01MSD (Column 1)	Client Sample ID:		WC-11MSD								
	DICAMBA	187.4	0	138	ug/Kg	74		1		10	112	20
	DICHLORPROP	187.4	0	133	ug/Kg	71		8		10	113	20
	2,4-D	187.4	0	562	ug/Kg	300	*	3		10	144	20
	2,4,5-TP(Silvex)	187.4	0	144	ug/Kg	77		1		10	114	20
	2,4,5-T	187.4	0	159	ug/Kg	85		2		10	115	20
	2,4-DB	187.4	0	126	ug/Kg	67		6		10	140	20
	Dinoseb	187.4	0	0	ug/Kg	0	*	0		10	118	20
Lab Sample ID:	Q2493-01MSD (Column 2)	Client Sample ID:		WC-11MSD								
	DICAMBA	187.4	0	133	ug/Kg	71		0		10	112	20
	DICHLORPROP	187.4	0	143	ug/Kg	76		0		10	113	20
	2,4-D	187.4	0	128	ug/Kg	68	*	3		10	144	20
	2,4,5-TP(Silvex)	187.4	0	143	ug/Kg	76		0		10	114	20
	2,4,5-T	187.4	0	142	ug/Kg	76		1		10	115	20
	2,4-DB	187.4	0	119	ug/Kg	64		3		10	140	20
	Dinoseb	187.4	0	0	ug/Kg	0	*	0		10	118	20