

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2487

Analytical Method: 8151A

Client: Walsh Construction Company II, LLC

Datafile : PS030964.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD			Limits	
						RPD	Qual	Qual	Low	High
PB168753BS (Column 1)	DICAMBA	166.6	172	ug/Kg	103				72	129
	DICHLORPROP	166.6	163	ug/Kg	98				77	135
	2,4-D	166.6	170	ug/Kg	102				65	144
	2,4,5-TP(Silvex)	166.6	176	ug/Kg	106				74	146
	2,4,5-T	166.6	185	ug/Kg	111				77	134
	2,4-DB	166.6	193	ug/Kg	116				72	122
	Dinoseb	166.6	168	ug/Kg	101				74	132
PB168753BS (Column 2)	DICAMBA	166.6	162	ug/Kg	97				72	129
	DICHLORPROP	166.6	159	ug/Kg	95				77	135
	2,4-D	166.6	153	ug/Kg	92				65	144
	2,4,5-TP(Silvex)	166.6	167	ug/Kg	100				74	146
	2,4,5-T	166.6	168	ug/Kg	101				77	134
	2,4-DB	166.6	161	ug/Kg	97				72	122
	Dinoseb	166.6	153	ug/Kg	92				74	132