

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/04/2025

Initial Calibration Date(s): 07/01/2025

07/01/2025

Continuing Calib Time: 14:58

Initial Calibration Time(s): 14:04

21:30

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1242-1       | (1) | 5.64       | 5.65      | 5.55      | 5.75  | 0.01       |
| Aroclor-1242-2       | (2) | 5.66       | 5.67      | 5.57      | 5.77  | 0.01       |
| Aroclor-1242-3       | (3) | 5.72       | 5.73      | 5.63      | 5.83  | 0.01       |
| Aroclor-1242-4       | (4) | 5.82       | 5.83      | 5.73      | 5.93  | 0.01       |
| Aroclor-1242-5       | (5) | 6.55       | 6.56      | 6.46      | 6.66  | 0.01       |
| Tetrachloro-m-xylene |     | 4.49       | 4.49      | 4.39      | 4.59  | 0.00       |
| Decachlorobiphenyl   |     | 10.18      | 10.19     | 10.09     | 10.29 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/04/2025

Initial Calibration Date(s): 07/01/2025

07/01/2025

Continuing Calib Time: 14:58

Initial Calibration Time(s): 14:04

21:30

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.86       | 4.87      | 4.77      | 4.97 | 0.01       |
| Aroclor-1242-2 (2)   | 4.88       | 4.88      | 4.78      | 4.98 | 0.00       |
| Aroclor-1242-3 (3)   | 5.05       | 5.06      | 4.96      | 5.16 | 0.01       |
| Aroclor-1242-4 (4)   | 5.14       | 5.14      | 5.04      | 5.24 | 0.00       |
| Aroclor-1242-5 (5)   | 5.66       | 5.67      | 5.57      | 5.77 | 0.01       |
| Tetrachloro-m-xylene | 3.78       | 3.79      | 3.69      | 3.89 | 0.01       |
| Decachlorobiphenyl   | 8.78       | 8.79      | 8.69      | 8.89 | 0.01       |