

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/07/2025

Initial Calibration Date(s): 07/01/2025

07/01/2025

Continuing Calib Time: 09:16

Initial Calibration Time(s): 14:04

21:30

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1242-1 (1)	5.64	5.65	5.55	5.75	0.01
Aroclor-1242-2 (2)	5.66	5.67	5.57	5.77	0.01
Aroclor-1242-3 (3)	5.72	5.73	5.63	5.83	0.01
Aroclor-1242-4 (4)	5.82	5.83	5.73	5.93	0.01
Aroclor-1242-5 (5)	6.55	6.56	6.46	6.66	0.01
Tetrachloro-m-xylene	4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl	10.17	10.19	10.09	10.29	0.02

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/07/2025

Initial Calibration Date(s): 07/01/2025

07/01/2025

Continuing Calib Time: 09:16

Initial Calibration Time(s): 14:04

21:30

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.86	4.87	4.77	4.97	0.01
Aroclor-1242-2 (2)	4.87	4.88	4.78	4.98	0.01
Aroclor-1242-3 (3)	5.05	5.06	4.96	5.16	0.01
Aroclor-1242-4 (4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1242-5 (5)	5.66	5.67	5.57	5.77	0.01
Tetrachloro-m-xylene	3.78	3.79	3.69	3.89	0.01
Decachlorobiphenyl	8.78	8.79	8.69	8.89	0.01