

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/08/2025

Initial Calibration Date(s): 07/07/2025 07/08/2025

Continuing Calib Time: 14:55

Initial Calibration Time(s): 21:03 04:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	5.64	5.64	5.54	5.74	0.00
Aroclor-1248-2	(2)	5.91	5.91	5.81	6.01	0.00
Aroclor-1248-3	(3)	6.11	6.11	6.01	6.21	0.00
Aroclor-1248-4	(4)	6.51	6.51	6.41	6.61	0.00
Aroclor-1248-5	(5)	6.55	6.55	6.45	6.65	0.00
Tetrachloro-m-xylene		4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl		10.18	10.17	10.07	10.27	-0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/08/2025

Initial Calibration Date(s): 07/07/2025 07/08/2025

Continuing Calib Time: 14:55

Initial Calibration Time(s): 21:03 04:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.86	4.86	4.76	4.96	0.00
Aroclor-1248-2	(2)	5.10	5.09	4.99	5.19	0.00
Aroclor-1248-3	(3)	5.14	5.14	5.04	5.24	0.00
Aroclor-1248-4	(4)	5.31	5.31	5.21	5.41	0.00
Aroclor-1248-5	(5)	5.70	5.70	5.60	5.80	0.00
Tetrachloro-m-xylene		3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl		8.78	8.78	8.68	8.88	0.00