

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/08/2025

Initial Calibration Date(s): 07/07/2025

07/08/2025

Continuing Calib Time: 20:24

Initial Calibration Time(s): 21:03

04:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1268-1 (1)	8.69	8.69	8.59	8.79	0.00
Aroclor-1268-2 (2)	8.79	8.79	8.69	8.89	0.00
Aroclor-1268-3 (3)	9.01	9.01	8.91	9.11	0.00
Aroclor-1268-4 (4)	9.43	9.43	9.33	9.53	0.00
Aroclor-1268-5 (5)	9.84	9.84	9.74	9.94	0.00
Tetrachloro-m-xylene	4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl	10.18	10.18	10.08	10.28	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q2487

Continuing Calib Date: 07/08/2025

Initial Calibration Date(s): 07/07/2025 07/08/2025

Continuing Calib Time: 20:24

Initial Calibration Time(s): 21:03 04:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1268-1 (1)	7.67	7.67	7.57	7.77	0.00
Aroclor-1268-2 (2)	7.74	7.74	7.64	7.84	0.00
Aroclor-1268-3 (3)	7.94	7.94	7.84	8.04	0.00
Aroclor-1268-4 (4)	8.23	8.23	8.13	8.33	0.00
Aroclor-1268-5 (5)	8.53	8.53	8.43	8.63	0.00
Tetrachloro-m-xylene	3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl	8.78	8.78	8.68	8.88	0.00