

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073536.D **Time Analyzed:** 09:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.638	5.544	5.744	507.550	500.000	1.5
Aroclor-1016-2	5.660	5.566	5.766	523.570	500.000	4.7
Aroclor-1016-3	5.722	5.629	5.829	501.790	500.000	0.4
Aroclor-1016-4	5.819	5.726	5.926	500.940	500.000	0.2
Aroclor-1016-5	6.111	6.018	6.218	504.930	500.000	1.0
Aroclor-1260-1	7.228	7.134	7.334	466.110	500.000	-6.8
Aroclor-1260-2	7.482	7.389	7.589	477.030	500.000	-4.6
Aroclor-1260-3	7.840	7.746	7.946	467.830	500.000	-6.4
Aroclor-1260-4	8.064	7.971	8.171	441.050	500.000	-11.8
Aroclor-1260-5	8.381	8.289	8.489	439.440	500.000	-12.1
Decachlorobiphenyl	10.176	10.086	10.286	45.070	50.000	-9.9
Tetrachloro-m-xylene	4.488	4.393	4.593	51.640	50.000	3.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073536.D **Time Analyzed:** 09:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.857	4.764	4.964	512.540	500.000	2.5
Aroclor-1016-2	4.875	4.782	4.982	521.980	500.000	4.4
Aroclor-1016-3	5.051	4.958	5.158	535.360	500.000	7.1
Aroclor-1016-4	5.093	5.000	5.200	535.980	500.000	7.2
Aroclor-1016-5	5.306	5.214	5.414	540.480	500.000	8.1
Aroclor-1260-1	6.336	6.244	6.444	498.020	500.000	-0.4
Aroclor-1260-2	6.525	6.433	6.633	469.310	500.000	-6.1
Aroclor-1260-3	6.677	6.585	6.785	491.890	500.000	-1.6
Aroclor-1260-4	7.145	7.054	7.254	455.460	500.000	-8.9
Aroclor-1260-5	7.388	7.297	7.497	460.660	500.000	-7.9
Decachlorobiphenyl	8.779	8.690	8.890	44.680	50.000	-10.6
Tetrachloro-m-xylene	3.779	3.685	3.885	48.200	50.000	-3.6