

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073549.D **Time Analyzed:** 14:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.644	5.544	5.744	468.330	500.000	-6.3
Aroclor-1016-2	5.665	5.566	5.766	482.830	500.000	-3.4
Aroclor-1016-3	5.727	5.629	5.829	466.660	500.000	-6.7
Aroclor-1016-4	5.825	5.726	5.926	469.690	500.000	-6.1
Aroclor-1016-5	6.117	6.018	6.218	460.050	500.000	-8.0
Aroclor-1260-1	7.234	7.134	7.334	427.660	500.000	-14.5
Aroclor-1260-2	7.487	7.389	7.589	426.160	500.000	-14.8
Aroclor-1260-3	7.846	7.746	7.946	415.900	500.000	-16.8
Aroclor-1260-4	8.068	7.971	8.171	418.970	500.000	-16.2
Aroclor-1260-5	8.385	8.289	8.489	394.800	500.000	-21.0
Decachlorobiphenyl	10.183	10.086	10.286	40.670	50.000	-18.7
Tetrachloro-m-xylene	4.493	4.393	4.593	47.700	50.000	-4.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/07/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073549.D **Time Analyzed:** 14:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.862	4.764	4.964	584.980	500.000	17.0
Aroclor-1016-2	4.879	4.782	4.982	598.000	500.000	19.6
Aroclor-1016-3	5.056	4.958	5.158	609.830	500.000	22.0
Aroclor-1016-4	5.098	5.000	5.200	611.310	500.000	22.3
Aroclor-1016-5	5.311	5.214	5.414	612.950	500.000	22.6
Aroclor-1260-1	6.341	6.244	6.444	573.050	500.000	14.6
Aroclor-1260-2	6.530	6.433	6.633	529.440	500.000	5.9
Aroclor-1260-3	6.682	6.585	6.785	574.180	500.000	14.8
Aroclor-1260-4	7.151	7.054	7.254	522.220	500.000	4.4
Aroclor-1260-5	7.393	7.297	7.497	521.380	500.000	4.3
Decachlorobiphenyl	8.786	8.690	8.890	50.840	50.000	1.7
Tetrachloro-m-xylene	3.783	3.685	3.885	57.730	50.000	15.5