

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073589.D **Time Analyzed:** 09:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.638	5.539	5.739	524.700	500.000	4.9
Aroclor-1016-2	5.659	5.560	5.760	536.300	500.000	7.3
Aroclor-1016-3	5.721	5.622	5.822	535.870	500.000	7.2
Aroclor-1016-4	5.819	5.720	5.920	544.880	500.000	9.0
Aroclor-1016-5	6.111	6.012	6.212	568.730	500.000	13.7
Aroclor-1260-1	7.228	7.129	7.329	545.040	500.000	9.0
Aroclor-1260-2	7.481	7.383	7.583	476.820	500.000	-4.6
Aroclor-1260-3	7.840	7.740	7.940	543.140	500.000	8.6
Aroclor-1260-4	8.063	7.964	8.164	551.350	500.000	10.3
Aroclor-1260-5	8.381	8.283	8.483	558.150	500.000	11.6
Decachlorobiphenyl	10.175	10.076	10.276	57.050	50.000	14.1
Tetrachloro-m-xylene	4.487	4.388	4.588	55.220	50.000	10.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073589.D **Time Analyzed:** 09:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.857	4.758	4.958	520.400	500.000	4.1
Aroclor-1016-2	4.874	4.775	4.975	521.200	500.000	4.2
Aroclor-1016-3	5.050	4.951	5.151	528.680	500.000	5.7
Aroclor-1016-4	5.092	4.993	5.193	521.660	500.000	4.3
Aroclor-1016-5	5.306	5.207	5.407	528.720	500.000	5.7
Aroclor-1260-1	6.335	6.237	6.437	518.210	500.000	3.6
Aroclor-1260-2	6.524	6.425	6.625	509.780	500.000	2.0
Aroclor-1260-3	6.676	6.577	6.777	521.140	500.000	4.2
Aroclor-1260-4	7.145	7.046	7.246	537.420	500.000	7.5
Aroclor-1260-5	7.387	7.289	7.489	533.110	500.000	6.6
Decachlorobiphenyl	8.779	8.680	8.880	58.250	50.000	16.5
Tetrachloro-m-xylene	3.778	3.678	3.878	52.350	50.000	4.7