

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073605.D **Time Analyzed:** 14:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.639	5.539	5.739	483.260	500.000	-3.3
Aroclor-1016-2	5.660	5.560	5.760	491.850	500.000	-1.6
Aroclor-1016-3	5.722	5.622	5.822	486.530	500.000	-2.7
Aroclor-1016-4	5.820	5.720	5.920	495.970	500.000	-0.8
Aroclor-1016-5	6.112	6.012	6.212	511.800	500.000	2.4
Aroclor-1260-1	7.229	7.129	7.329	489.510	500.000	-2.1
Aroclor-1260-2	7.483	7.383	7.583	428.670	500.000	-14.3
Aroclor-1260-3	7.840	7.740	7.940	480.540	500.000	-3.9
Aroclor-1260-4	8.064	7.964	8.164	484.260	500.000	-3.1
Aroclor-1260-5	8.381	8.283	8.483	497.720	500.000	-0.5
Decachlorobiphenyl	10.175	10.076	10.276	49.880	50.000	-0.2
Tetrachloro-m-xylene	4.488	4.388	4.588	52.600	50.000	5.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073605.D **Time Analyzed:** 14:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.859	4.758	4.958	525.770	500.000	5.2
Aroclor-1016-2	4.876	4.775	4.975	515.420	500.000	3.1
Aroclor-1016-3	5.053	4.951	5.151	521.150	500.000	4.2
Aroclor-1016-4	5.095	4.993	5.193	513.620	500.000	2.7
Aroclor-1016-5	5.309	5.207	5.407	562.240	500.000	12.4
Aroclor-1260-1	6.338	6.237	6.437	492.220	500.000	-1.6
Aroclor-1260-2	6.527	6.425	6.625	490.070	500.000	-2.0
Aroclor-1260-3	6.678	6.577	6.777	496.110	500.000	-0.8
Aroclor-1260-4	7.147	7.046	7.246	497.210	500.000	-0.6
Aroclor-1260-5	7.390	7.289	7.489	490.430	500.000	-1.9
Decachlorobiphenyl	8.781	8.680	8.880	55.810	50.000	11.6
Tetrachloro-m-xylene	3.781	3.678	3.878	53.410	50.000	6.8