

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070125

Review By	yogesh	Review On	7/1/2025 1:43:36 PM
Supervise By	mohammad	Supervise On	7/2/2025 7:44:45 AM
SubDirectory	PP070125	HP Acquire Method	HP Processing Method PP070125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP073411.D	01 Jul 2025 13:30		YP\AJ	Ok
2	I.BLK	I.BLK	PP073412.D	01 Jul 2025 13:46		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP073413.D	01 Jul 2025 14:04		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP073414.D	01 Jul 2025 14:21		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP073415.D	01 Jul 2025 14:37		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP073416.D	01 Jul 2025 14:54		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP073417.D	01 Jul 2025 15:10		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP073418.D	01 Jul 2025 15:26		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP073419.D	01 Jul 2025 15:43		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP073420.D	01 Jul 2025 15:59		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PP073421.D	01 Jul 2025 16:16		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PP073422.D	01 Jul 2025 16:33		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PP073423.D	01 Jul 2025 16:49		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PP073424.D	01 Jul 2025 17:05		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP073425.D	01 Jul 2025 17:22		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PP073426.D	01 Jul 2025 17:39		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PP073427.D	01 Jul 2025 17:55		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PP073428.D	01 Jul 2025 18:12		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070125

Review By	yogesh	Review On	7/1/2025 1:43:36 PM
Supervise By	mohammad	Supervise On	7/2/2025 7:44:45 AM
SubDirectory	PP070125	HP Acquire Method	HP Processing Method PP070125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP073429.D	01 Jul 2025 18:28		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP073430.D	01 Jul 2025 18:45		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PP073431.D	01 Jul 2025 19:01		YPIAJ	Ok
22	AR1254ICC500	AR1254ICC500	PP073432.D	01 Jul 2025 19:18		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP073433.D	01 Jul 2025 19:35		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP073434.D	01 Jul 2025 19:51		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP073435.D	01 Jul 2025 20:08		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP073436.D	01 Jul 2025 20:24		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP073437.D	01 Jul 2025 20:41		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP073438.D	01 Jul 2025 20:57		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP073439.D	01 Jul 2025 21:14		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP073440.D	01 Jul 2025 21:30		YPIAJ	Ok,M
31	PP070125ICV500	ICVPP070125	PP073441.D	01 Jul 2025 21:47		YPIAJ	Ok
32	AR1242ICV500	ICVPP070125AR1242	PP073442.D	01 Jul 2025 22:20		YPIAJ	Ok
33	AR1248ICV500	ICVPP070125AR1248	PP073443.D	01 Jul 2025 22:53		YPIAJ	Ok
34	AR1254ICV500	ICVPP070125AR1254	PP073444.D	01 Jul 2025 23:26		YPIAJ	Ok
35	AR1268ICV500	ICVPP070125AR1268	PP073445.D	02 Jul 2025 00:00		YPIAJ	Ok
36	DDT ANALOGUE	DDT ANALOGUE	PP073446.D	02 Jul 2025 07:34		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070425

Review By	yogesh	Review On	7/7/2025 1:32:01 PM
Supervise By	mohammad	Supervise On	7/8/2025 9:14:39 AM
SubDirectory	PP070425	HP Acquire Method	HP Processing Method PP070125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP073503.D	04 Jul 2025 03:01		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP073504.D	04 Jul 2025 03:34		YP\AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP073505.D	04 Jul 2025 04:06		YP\AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PP073506.D	04 Jul 2025 04:23		YP\AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PP073507.D	04 Jul 2025 04:39		YP\AJ	Ok,M
6	I.BLK	I.BLK	PP073508.D	04 Jul 2025 04:55		YP\AJ	Ok,M
7	PB168724BL	PB168724BL	PP073509.D	04 Jul 2025 05:12		YP\AJ	Ok
8	PB168724BS	PB168724BS	PP073510.D	04 Jul 2025 05:28		YP\AJ	Ok,M
9	Q2500-01MS	X600-B2MS	PP073511.D	04 Jul 2025 05:44		YP\AJ	Ok,M
10	Q2500-01MSD	X600-B2MSD	PP073512.D	04 Jul 2025 06:01		YP\AJ	Ok
11	Q2487-09	G4(0-6)	PP073513.D	04 Jul 2025 06:17	AR1254 ccc fail, AR1260+1254 Hit	YP\AJ	Not Ok
12	Q2487-10	G4(6-12)	PP073514.D	04 Jul 2025 06:33		YP\AJ	Ok,M
13	Q2487-11	G3(0-6)	PP073515.D	04 Jul 2025 06:50	AR1260 Hit	YP\AJ	Ok,M
14	Q2487-12	G3(6-12)	PP073516.D	04 Jul 2025 07:06		YP\AJ	Ok
15	Q2487-13	G2(0-6)	PP073517.D	04 Jul 2025 07:22	AR1260 Hit	YP\AJ	Ok,M
16	Q2487-14	G2(6-12)	PP073518.D	04 Jul 2025 07:38		YP\AJ	Ok
17	Q2487-15	G1(0-6)	PP073519.D	04 Jul 2025 07:55	AR1254+1268 Hit , AR1254 CCC fail	YP\AJ	Not Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070425

Review By	yogesh	Review On	7/7/2025 1:32:01 PM
Supervise By	mohammad	Supervise On	7/8/2025 9:14:39 AM
SubDirectory	PP070425	HP Acquire Method	HP Processing Method PP070125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1660CCC500	AR1660CCC500	PP073520.D	04 Jul 2025 09:16	AR1016-5 high in 2nd column	YPIAJ	Ok,M
19	AR1242CCC500	AR1242CCC500	PP073521.D	04 Jul 2025 11:10	AR1242-1,2,3 and DCB low in 1st column	YPIAJ	Not Ok
20	AR1248CCC500	AR1248CCC500	PP073522.D	04 Jul 2025 11:26	DCB Low in 1st column	YPIAJ	Ok,M
21	AR1254CCC500	AR1254CCC500	PP073523.D	04 Jul 2025 11:42	AR1254-3,4,5 and DCB low in 1st column	YPIAJ	Not Ok
22	I.BLK	I.BLK	PP073524.D	04 Jul 2025 11:59		YPIAJ	Ok
23	Q2487-16	G1(6-12)	PP073525.D	04 Jul 2025 12:15	end ccc fail	YPIAJ	Not Ok
24	Q2491-01	EO-1-070225	PP073526.D	04 Jul 2025 12:31	end ccc fail	YPIAJ	Not Ok
25	Q2491-01MS	EO-1-070225MS	PP073527.D	04 Jul 2025 12:48	end ccc fail	YPIAJ	Not Ok
26	Q2491-01MSD	EO-1-070225MSD	PP073528.D	04 Jul 2025 13:04	end ccc fail	YPIAJ	Not Ok
27	Q2493-01	WC-11	PP073529.D	04 Jul 2025 13:20	end ccc fail	YPIAJ	Not Ok
28	AR1660CCC500	AR1660CCC500	PP073530.D	04 Jul 2025 14:41	AR1260 and DCB low in first column	YPIAJ	Not Ok
29	AR1242CCC500	AR1242CCC500	PP073531.D	04 Jul 2025 14:58	AR1242 low in 1st column	YPIAJ	Not Ok
30	AR1248CCC500	AR1248CCC500	PP073532.D	04 Jul 2025 15:14	AR1248,DCB and TCMX low in 1st column	YPIAJ	Not Ok
31	AR1254CCC500	AR1254CCC500	PP073533.D	04 Jul 2025 15:30	AR1254,DCB and TCMX low in 1st column	YPIAJ	Not Ok
32	I.BLK	I.BLK	PP073534.D	04 Jul 2025 15:47		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070725

Review By	yogesh	Review On	7/7/2025 11:19:16 AM
Supervise By		Supervise On	
SubDirectory	PP070725	HP Acquire Method	HP Processing Method PP070125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP073535.D	07 Jul 2025 08:43		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP073536.D	07 Jul 2025 09:00		YPIAJ	Ok,NS
3	AR1242CCC500	AR1242CCC500	PP073537.D	07 Jul 2025 09:16		YPIAJ	Ok,NS
4	AR1248CCC500	AR1248CCC500	PP073538.D	07 Jul 2025 09:33		YPIAJ	Ok,NS
5	AR1254CCC500	AR1254CCC500	PP073539.D	07 Jul 2025 09:49		YPIAJ	Ok,NS
6	I.BLK	I.BLK	PP073540.D	07 Jul 2025 10:05		YPIAJ	Ok
7	DDT ANALOGUE	DDT ANALOGUE	PP073541.D	07 Jul 2025 10:22		YPIAJ	Ok
8	Q2491-01	EO-1-070225	PP073542.D	07 Jul 2025 10:38		YPIAJ	Ok,NS
9	Q2491-01MS	EO-1-070225MS	PP073543.D	07 Jul 2025 10:54		YPIAJ	Ok,NS
10	Q2491-01MSD	EO-1-070225MSD	PP073544.D	07 Jul 2025 11:11		YPIAJ	Ok,NS
11	Q2493-01	WC-11	PP073545.D	07 Jul 2025 11:27		YPIAJ	Ok
12	Q2487-09	G4(0-6)	PP073546.D	07 Jul 2025 11:44	AR1260 Hit	YPIAJ	Ok,NS
13	Q2487-11	G3(0-6)	PP073547.D	07 Jul 2025 12:00	AR1260 Hit	YPIAJ	Ok,NS
14	Q2487-16	G1(6-12)	PP073548.D	07 Jul 2025 12:16		YPIAJ	Ok,NS
15	AR1660CCC500	AR1660CCC500	PP073549.D	07 Jul 2025 14:53		YPIAJ	Ok,NS
16	AR1242CCC500	AR1242CCC500	PP073550.D	07 Jul 2025 15:26		YPIAJ	Ok,NS
17	I.BLK	I.BLK	PP073551.D	07 Jul 2025 15:43		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070825

Review By	yogesh	Review On	7/7/2025 4:17:13 PM
Supervise By	mohammad	Supervise On	7/9/2025 1:51:14 AM
SubDirectory	PP070825	HP Acquire Method	HP Processing Method PP070825

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP073552.D	07 Jul 2025 20:13		YP\AJ	Ok
2	I.BLK	I.BLK	PP073553.D	07 Jul 2025 20:30		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP073554.D	07 Jul 2025 21:03		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP073555.D	07 Jul 2025 21:19		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP073556.D	07 Jul 2025 21:35		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP073557.D	07 Jul 2025 21:52		YP\AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PP073558.D	07 Jul 2025 22:08		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP073559.D	07 Jul 2025 22:24		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP073560.D	07 Jul 2025 22:41		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP073561.D	07 Jul 2025 22:57		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PP073562.D	07 Jul 2025 23:14		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PP073563.D	07 Jul 2025 23:30		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PP073564.D	07 Jul 2025 23:46		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PP073565.D	08 Jul 2025 00:03		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP073566.D	08 Jul 2025 00:19		YP\AJ	Ok
16	AR1248ICCC750	AR1248ICCC750	PP073567.D	08 Jul 2025 00:35		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PP073568.D	08 Jul 2025 00:52		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PP073569.D	08 Jul 2025 01:08		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070825

Review By	yogesh	Review On	7/7/2025 4:17:13 PM
Supervise By	mohammad	Supervise On	7/9/2025 1:51:14 AM
SubDirectory	PP070825	HP Acquire Method	HP Processing Method PP070825
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

19	AR1248ICC050	AR1248ICC050	PP073570.D	08 Jul 2025 01:25		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP073571.D	08 Jul 2025 01:41		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PP073572.D	08 Jul 2025 01:57		YPIAJ	Ok
22	AR1254ICC500	AR1254ICC500	PP073573.D	08 Jul 2025 02:14		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP073574.D	08 Jul 2025 02:30		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP073575.D	08 Jul 2025 02:46		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP073576.D	08 Jul 2025 03:03		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP073577.D	08 Jul 2025 03:19		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP073578.D	08 Jul 2025 03:35		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP073579.D	08 Jul 2025 03:52		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP073580.D	08 Jul 2025 04:08		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP073581.D	08 Jul 2025 04:24		YPIAJ	Ok,M
31	PP070125ICV500	ICVPP070825	PP073582.D	08 Jul 2025 04:41		YPIAJ	Ok
32	AR1242ICV500	ICVPP070825AR1242	PP073583.D	08 Jul 2025 05:30		YPIAJ	Ok
33	AR1248ICV500	ICVPP070825AR1248	PP073584.D	08 Jul 2025 05:46		YPIAJ	Ok
34	AR1254ICV500	ICVPP070825AR1254	PP073585.D	08 Jul 2025 06:19		YPIAJ	Ok
35	AR1268ICV500	ICVPP070825AR1268	PP073586.D	08 Jul 2025 06:52		YPIAJ	Ok
36	DDT ANALOGUE	DDT ANALOGUE	PP073587.D	08 Jul 2025 07:24		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070925

Review By	yogesh	Review On	7/8/2025 1:09:22 PM
Supervise By		Supervise On	
SubDirectory	PP070925	HP Acquire Method	HP Processing Method PP070825

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP073588.D	08 Jul 2025 08:50		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP073589.D	08 Jul 2025 09:06		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP073590.D	08 Jul 2025 09:22		YPIAJ	Ok,NS
4	AR1248CCC500	AR1248CCC500	PP073591.D	08 Jul 2025 09:39		YPIAJ	Ok
5	AR1254CCC500	AR1254CCC500	PP073592.D	08 Jul 2025 09:55		YPIAJ	Ok
6	AR1268CCC500	AR1268CCC500	PP073593.D	08 Jul 2025 10:12		YPIAJ	Ok
7	I.BLK	I.BLK	PP073594.D	08 Jul 2025 10:28		YPIAJ	Ok
8	DDT ANALOGUE	DDT ANALOGUE	PP073595.D	08 Jul 2025 10:45		YPIAJ	Ok
9	PB168734BL	PB168734BL	PP073596.D	08 Jul 2025 11:05		YPIAJ	Ok
10	Q2476-01	50731	PP073597.D	08 Jul 2025 11:22	AR1242 Hit	YPIAJ	Dilution
11	Q2476-01DL	50731DL	PP073598.D	08 Jul 2025 11:38	AR1242 Hit	YPIAJ	Ok,NS
12	PB168735BL	PB168735BL	PP073599.D	08 Jul 2025 11:55		YPIAJ	Ok
13	Q2505-01	#62825	PP073600.D	08 Jul 2025 12:11	AR1242 Hit , DCB high in 2nd column	YPIAJ	Ok,NS
14	Q2505-02	#62525	PP073601.D	08 Jul 2025 12:27	DCB high in 2nd column , 1242 Hit	YPIAJ	Ok,NS
15	Q2508-02	AUD-25-0106	PP073602.D	08 Jul 2025 12:44		YPIAJ	Ok,NS
16	Q2487-15	G1(0-6)	PP073603.D	08 Jul 2025 13:00	AR1268+1260+1254 Hit , DCB high in 1st column	YPIAJ	Ok,NS
17	Q2509-02	AUD-25-0112	PP073604.D	08 Jul 2025 13:17		YPIAJ	Ok,NS

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070925

Review By	yogesh	Review On	7/8/2025 1:09:22 PM
Supervise By		Supervise On	
SubDirectory	PP070925	HP Acquire Method	HP Processing Method PP070825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1660CCC500	AR1660CCC500	PP073605.D	08 Jul 2025 14:22		YPIAJ	Ok
19	AR1242CCC500	AR1242CCC500	PP073606.D	08 Jul 2025 14:39		YPIAJ	Ok
20	AR1248CCC500	AR1248CCC500	PP073607.D	08 Jul 2025 14:55		YPIAJ	Ok
21	AR1254CCC500	AR1254CCC500	PP073608.D	08 Jul 2025 15:12		YPIAJ	Ok
22	AR1268CCC500	AR1268CCC500	PP073609.D	08 Jul 2025 15:28		YPIAJ	Ok
23	I.BLK	I.BLK	PP073610.D	08 Jul 2025 15:45		YPIAJ	Ok
24	PB168754BL	PB168754BL	PP073611.D	08 Jul 2025 16:01		YPIAJ	Ok
25	PB168754BS	PB168754BS	PP073612.D	08 Jul 2025 16:18		YPIAJ	Ok
26	PB168754BSD	PB168754BSD	PP073613.D	08 Jul 2025 16:34		YPIAJ	Ok
27	Q2506-01	#60470	PP073614.D	08 Jul 2025 16:50		YPIAJ	Ok
28	Q2511-01	#60266	PP073615.D	08 Jul 2025 17:07		YPIAJ	Ok,NS
29	Q2506-02	#60472	PP073616.D	08 Jul 2025 17:23	AR1242 Hit	YPIAJ	Ok
30	Q2509-01	AUD-25-0110-0111	PP073617.D	08 Jul 2025 17:40	DCB Low in 2nd column	YPIAJ	Ok
31	AR1660CCC500	AR1660CCC500	PP073618.D	08 Jul 2025 18:45		YPIAJ	Ok
32	AR1242CCC500	AR1242CCC500	PP073619.D	08 Jul 2025 19:34		YPIAJ	Ok
33	AR1248CCC500	AR1248CCC500	PP073620.D	08 Jul 2025 19:51		YPIAJ	Ok
34	AR1254CCC500	AR1254CCC500	PP073621.D	08 Jul 2025 20:07		YPIAJ	Ok
35	AR1268CCC500	AR1268CCC500	PP073622.D	08 Jul 2025 20:24		YPIAJ	Ok
36	I.BLK	I.BLK	PP073623.D	08 Jul 2025 20:40		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070925

Review By	yogesh	Review On	7/8/2025 1:09:22 PM
Supervise By		Supervise On	
SubDirectory	PP070925	HP Acquire Method	HP Processing Method PP070825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

M : Manual Integration