

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2487</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Walsh Construction Company II, LLC</u>	DataFile :	<u>PP073544.D</u>

		Sample				Rec		RPD		Limits	
Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2491-01MSD	Client Sample ID:		EO-1-070225MSD							
	(Column 1)										
	AR1016	180.3	0	139	ug/kg	77	0		55	146	15
	AR1260	180.3	0	113	ug/kg	63	2		54	119	15
Lab Sample ID:	Q2491-01MSD	Client Sample ID:		EO-1-070225MSD							
	(Column 2)										
	AR1016	180.3	0	150	ug/kg	83	2		55	146	15
	AR1260	180.3	0	129	ug/kg	72	0		54	119	15