

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: WALS01
SDG NO.: Q2487

Calibration Date(s): 06/17/2025 06/17/2025
Calibration Times: 15:52 16:47

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD088993.D</u>	CF 075 =	<u>PD088994.D</u>		
CF 050 =		<u>PD088995.D</u>	CF 025 =	<u>PD088996.D</u>	CF 005 =	<u>PD088997.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	3557450000	3439230000	3414120000	3228670000	3520780000	3432050000	4
4,4'-DDE	4564130000	4401240000	4312380000	4092310000	4435110000	4361030000	4
4,4'-DDT	3928430000	3802000000	3763430000	3580670000	3866490000	3788210000	3
Aldrin	5504150000	5312640000	5254800000	4989300000	5504000000	5312980000	4
alpha-BHC	6295370000	6031790000	5874750000	5417910000	5525530000	5829070000	6
alpha-Chlordane	4885060000	4758900000	4734630000	4584390000	5229700000	4838540000	5
beta-BHC	2136750000	2100860000	2137680000	2144670000	2515730000	2207140000	8
Decachlorobiphenyl	3598890000	3628150000	3773200000	3897960000	4739940000	3927630000	12
delta-BHC	5546280000	5292010000	5182350000	4751190000	4892690000	5132900000	6
Dieldrin	4933650000	4797080000	4750270000	4527510000	4978810000	4797470000	4
Endosulfan I	4511200000	4405770000	4412600000	4289200000	4918470000	4507450000	5
Endosulfan II	3829260000	3982180000	4002080000	3816490000	4553000000	4036600000	7
Endosulfan sulfate	3786900000	3718160000	3741600000	3667440000	4271990000	3837220000	6
Endrin	4238140000	4109540000	4096430000	3878120000	4305700000	4125590000	4
Endrin aldehyde	2964260000	2935670000	2964230000	2966830000	3488720000	3063940000	8
Endrin ketone	4050100000	3970810000	4004900000	3895960000	4443590000	4073070000	5
gamma-BHC (Lindane)	5900830000	5686510000	5566590000	5221690000	5532800000	5581680000	4
gamma-Chlordane	4869330000	4801300000	4716110000	4508810000	5045020000	4788110000	4
Heptachlor	5621840000	5429540000	5358570000	5116310000	5733550000	5451960000	4
Heptachlor epoxide	4794850000	4673610000	4749650000	4554690000	5346870000	4823940000	6
Methoxychlor	1926110000	1918890000	1971480000	1982080000	2304580000	2020630000	8
Tetrachloro-m-xylene	2812300000	2762040000	2767280000	2751820000	3172230000	2853130000	6

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: WALS01
SDG NO.: Q2487

Calibration Date(s): 06/17/2025 06/17/2025
Calibration Times: 15:52 16:47

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PD088993.D</u>	CF 075 = <u>PD088994.D</u>				
CF 050 = <u>PD088995.D</u>		CF 025 = <u>PD088996.D</u>	CF 005 = <u>PD088997.D</u>				
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	17754100000	17899700000	18191100000	18805800000	22996200000	19129400000	11
4,4'-DDE	21343500000	21440200000	21823600000	22564100000	27409000000	22916100000	11
4,4'-DDT	19391000000	19327800000	19594100000	20005200000	23078500000	20279300000	8
Aldrin	22719200000	22747700000	23159400000	23852800000	28607200000	24217300000	10
alpha-BHC	25649000000	25438700000	25849200000	26153100000	30895500000	26797100000	9
alpha-Chlordane	21191300000	21145300000	21570900000	22517200000	28059300000	22896800000	13
beta-BHC	9903630000	9918550000	10198200000	10624100000	13086800000	10746200000	12
Decachlorobiphenyl	17946500000	17939700000	18428600000	19427500000	25107500000	19770000000	15
delta-BHC	23614400000	23533700000	23918700000	24298300000	29003000000	24873600000	9
Dieldrin	21438700000	21622000000	22120600000	22941800000	27919200000	23208500000	12
Endosulfan I	19170300000	19201000000	19315500000	20752900000	25772600000	20842500000	14
Endosulfan II	18473200000	18573800000	19087100000	20002300000	24846800000	20196600000	13
Endosulfan sulfate	17925600000	18043500000	18492500000	19409200000	24234500000	19621100000	13
Endrin	20535400000	19906600000	20644700000	21276500000	26206500000	21713900000	12
Endrin aldehyde	13773600000	13920200000	14379400000	15195500000	19287900000	15311300000	15
Endrin ketone	19587400000	19824900000	20446500000	21535300000	26614600000	21601700000	13
gamma-BHC (Lindane)	23622800000	23442000000	23749900000	24221500000	28748400000	24756900000	9
gamma-Chlordane	22229100000	22022500000	22519500000	23493700000	29348800000	23922700000	13
Heptachlor	23291600000	23365200000	23922200000	24716000000	29896800000	25038400000	11
Heptachlor epoxide	20080600000	20251900000	20827300000	21695300000	26635200000	21898100000	12
Methoxychlor	9712450000	9897880000	10259100000	10830900000	13140100000	10768100000	13
Tetrachloro-m-xylene	15879800000	15866400000	15978000000	16730900000	20310700000	16953200000	11