

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: WALS01

Lab Code: ACE SDG NO.: Q2487

Instrument ID: ECD_D Date(s) Analyzed: 06/17/2025 06/17/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.24	6.14	6.34	33601800
		2	6.44	6.34	6.54	47528600
		3	7.15	7.05	7.25	87292700
		4	7.56	7.46	7.66	111357000
		5	7.93	7.83	8.03	63112600

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance
Contract: WALS01

Lab Code: ACE
SDG NO.: Q2487

Instrument ID: ECD_D
Date(s) Analyzed: 06/17/2025 06/17/2025

GC Column: ZB-MR2
ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.47	5.37	5.57	158160000
		2	5.65	5.55	5.75	108820000
		3	6.76	6.66	6.86	511856000
		4	7.20	7.10	7.30	350145000
		5	7.33	7.23	7.43	253920000