

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/03/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089328.D **Time Analyzed:** 09:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.705	6.605	6.805	54.350	50.000	8.7
4,4'-DDE	6.196	6.096	6.296	51.990	50.000	4.0
4,4'-DDT	7.021	6.921	7.121	43.940	50.000	-12.1
Aldrin	5.271	5.171	5.371	55.520	50.000	11.0
alpha-BHC	4.000	3.899	4.099	56.340	50.000	12.7
alpha-Chlordane	6.027	5.927	6.127	54.210	50.000	8.4
beta-BHC	4.516	4.415	4.615	53.430	50.000	6.9
Decachlorobiphenyl	9.073	8.972	9.172	46.920	50.000	-6.2
delta-BHC	4.765	4.664	4.864	58.070	50.000	16.1
Dieldrin	6.347	6.247	6.447	54.110	50.000	8.2
Endosulfan I	6.075	5.975	6.175	54.180	50.000	8.4
Endosulfan II	6.786	6.686	6.886	51.490	50.000	3.0
Endosulfan sulfate	7.150	7.049	7.249	49.420	50.000	-1.2
Endrin	6.574	6.475	6.675	49.350	50.000	-1.3
Endrin aldehyde	6.915	6.815	7.015	49.610	50.000	-0.8
Endrin ketone	7.630	7.530	7.730	49.570	50.000	-0.9
gamma-BHC (Lindane)	4.331	4.230	4.430	55.920	50.000	11.8
gamma-Chlordane	5.946	5.846	6.046	55.190	50.000	10.4
Heptachlor	4.930	4.829	5.029	52.740	50.000	5.5
Heptachlor epoxide	5.691	5.591	5.791	53.220	50.000	6.4
Methoxychlor	7.493	7.393	7.593	49.780	50.000	-0.4
Tetrachloro-m-xylene	3.550	3.450	3.650	55.630	50.000	11.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/03/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089328.D **Time Analyzed:** 09:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.929	5.830	6.030	52.750	50.000	5.5
4,4'-DDE	5.374	5.275	5.475	52.830	50.000	5.7
4,4'-DDT	6.183	6.084	6.284	44.690	50.000	-10.6
Aldrin	4.368	4.269	4.469	54.130	50.000	8.3
alpha-BHC	3.393	3.293	3.493	55.810	50.000	11.6
alpha-Chlordane	5.189	5.091	5.291	53.210	50.000	6.4
beta-BHC	4.025	3.926	4.126	51.100	50.000	2.2
Decachlorobiphenyl	8.072	7.972	8.172	44.450	50.000	-11.1
delta-BHC	4.261	4.162	4.362	53.310	50.000	6.6
Dieldrin	5.511	5.413	5.613	53.310	50.000	6.6
Endosulfan I	5.246	5.147	5.347	53.080	50.000	6.2
Endosulfan II	6.080	5.981	6.181	51.600	50.000	3.2
Endosulfan sulfate	6.482	6.383	6.583	49.050	50.000	-1.9
Endrin	5.789	5.689	5.889	54.650	50.000	9.3
Endrin aldehyde	6.258	6.159	6.359	49.950	50.000	-0.1
Endrin ketone	6.990	6.892	7.092	48.890	50.000	-2.2
gamma-BHC (Lindane)	3.728	3.630	3.830	54.280	50.000	8.6
gamma-Chlordane	5.123	5.026	5.226	52.410	50.000	4.8
Heptachlor	4.082	3.983	4.183	50.420	50.000	0.8
Heptachlor epoxide	4.872	4.773	4.973	56.110	50.000	12.2
Methoxychlor	6.754	6.655	6.855	49.550	50.000	-0.9
Tetrachloro-m-xylene	2.881	2.780	2.980	57.330	50.000	14.7