

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/03/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089338.D **Time Analyzed:** 16:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.706	6.605	6.805	54.570	50.000	9.1
4,4'-DDE	6.196	6.096	6.296	50.560	50.000	1.1
4,4'-DDT	7.022	6.921	7.121	44.490	50.000	-11.0
Aldrin	5.271	5.171	5.371	54.740	50.000	9.5
alpha-BHC	4.000	3.899	4.099	55.900	50.000	11.8
alpha-Chlordane	6.027	5.927	6.127	52.800	50.000	5.6
beta-BHC	4.516	4.415	4.615	52.760	50.000	5.5
Decachlorobiphenyl	9.074	8.972	9.172	46.330	50.000	-7.3
delta-BHC	4.765	4.664	4.864	57.140	50.000	14.3
Dieldrin	6.348	6.247	6.447	52.530	50.000	5.1
Endosulfan I	6.075	5.975	6.175	52.890	50.000	5.8
Endosulfan II	6.787	6.686	6.886	51.760	50.000	3.5
Endosulfan sulfate	7.151	7.049	7.249	50.190	50.000	0.4
Endrin	6.575	6.475	6.675	49.270	50.000	-1.5
Endrin aldehyde	6.916	6.815	7.015	50.280	50.000	0.6
Endrin ketone	7.632	7.530	7.730	50.190	50.000	0.4
gamma-BHC (Lindane)	4.331	4.230	4.430	55.390	50.000	10.8
gamma-Chlordane	5.947	5.846	6.046	53.600	50.000	7.2
Heptachlor	4.930	4.829	5.029	52.380	50.000	4.8
Heptachlor epoxide	5.691	5.591	5.791	52.180	50.000	4.4
Methoxychlor	7.494	7.393	7.593	50.700	50.000	1.4
Tetrachloro-m-xylene	3.551	3.450	3.650	55.250	50.000	10.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q2487
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/03/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089338.D **Time Analyzed:** 16:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.929	5.830	6.030	52.350	50.000	4.7
4,4'-DDE	5.375	5.275	5.475	52.440	50.000	4.9
4,4'-DDT	6.184	6.084	6.284	45.480	50.000	-9.0
Aldrin	4.369	4.269	4.469	54.440	50.000	8.9
alpha-BHC	3.393	3.293	3.493	56.730	50.000	13.5
alpha-Chlordane	5.190	5.091	5.291	52.940	50.000	5.9
beta-BHC	4.025	3.926	4.126	51.210	50.000	2.4
Decachlorobiphenyl	8.072	7.972	8.172	46.160	50.000	-7.7
delta-BHC	4.262	4.162	4.362	54.020	50.000	8.0
Dieldrin	5.511	5.413	5.613	52.900	50.000	5.8
Endosulfan I	5.246	5.147	5.347	52.850	50.000	5.7
Endosulfan II	6.080	5.981	6.181	51.150	50.000	2.3
Endosulfan sulfate	6.482	6.383	6.583	50.040	50.000	0.1
Endrin	5.789	5.689	5.889	54.140	50.000	8.3
Endrin aldehyde	6.259	6.159	6.359	50.360	50.000	0.7
Endrin ketone	6.991	6.892	7.092	49.800	50.000	-0.4
gamma-BHC (Lindane)	3.728	3.630	3.830	55.020	50.000	10.0
gamma-Chlordane	5.125	5.026	5.226	53.820	50.000	7.6
Heptachlor	4.082	3.983	4.183	52.110	50.000	4.2
Heptachlor epoxide	4.872	4.773	4.973	56.280	50.000	12.6
Methoxychlor	6.754	6.655	6.855	52.330	50.000	4.7
Tetrachloro-m-xylene	2.881	2.780	2.980	57.920	50.000	15.8