

Surrogate Summary

SDG No.: Q2487

Client: Walsh Construction Company II, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD088990.D	PIBLK-PD088990.D	Decachlorobiphen	1	20	18.2	91		57	171
		Tetrachloro-m-xyl	1	20	16.1	80		61	148
		Decachlorobiphen	2	20	18.6	93		57	171
		Tetrachloro-m-xyl	2	20	17.7	88		61	148
I.BLK-PD089326.D	PIBLK-PD089326.D	Decachlorobiphen	1	20	19.1	96		57	171
		Tetrachloro-m-xyl	1	20	19.5	97		61	148
		Decachlorobiphen	2	20	18.0	90		57	171
		Tetrachloro-m-xyl	2	20	21.9	109		61	148
PB168725BL	PB168725BL	Decachlorobiphen	1	20	18.8	94		20	144
		Tetrachloro-m-xyl	1	20	18.4	92		19	148
		Decachlorobiphen	2	20	19.3	97		20	144
		Tetrachloro-m-xyl	2	20	20.8	104		19	148
PB168725BS	PB168725BS	Decachlorobiphen	1	20	18.4	92		20	144
		Tetrachloro-m-xyl	1	20	17.6	88		19	148
		Decachlorobiphen	2	20	18.9	95		20	144
		Tetrachloro-m-xyl	2	20	19.3	96		19	148
Q2487-09	G4(0-6)	Decachlorobiphen	1	20	10.1	51		20	144
		Tetrachloro-m-xyl	1	20	11.4	57		19	148
		Decachlorobiphen	2	20	8.83	44		20	144
		Tetrachloro-m-xyl	2	20	12.9	64		19	148
Q2487-10	G4(6-12)	Decachlorobiphen	1	20	6.38	32		20	144
		Tetrachloro-m-xyl	1	20	11.8	59		19	148
		Decachlorobiphen	2	20	5.93	30		20	144
		Tetrachloro-m-xyl	2	20	13.0	65		19	148
I.BLK-PD089337.D	PIBLK-PD089337.D	Decachlorobiphen	1	20	19.0	95		57	171
		Tetrachloro-m-xyl	1	20	19.8	99		61	148
		Decachlorobiphen	2	20	19.1	95		57	171
		Tetrachloro-m-xyl	2	20	22.4	112		61	148
I.BLK-PD089340.D	PIBLK-PD089340.D	Decachlorobiphen	1	20	20.0	100		57	171
		Tetrachloro-m-xyl	1	20	18.7	93		61	148
		Decachlorobiphen	2	20	20.0	100		57	171
		Tetrachloro-m-xyl	2	20	21.2	106		61	148
Q2487-11	G3(0-6)	Decachlorobiphen	1	20	16.7	84		20	144
		Tetrachloro-m-xyl	1	20	10.6	53		19	148
		Decachlorobiphen	2	20	9.73	49		20	144
		Tetrachloro-m-xyl	2	20	13.8	69		19	148
Q2487-12	G3(6-12)	Decachlorobiphen	1	20	11.6	58		20	144
		Tetrachloro-m-xyl	1	20	16.3	81		19	148
		Decachlorobiphen	2	20	8.80	44		20	144
		Tetrachloro-m-xyl	2	20	18.6	93		19	148
Q2487-13	G2(0-6)	Decachlorobiphen	1	20	17.2	86		20	144

Surrogate Summary

SDG No.: Q2487

Client: Walsh Construction Company II, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q2487-13	G2(0-6)	Tetrachloro-m-xyl	1	20	10.2	51		19	148
		Decachlorobiphen	2	20	8.79	44		20	144
Q2487-14	G2(6-12)	Tetrachloro-m-xyl	2	20	14.1	71		19	148
		Decachlorobiphen	1	20	8.97	45		20	144
		Tetrachloro-m-xyl	1	20	14.1	71		19	148
		Decachlorobiphen	2	20	7.01	35		20	144
Q2487-15	G1(0-6)	Tetrachloro-m-xyl	2	20	15.6	78		19	148
		Decachlorobiphen	1	20	25.0	125		20	144
		Tetrachloro-m-xyl	1	20	10.7	54		19	148
		Decachlorobiphen	2	20	13.7	69		20	144
Q2487-16	G1(6-12)	Tetrachloro-m-xyl	2	20	14.5	72		19	148
		Decachlorobiphen	1	20	7.13	36		20	144
		Tetrachloro-m-xyl	1	20	10.0	50		19	148
		Decachlorobiphen	2	20	6.03	30		20	144
Q2493-01MS	WC-11MS	Tetrachloro-m-xyl	2	20	12.3	61		19	148
		Decachlorobiphen	1	20	12.6	63		20	144
		Tetrachloro-m-xyl	1	20	16.2	81		19	148
		Decachlorobiphen	2	20	10.9	55		20	144
Q2493-01MSD	WC-11MSD	Tetrachloro-m-xyl	2	20	18.4	92		19	148
		Decachlorobiphen	1	20	13.3	66		20	144
		Tetrachloro-m-xyl	1	20	16.4	82		19	148
		Decachlorobiphen	2	20	11.4	57		20	144
I.BLK-PD089356.D	PIBLK-PD089356.D	Tetrachloro-m-xyl	2	20	18.0	90		19	148
		Decachlorobiphen	1	20	17.9	89		57	171
		Tetrachloro-m-xyl	1	20	19.0	95		61	148
		Decachlorobiphen	2	20	16.4	82		57	171
		Tetrachloro-m-xyl	2	20	22.2	111		61	148