

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2487</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Walsh Construction Company II, LLC</u>	DataFile :	<u>PD089351.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits	
			Result								High	RPD
Lab Sample ID:	Q2493-01MS (Column 1)	Client Sample ID:		WC-11MS								
	alpha-BHC	18.78	0	16.6	ug/kg	88				60	144	
	beta-BHC	18.78	0	16.3	ug/kg	87				54	143	
	delta-BHC	18.78	0	17.6	ug/kg	94				29	151	
	gamma-BHC (Lindane)	18.78	0	17.1	ug/kg	91				61	140	
	Heptachlor	18.78	0	16.5	ug/kg	88				63	135	
	Aldrin	18.78	0	15.8	ug/kg	84				49	139	
	Heptachlor epoxide	18.78	0	15.2	ug/kg	81				41	156	
	Endosulfan I	18.78	0	15.1	ug/kg	80				56	142	
	Dieldrin	18.78	0	14.6	ug/kg	78				47	161	
	4,4'-DDE	18.78	0	14.8	ug/kg	79				55	136	
	Endrin	18.78	0	14.5	ug/kg	77				57	139	
	Endosulfan II	18.78	0	13.7	ug/kg	73				40	163	
	4,4'-DDD	18.78	0	14.3	ug/kg	76				47	163	
	Endosulfan sulfate	18.78	0	13.6	ug/kg	72				62	139	
	4,4'-DDT	18.78	0.2176	13.7	ug/kg	79				51	146	
	Methoxychlor	18.78	0	12.3	ug/kg	65				54	136	
	Endrin ketone	18.78	0	13.3	ug/kg	71				60	129	
	Endrin aldehyde	18.78	0	14.1	ug/kg	75				59	132	
	alpha-Chlordane	18.78	0	15.1	ug/kg	80				39	166	
	gamma-Chlordane	18.78	0	15.0	ug/kg	80				44	175	
Lab Sample ID:	Q2493-01MS (Column 2)	Client Sample ID:		WC-11MS								
	alpha-BHC	18.78	0	18.3	ug/kg	97				60	144	
	beta-BHC	18.78	0	18.1	ug/kg	96				54	143	
	delta-BHC	18.78	0	18.1	ug/kg	96				29	151	
	gamma-BHC (Lindane)	18.78	0	18.0	ug/kg	96				61	140	
	Heptachlor	18.78	0	17.9	ug/kg	95				63	135	
	Aldrin	18.78	0	17.9	ug/kg	95				49	139	
	Heptachlor epoxide	18.78	0	18.2	ug/kg	97				41	156	
	Endosulfan I	18.78	0	17.0	ug/kg	91				56	142	
	Dieldrin	18.78	0	16.7	ug/kg	89				47	161	
	4,4'-DDE	18.78	0.2326	16.8	ug/kg	89				55	136	
	Endrin	18.78	0	16.3	ug/kg	87				57	139	
	Endosulfan II	18.78	0	15.4	ug/kg	82				40	163	
	4,4'-DDD	18.78	0	14.7	ug/kg	78				47	163	
	Endosulfan sulfate	18.78	0	14.7	ug/kg	78				62	139	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2487</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Walsh Construction Company II, LLC</u>	DataFile :	<u>PD089351.D</u>

Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits	
		Result	Result							High	RPD
4,4'-DDT	18.78	0.3827	15.3	ug/kg	81				51	146	
Methoxychlor	18.78	0	13.3	ug/kg	71				54	136	
Endrin ketone	18.78	0	12.6	ug/kg	67				60	129	
Endrin aldehyde	18.78	0	14.9	ug/kg	79				59	132	
alpha-Chlordane	18.78	0	17.1	ug/kg	91				39	166	
gamma-Chlordane	18.78	0	17.2	ug/kg	92				44	175	