

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2487</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Walsh Construction Company II, LLC</u>	Datafile :	<u>PD089334.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168725BS (Column 1)	alpha-BHC	16.66	18.2	ug/kg	109				84	123	
	beta-BHC	16.66	17.1	ug/kg	103				82	123	
	delta-BHC	16.66	19.0	ug/kg	114				83	126	
	gamma-BHC (Lindane)	16.66	18.0	ug/kg	108				83	125	
	Heptachlor	16.66	17.4	ug/kg	104				83	122	
	Aldrin	16.66	18.3	ug/kg	110				82	124	
	Heptachlor epoxide	16.66	17.9	ug/kg	107				83	120	
	Endosulfan I	16.66	18.1	ug/kg	109				81	124	
	Dieldrin	16.66	18.3	ug/kg	110				85	121	
	4,4'-DDE	16.66	18.1	ug/kg	109				81	123	
	Endrin	16.66	16.1	ug/kg	97				76	130	
	Endosulfan II	16.66	18.4	ug/kg	110				80	125	
	4,4'-DDD	16.66	19.1	ug/kg	115				80	131	
	Endosulfan sulfate	16.66	18.0	ug/kg	108				81	122	
	4,4'-DDT	16.66	15.0	ug/kg	90				70	129	
	Methoxychlor	16.66	14.0	ug/kg	84				60	119	
	Endrin ketone	16.66	18.2	ug/kg	109				77	132	
	Endrin aldehyde	16.66	18.3	ug/kg	110				79	124	
	alpha-Chlordane	16.66	18.0	ug/kg	108				84	120	
	gamma-Chlordane	16.66	18.5	ug/kg	111				83	122	
PB168725BS (Column 2)	alpha-BHC	16.66	18.5	ug/kg	111				84	123	
	beta-BHC	16.66	17.9	ug/kg	107				82	123	
	delta-BHC	16.66	18.3	ug/kg	110				83	126	
	gamma-BHC (Lindane)	16.66	18.1	ug/kg	109				83	125	
	Heptachlor	16.66	17.4	ug/kg	104				83	122	
	Aldrin	16.66	18.4	ug/kg	110				82	124	
	Heptachlor epoxide	16.66	18.3	ug/kg	110				83	120	
	Endosulfan I	16.66	18.2	ug/kg	109				81	124	
	Dieldrin	16.66	18.1	ug/kg	109				85	121	
	4,4'-DDE	16.66	18.0	ug/kg	108				81	123	
	Endrin	16.66	16.0	ug/kg	96				76	130	
	Endosulfan II	16.66	18.1	ug/kg	109				80	125	
	4,4'-DDD	16.66	18.4	ug/kg	110				80	131	
	Endosulfan sulfate	16.66	18.0	ug/kg	108				81	122	
	4,4'-DDT	16.66	15.7	ug/kg	94				70	129	
	Methoxychlor	16.66	14.7	ug/kg	88				60	119	
	Endrin ketone	16.66	18.4	ug/kg	110				77	132	
	Endrin aldehyde	16.66	18.1	ug/kg	109				79	124	
	alpha-Chlordane	16.66	18.0	ug/kg	108				84	120	
	gamma-Chlordane	16.66	18.1	ug/kg	109				83	122	