

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG No.: Q2487

Instrument ID: BNA_F

Calibration Date(s): 06/19/2025 06/19/2025

Calibration Time(s): 17:07 20:40

LAB FILE ID:		RRF2.5 = BF142788.D			RRF005 = BF142789.D		RRF010 = BF142790.D	
		RRF020 = BF142791.D			RRF040 = BF142792.D		RRF050 = BF142793.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.238	1.230	1.230	1.149	1.263	1.201	4.3
Benzaldehyde			0.988	0.943	0.757	0.832	0.841	15.0
Phenol-d6		1.512	1.481	1.423	1.340	1.480	1.417	5.5
Phenol		1.648	1.653	1.591	1.516	1.651	1.575	5.3
bis(2-Chloroethyl) ether		1.167	1.133	1.134	1.072	1.183	1.126	3.7
2-Chlorophenol		1.362	1.303	1.292	1.245	1.366	1.296	4.1
2-Methylphenol		1.015	0.997	0.980	0.938	1.043	0.982	4.1
2,2-oxybis(1-Chloropropane)		1.975	1.869	1.815	1.723	1.879	1.809	6.1
Acetophenone		0.464	0.451	0.459	0.418	0.461	0.441	5.3
3+4-Methylphenols			1.292	1.262	1.176	1.312	1.224	6.4
n-Nitroso-di-n-propylamine	0.831	0.879	0.840	0.833	0.787	0.847	0.824	4.3
Nitrobenzene-d5		0.369	0.362	0.373	0.354	0.386	0.365	3.5
Hexachloroethane		0.533	0.521	0.507	0.499	0.539	0.511	3.9
Nitrobenzene		0.345	0.325	0.332	0.319	0.349	0.330	3.9
Isophorone		0.611	0.582	0.592	0.555	0.616	0.585	4.2
2-Nitrophenol		0.169	0.172	0.182	0.178	0.194	0.180	4.8
2,4-Dimethylphenol		0.322	0.310	0.318	0.300	0.329	0.311	4.0
bis(2-Chloroethoxy) methane		0.386	0.373	0.383	0.355	0.390	0.372	4.1
2,4-Dichlorophenol		0.291	0.280	0.290	0.273	0.298	0.282	3.8
Naphthalene		1.051	1.002	1.004	0.934	1.017	0.979	5.5
4-Chloroaniline		0.414	0.408	0.409	0.379	0.421	0.398	5.3
Hexachlorobutadiene		0.195	0.188	0.198	0.184	0.200	0.190	4.3
Caprolactam			0.073	0.075	0.072	0.081	0.075	5.8
4-Chloro-3-methylphenol		0.300	0.275	0.286	0.269	0.298	0.281	5.2
2-Methylnaphthalene		0.655	0.621	0.629	0.579	0.631	0.607	6.1
Hexachlorocyclopentadiene			0.235	0.308	0.336	0.365	0.335	17.0
2,4,6-Trichlorophenol		0.385	0.380	0.394	0.365	0.408	0.385	3.7
2-Fluorobiphenyl		1.740	1.636	1.592	1.430	1.500	1.522	9.2
2,4,5-Trichlorophenol		0.407	0.403	0.421	0.391	0.416	0.405	3.0
1,1-Biphenyl		1.701	1.616	1.637	1.505	1.609	1.580	5.3
2-Chloronaphthalene		1.300	1.202	1.208	1.129	1.209	1.186	5.5
2-Nitroaniline		0.321	0.329	0.337	0.324	0.358	0.332	3.9
Dimethylphthalate		1.372	1.334	1.330	1.228	1.335	1.295	5.2
Acenaphthylene		2.110	2.006	2.011	1.861	1.995	1.952	5.6

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG No.: Q2487

Instrument ID: BNA_F

Calibration Date(s): 06/19/2025 06/19/2025

Calibration Time(s): 17:07 20:40

LAB FILE ID:		RRF2.5 = BF142788.D			RRF005 = BF142789.D		RRF010 = BF142790.D	
		RRF020 = BF142791.D			RRF040 = BF142792.D		RRF050 = BF142793.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.297	0.279	0.290	0.277	0.294	0.284	3.6
3-Nitroaniline		0.311	0.311	0.316	0.302	0.338	0.313	4.5
Acenaphthene		1.247	1.199	1.241	1.142	1.228	1.191	4.4
2,4-Dinitrophenol			0.095	0.129	0.142	0.173	0.142	19.6
4-Nitrophenol			0.189	0.218	0.208	0.238	0.214	8.3
Dibenzofuran		1.844	1.772	1.763	1.626	1.752	1.708	6.1
2,4-Dinitrotoluene		0.373	0.378	0.397	0.367	0.407	0.378	5.9
Diethylphthalate		1.328	1.302	1.319	1.211	1.318	1.269	5.5
4-Chlorophenyl-phenylether		0.713	0.656	0.669	0.599	0.640	0.636	7.6
Fluorene		1.464	1.412	1.395	1.260	1.362	1.334	7.6
4-Nitroaniline		0.279	0.276	0.297	0.272	0.316	0.286	6.5
4,6-Dinitro-2-methylphenol			0.099	0.119	0.119	0.134	0.121	10.0
n-Nitrosodiphenylamine		0.736	0.686	0.710	0.670	0.717	0.693	4.1
2,4,6-Tribromophenol		0.208	0.207	0.214	0.200	0.219	0.206	5.0
4-Bromophenyl-phenylether		0.231	0.221	0.235	0.218	0.236	0.228	3.0
Hexachlorobenzene		0.263	0.248	0.257	0.241	0.262	0.253	3.3
Atrazine		0.176	0.172	0.180	0.174	0.195	0.179	4.5
Pentachlorophenol			0.102	0.119	0.125	0.142	0.126	11.3
Phenanthrene		1.183	1.113	1.115	1.031	1.122	1.083	6.2
Anthracene		1.226	1.123	1.149	1.064	1.140	1.111	6.3
Carbazole		1.040	0.986	0.999	0.900	1.002	0.959	6.7
Di-n-butylphthalate		0.997	1.006	1.045	0.946	1.073	0.999	4.9
Fluoranthene		1.136	1.088	1.090	0.954	1.059	1.028	8.4
Pyrene		2.110	1.940	1.968	1.742	2.030	1.875	11.4
Terphenyl-d14		1.691	1.556	1.533	1.327	1.535	1.447	13.3
Butylbenzylphthalate		0.478	0.470	0.538	0.547	0.613	0.544	10.0
3,3-Dichlorobenzidine			0.398	0.440	0.429	0.476	0.438	5.7
Benzo (a) anthracene		1.392	1.358	1.366	1.265	1.403	1.349	4.0
Chrysene		1.243	1.191	1.208	1.164	1.295	1.204	4.2
Bis (2-ethylhexyl) phthalate		0.650	0.675	0.767	0.803	0.861	0.782	11.5
Di-n-octyl phthalate			1.207	1.352	1.463	1.610	1.475	11.5
Benzo (b) fluoranthene		1.164	1.220	1.226	1.072	1.152	1.157	5.6
Benzo (k) fluoranthene		1.206	1.004	1.035	1.037	1.193	1.083	7.5
Benzo (a) pyrene		1.142	1.065	1.124	1.069	1.201	1.118	4.4
Indeno (1,2,3-cd) pyrene		1.528	1.488	1.545	1.474	1.648	1.523	4.4

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG No.: Q2487

Instrument ID: BNA_F

Calibration Date(s): 06/19/2025 06/19/2025

Calibration Time(s): 17:07 20:40

LAB FILE ID:		RRF2.5 = BF142788.D			RRF005 = BF142789.D		RRF010 = BF142790.D	
		RRF020 = BF142791.D			RRF040 = BF142792.D		RRF050 = BF142793.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo (a,h) anthracene		1.249	1.214	1.293	1.180	1.337	1.238	5.1
Benzo (g,h,i) perylene		1.239	1.198	1.266	1.188	1.332	1.234	4.7
1,2,4,5-Tetrachlorobenzene		0.626	0.609	0.602	0.557	0.598	0.590	4.3
1,4-Dioxane		0.479	0.499	0.471	0.461	0.509	0.480	3.6
2,3,4,6-Tetrachlorophenol		0.327	0.327	0.339	0.312	0.345	0.326	4.4