

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>WALS01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2487</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>07/08/2025 10:54</u>
Lab File ID:	<u>BF143025.D</u>	Init. Calib. Date(s):	<u>06/19/2025 06/19/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>17:07 20:40</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.201	1.183		-1.5	
Benzaldehyde	0.841	0.903		7.4	
Phenol-d6	1.417	1.384		-2.3	
Phenol	1.575	1.550		-1.6	20.0
bis(2-Chloroethyl)ether	1.126	1.102		-2.1	
2-Chlorophenol	1.296	1.276		-1.5	
2-Methylphenol	0.982	0.973		-0.9	
2,2-oxybis(1-Chloropropane)	1.809	1.645		-9.1	
Acetophenone	0.441	0.431		-2.3	
3+4-Methylphenols	1.224	1.225		0.1	
n-Nitroso-di-n-propylamine	0.824	0.805	0.050	-2.3	
Nitrobenzene-d5	0.365	0.385		5.5	
Hexachloroethane	0.511	0.498		-2.5	
Nitrobenzene	0.330	0.315		-4.5	
Isophorone	0.585	0.566		-3.2	
2-Nitrophenol	0.180	0.181		0.6	20.0
2,4-Dimethylphenol	0.311	0.301		-3.2	
bis(2-Chloroethoxy)methane	0.372	0.361		-3.0	
2,4-Dichlorophenol	0.282	0.275		-2.5	20.0
Naphthalene	0.979	0.952		-2.8	
4-Chloroaniline	0.398	0.391		-1.8	
Hexachlorobutadiene	0.190	0.185		-2.6	20.0
Caprolactam	0.075	0.078		4.0	
4-Chloro-3-methylphenol	0.281	0.276		-1.8	20.0
2-Methylnaphthalene	0.607	0.594		-2.1	
Hexachlorocyclopentadiene	0.335	0.341	0.050	1.8	
2,4,6-Trichlorophenol	0.385	0.367		-4.7	20.0
2-Fluorobiphenyl	1.522	1.552		2.0	
2,4,5-Trichlorophenol	0.405	0.383		-5.4	
1,1-Biphenyl	1.580	1.507		-4.6	
2-Chloronaphthalene	1.186	1.125		-5.1	
2-Nitroaniline	0.332	0.326		-1.8	
Dimethylphthalate	1.295	1.287		-0.6	
Acenaphthylene	1.952	1.854		-5.0	
2,6-Dinitrotoluene	0.284	0.279		-1.8	
3-Nitroaniline	0.313	0.304		-2.9	
Acenaphthene	1.191	1.180		-0.9	20.0
2,4-Dinitrophenol	0.142	0.188	0.050	32.4	
4-Nitrophenol	0.214	0.259	0.050	21.0	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>WALS01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2487</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>07/08/2025 10:54</u>
Lab File ID:	<u>BF143025.D</u>	Init. Calib. Date(s):	<u>06/19/2025 06/19/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>17:07 20:40</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.708	1.634		-4.3	
2,4-Dinitrotoluene	0.378	0.376		-0.5	
Diethylphthalate	1.269	1.241		-2.2	
4-Chlorophenyl-phenylether	0.636	0.623		-2.0	
Fluorene	1.334	1.302		-2.4	
4-Nitroaniline	0.286	0.282		-1.4	
4,6-Dinitro-2-methylphenol	0.121	0.120		-0.8	
n-Nitrosodiphenylamine	0.693	0.683		-1.4	20.0
2,4,6-Tribromophenol	0.206	0.200		-2.9	
4-Bromophenyl-phenylether	0.228	0.230		0.9	
Hexachlorobenzene	0.253	0.251		-0.8	
Atrazine	0.179	0.189		5.6	
Pentachlorophenol	0.126	0.185		46.8	20.0
Phenanthrene	1.083	1.051		-3.0	
Anthracene	1.111	1.089		-2.0	
Carbazole	0.959	0.908		-5.3	
Di-n-butylphthalate	0.999	1.022		2.3	
Fluoranthene	1.028	0.959		-6.7	20.0
Pyrene	1.875	2.136		13.9	
Terphenyl-d14	1.447	1.536		6.2	
Butylbenzylphthalate	0.544	0.573		5.3	
3,3-Dichlorobenzidine	0.438	0.444		1.4	
Benzo (a) anthracene	1.349	1.310		-2.9	
Chrysene	1.204	1.161		-3.6	
Bis(2-ethylhexyl)phthalate	0.782	0.775		-0.9	
Di-n-octyl phthalate	1.475	1.479		0.3	20.0
Benzo (b) fluoranthene	1.157	1.133		-2.1	
Benzo (k) fluoranthene	1.083	0.990		-8.6	
Benzo (a) pyrene	1.118	1.090		-2.5	20.0
Indeno (1,2,3-cd) pyrene	1.523	1.509		-0.9	
Dibenzo (a,h) anthracene	1.238	1.206		-2.6	
Benzo (g,h,i) perylene	1.234	1.198		-2.9	
1,2,4,5-Tetrachlorobenzene	0.590	0.574		-2.7	
1,4-Dioxane	0.480	0.434		-9.6	20.0
2,3,4,6-Tetrachlorophenol	0.326	0.307		-5.8	

All other compounds must meet a minimum RRF of 0.010.