

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2487

Client ID : SSTDCCC040

Date Analyzed: 07/08/2025

Lab File ID: BF143025.D

Time Analyzed: 10:54

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	66813	6.875	257725	8.16	136330	9.92
UPPER LIMIT	133626	7.375	515450	8.663	272660	10.422
LOWER LIMIT	33406.5	6.375	128863	7.663	68165	9.422
EPA SAMPLE NO.						
01 PB168722BS	56971	6.88	220160	8.16	116094	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.