

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2487

Client ID : SSTDCCC040

Date Analyzed: 07/10/2025

Lab File ID: BP025082.D

Time Analyzed: 14:04

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	409438	7.425	1701300	10.18	1080410	14.08
UPPER LIMIT	818876	7.925	3402600	10.678	2160820	14.584
LOWER LIMIT	204719	6.925	850650	9.678	540205	13.584
EPA SAMPLE NO.						
01 G2 (0-6) DL	333631	7.44	1470190	10.19	1059540	14.09

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.