

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2487

Client ID : SSTDCCC040

Date Analyzed: 07/15/2025

Lab File ID: BP025138.D

Time Analyzed: 09:36

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	539286	7.443	1912690	10.18	1092260	14.08
UPPER LIMIT	1078570	7.943	3825380	10.684	2184520	14.584
LOWER LIMIT	269643	6.943	956345	9.684	546130	13.584
EPA SAMPLE NO.						
01 G2 (0-6) DL2	679768	7.44	2796860	10.19	1841260	14.08

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.