

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2487

Client ID: SSTDCCC040

Date Analyzed: 07/10/2025

Lab File ID: BP025082.D

Time Analyzed: 14:04

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	2198800	16.901	2306390	21.336	2492510	24.483
UPPER LIMIT	4397600	17.401	4612780	21.836	4985020	24.983
LOWER LIMIT	1099400	16.401	1153200	20.836	1246260	23.983
EPA SAMPLE NO.						
G2 (0-6) DL	2622860	16.91	3703130	21.34	4039720	24.48

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.