

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2487

Client ID: SSTDCCC040

Date Analyzed: 07/15/2025

Lab File ID: BP025138.D

Time Analyzed: 09:36

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	2112950	16.89	2423210	21.336	2977610	24.466
UPPER LIMIT	4225900	17.39	4846420	21.836	5955220	24.966
LOWER LIMIT	1056480	16.39	1211610	20.836	1488810	23.966
EPA SAMPLE NO.						
01 G2 (0-6)DL2	3743450	16.91	3883440	21.34	4506340	24.50

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.