

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF143002.D
Instrument ID: BNA_F

Contract: WALS01
SDG NO.: Q2487
DFTPP Injection Date: 07/07/2025
DFTPP Injection Time: 11:12

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	29.4
70	Less than 2.0% of mass 69	0.0 (0.0) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	15.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19 (19) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF143003.D	07/07/2025	12:12
PB168722BL	PB168722BL	BF143004.D	07/07/2025	12:42
WC-11MS	Q2493-01MS	BF143006.D	07/07/2025	13:47
WC-11MSD	Q2493-01MSD	BF143007.D	07/07/2025	14:17
G3 (6-12)	Q2487-12	BF143008.D	07/07/2025	14:47
G4 (6-12)	Q2487-10	BF143009.D	07/07/2025	15:17
G2 (6-12)	Q2487-14	BF143010.D	07/07/2025	15:48